

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090421\
 Data File : BN016194.D
 Acq On : 04 Sep 2021 15:01
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN090421

Quant Time: Sep 06 02:12:13 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN090421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 06 02:08:58 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2	1,4-Dioxane	0.128	0.131	-2.3	94	0.00
3	n-Nitrosodimethylamine	0.311	0.297	4.5	96	0.00
4 S	2-Fluorophenol	1.129	1.147	-1.6	100	0.00
5 S	Phenol-d6	1.443	1.457	-1.0	101	0.00
6	bis(2-Chloroethyl)ether	1.044	1.072	-2.7	100	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	98	0.00
8 S	Nitrobenzene-d5	0.321	0.310	3.4	101	0.00
9	Naphthalene	1.059	1.076	-1.6	100	0.00
10	Hexachlorobutadiene	0.215	0.221	-2.8	100	0.00
11 SURR	2-Methylnaphthalene-d10	0.637	0.650	-2.0	99	0.00
12	2-Methylnaphthalene	0.714	0.730	-2.2	100	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	97	0.00
14 S	2,4,6-Tribromophenol	0.179	0.163	8.9	103	0.00
15 S	2-Fluorobiphenyl	1.508	1.480	1.9	98	0.00
16	Acenaphthylene	1.794	1.792	0.1	99	0.00
17	Acenaphthene	1.136	1.133	0.3	98	0.00
18	Fluorene	1.427	1.426	0.1	99	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	96	0.00
20	4,6-Dinitro-2-methylphenol	0.028	0.011	60.7#	224#	0.00
21	4-Bromophenyl-phenylether	0.211	0.207	1.9	96	0.00
22	Hexachlorobenzene	0.249	0.249	0.0	97	0.00
23	Atrazine	0.214	0.206	3.7	97	0.00
24	Pentachlorophenol	0.064	0.039	39.1#	137	-0.01
25	Phenanthrene	1.120	1.126	-0.5	98	0.00
26	Anthracene	1.068	1.060	0.7	98	-0.01
27 SURR	Fluoranthene-d10	1.201	1.205	-0.3	97	0.00
28	Fluoranthene	1.328	1.350	-1.7	97	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	98	0.01
30	Pyrene	1.303	1.279	1.8	97	0.00
31 S	Terphenyl-d14	1.011	0.959	5.1	97	0.00
32	Benzo(a)anthracene	1.343	1.305	2.8	98	0.00
33	Chrysene	1.305	1.290	1.1	99	0.00
34	Bis(2-ethylhexyl)phthalate	0.904	0.850	6.0	98	0.00
35 I	Perylene-d12	1.000	1.000	0.0	99	0.00
36	Indeno(1,2,3-cd)pyrene	1.113	1.079	3.1	111	0.00
37	Benzo(b)fluoranthene	1.366	1.388	-1.6	100	0.00
38	Benzo(k)fluoranthene	1.251	1.283	-2.6	101	0.00
39 C	Benzo(a)pyrene	1.212	1.212	0.0	101	0.00
40	Dibenzo(a,h)anthracene	0.929	0.905	2.6	113	0.00
41	Benzo(g,h,i)perylene	0.935	0.899	3.9	110	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN090421