

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090421\
 Data File : BN016222.D
 Acq On : 05 Sep 2021 09:46
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 06 06:22:58 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN090421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 06 02:08:58 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	116	0.00
2	1,4-Dioxane	0.128	0.115	10.2	97	0.00
3	n-Nitrosodimethylamine	0.311	0.308	1.0	117	0.00
4 S	2-Fluorophenol	1.129	1.108	1.9	113	0.00
5 S	Phenol-d6	1.443	1.408	2.4	115	0.02
6	bis(2-Chloroethyl)ether	1.044	1.083	-3.7	119	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	117	0.00
8 S	Nitrobenzene-d5	0.321	0.297	7.5	115	0.00
9	Naphthalene	1.059	1.045	1.3	116	0.00
10	Hexachlorobutadiene	0.215	0.217	-0.9	117	0.00
11 SURR	2-Methylnaphthalene-d10	0.637	0.653	-2.5	119	0.00
12	2-Methylnaphthalene	0.714	0.735	-2.9	119	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	120	0.00
14 S	2,4,6-Tribromophenol	0.179	0.193	-7.8	150	0.00
15 S	2-Fluorobiphenyl	1.508	1.455	3.5	119	0.00
16	Acenaphthylene	1.794	1.746	2.7	119	0.00
17	Acenaphthene	1.136	1.118	1.6	119	0.00
18	Fluorene	1.427	1.423	0.3	121	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	123	0.00
20	4,6-Dinitro-2-methylphenol	0.028	0.017	39.3#	462#	0.00
21	4-Bromophenyl-phenylether	0.211	0.206	2.4	123	0.00
22	Hexachlorobenzene	0.249	0.239	4.0	119	0.00
23	Atrazine	0.214	0.192	10.3	116	0.00
24	Pentachlorophenol	0.064	0.053	17.2	238#	-0.01
25	Phenanthrene	1.120	1.094	2.3	121	0.00
26	Anthracene	1.068	1.041	2.5	123	-0.01
27 SURR	Fluoranthene-d10	1.201	1.193	0.7	122	0.00
28	Fluoranthene	1.328	1.328	0.0	123	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	124	0.00
30	Pyrene	1.303	1.280	1.8	124	0.00
31 S	Terphenyl-d14	1.011	0.946	6.4	121	0.00
32	Benzo(a)anthracene	1.343	1.303	3.0	124	0.00
33	Chrysene	1.305	1.285	1.5	125	0.00
34	Bis(2-ethylhexyl)phthalate	0.904	0.803	11.2	117	0.00
35 I	Perylene-d12	1.000	1.000	0.0	118	0.00
36	Indeno(1,2,3-cd)pyrene	1.113	1.102	1.0	135	0.00
37	Benzo(b)fluoranthene	1.366	1.392	-1.9	120	0.00
38	Benzo(k)fluoranthene	1.251	1.309	-4.6	123	0.00
39 C	Benzo(a)pyrene	1.212	1.248	-3.0	124	0.00
40	Dibenzo(a,h)anthracene	0.929	0.942	-1.4	140	0.00
41	Benzo(g,h,i)perylene	0.935	0.848	9.3	124	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4