

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090421\
 Data File : BN016195.D
 Acq On : 04 Sep 2021 16:14
 Operator : CG/JU
 Sample : PB138841BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB138841BL

Quant Time: Sep 06 02:12:27 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN090421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 06 02:08:58 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.873	152	13512	0.400	ng	0.00
7) Naphthalene-d8	10.660	136	71362	0.400	ng	0.00
13) Acenaphthene-d10	14.497	164	41299	0.400	ng	0.00
19) Phenanthrene-d10	17.249	188	81993	0.400	ng	0.00
29) Chrysene-d12	21.429	240	85008	0.400	ng	0.00
35) Perylene-d12	23.823	264	75787	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.475	112	12728	0.334	ng	0.02
5) Phenol-d6	7.043	99	15310	0.314	ng	0.00
8) Nitrobenzene-d5	9.025	82	19765	0.345	ng	0.00
11) 2-Methylnaphthalene-d10	12.252	152	40020	0.352	ng	0.00
14) 2,4,6-Tribromophenol	15.994	330	4573	0.248	ng	0.00
15) 2-Fluorobiphenyl	13.127	172	55901	0.359	ng	0.00
27) Fluoranthene-d10	19.276	212	87029	0.353	ng	0.00
31) Terphenyl-d14	19.872	244	80012	0.372	ng	0.00


Target Compounds	Qvalue
-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090421\
Data File : BN016195.D
Acq On : 04 Sep 2021 16:14
Operator : CG/JU
Sample : PB138841BL
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB138841BL

Quant Time: Sep 06 02:12:27 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN090421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 06 02:08:58 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.873 min Scan# 520
Delta R.T. -0.000 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14




Instrument :
BNA_N
ClientSampleId :
PB138841BL


Tgt Ion:152 Resp: 13512
Ion Ratio Lower Upper
152 100
150 152.9 125.3 187.9
115 59.3 48.4 72.6



#4
2-Fluorophenol
Concen: 0.334 ng
RT: 5.475 min Scan# 260
Delta R.T. 0.018 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14

Tgt Ion:112 Resp: 12728
Ion Ratio Lower Upper
112 100
64 59.3 46.8 70.2
63 28.8 22.7 34.1









#5
Phenol-d6
Concen: 0.314 ng
RT: 7.043 min Scan# 430
Delta R.T. 0.009 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14




Instrument :
BNA_N
ClientSampleId :
PB138841BL


Tgt Ion:	99	Resp:	15310
Ion	Ratio	Lower	Upper
99	100		
42	16.4	13.8	20.6
71	30.0	24.3	36.5



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.660 min Scan# 753
Delta R.T. 0.000 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14

Tgt Ion:	136	Resp:	71362
Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.9	13.3
54	9.0	7.2	10.8
68	6.4	5.2	7.8









#8
Nitrobenzene-d5
Concen: 0.345 ng
RT: 9.025 min Scan# 624
Delta R.T. 0.000 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14




Instrument :
BNA_N
ClientSampleId :
PB138841BL


Tgt Ion: 82 Resp: 19765
Ion Ratio Lower Upper
82 100
128 47.5 39.3 58.9
54 51.5 40.2 60.4



#11
2-Methylnaphthalene-d10
Concen: 0.352 ng
RT: 12.252 min Scan# 987
Delta R.T. 0.000 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14

Tgt Ion: 152 Resp: 40020
Ion Ratio Lower Upper
152 100
151 20.1 16.2 24.2









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.497 min Scan# 1201
Delta R.T. 0.000 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14




Instrument :
BNA_N
ClientSampleId :
PB138841BL


Tgt Ion:164 Resp: 41299
Ion Ratio Lower Upper
164 100
162 99.3 79.8 119.6
160 43.2 34.2 51.4



#14
2,4,6-Tribromophenol
Concen: 0.248 ng
RT: 15.994 min Scan# 1319
Delta R.T. 0.000 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14

Tgt Ion:330 Resp: 4573
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 31.4 25.4 38.2









#15
2-Fluorobiphenyl
Concen: 0.359 ng
RT: 13.127 min Scan# 1093
Delta R.T. 0.000 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14




Instrument :
BNA_N
ClientSampleId :
PB138841BL


Tgt Ion:	172	Resp:	55901
Ion	Ratio	Lower	Upper
172	100		
171	34.0	26.9	40.3
170	21.7	17.1	25.7



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.249 min Scan# 1422
Delta R.T. 0.000 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14

Tgt Ion:	188	Resp:	81993
Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	6.0	5.0	7.6









#27
Fluoranthene-d10
Concen: 0.353 ng
RT: 19.276 min Scan# 1739
Delta R.T. -0.005 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14




Instrument :
BNA_N
ClientSampleId :
PB138841BL


Tgt Ion:212 Resp: 87029
Ion Ratio Lower Upper
212 100
106 12.2 9.7 14.5
104 6.8 5.4 8.0



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.429 min Scan# 2020
Delta R.T. 0.000 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14

Tgt Ion:240 Resp: 85008
Ion Ratio Lower Upper
240 100
120 11.0 10.0 15.0
236 25.8 20.9 31.3









#31
Terphenyl-d14
Concen: 0.372 ng
RT: 19.872 min Scan# 1859
Delta R.T. 0.000 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14

Instrument :
BNA_N
ClientSampleId :
PB138841BL

Tgt Ion:244 Resp: 80012
Ion Ratio Lower Upper
244 100
212 6.9 5.5 8.3
122 10.8 9.0 13.4



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.823 min Scan# 2526
Delta R.T. -0.003 min
Lab File: BN016195.D
Acq: 04 Sep 2021 16:14

Tgt Ion:264 Resp: 75787
Ion Ratio Lower Upper
264 100
260 23.8 19.5 29.3
265 24.7 20.2 30.2

