

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN090519\
 Data File : BN007536.D
 Acq On : 05 Sep 2019 15:34
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 06 02:03:56 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 29 12:18:45 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	182#	-0.02
2	1,4-Dioxane	0.665	0.635	4.5	167#	0.00
3	n-Nitrosodimethylamine	0.309	0.202	34.6#	115	0.00
4 S	2-Fluorophenol	1.490	1.155	22.5	149	-0.02
5 S	Phenol-d6	1.953	1.478	24.3	158#	0.00
6	bis(2-Chloroethyl)ether	1.445	1.551	-7.3	197#	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	205#	-0.03
8 S	Nitrobenzene-d5	0.348	0.290	16.7	178#	-0.01
9	Naphthalene	1.143	1.127	1.4	197#	-0.01
10	Hexachlorobutadiene	0.234	0.202	13.7	170#	-0.01
11 SURR	2-Methylnaphthalene-d10	0.665	0.663	0.3	201#	-0.01
12	2-Methylnaphthalene	0.779	0.782	-0.4	200#	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	222#	-0.02
14 S	2,4,6-Tribromophenol	0.209	0.150	28.2#	162#	-0.01
15 S	2-Fluorobiphenyl	1.613	1.583	1.9	213#	-0.01
16	Acenaphthylene	2.367	1.950	17.6	181#	-0.01
17	Acenaphthene	1.385	1.336	3.5	209#	-0.01
18	Fluorene	1.751	1.715	2.1	213#	-0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	229#	-0.01
20	4-Bromophenyl-phenylether	0.158	0.028	82.3#	46#	0.00
21	Hexachlorobenzene	0.268	0.260	3.0	220#	-0.01
22	Pentachlorophenol	0.083	0.056	32.5#	164#	-0.01
23	Phenanthrene	1.203	1.212	-0.7	230#	-0.01
24	Anthracene	1.210	1.041	14.0	196#	-0.01
25 SURR	Fluoranthene-d10	1.288	1.194	7.3	205#	0.00
26	Fluoranthene	1.545	1.445	6.5	208#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	223#	-0.01
28	Pvrene	1.609	1.539	4.4	208#	0.00
29 S	Terphenyl-d14	1.047	0.991	5.3	204#	-0.01
30	Benzo(a)anthracene	1.571	1.414	10.0	197#	-0.01
31	Chrysene	1.516	1.513	0.2	213#	-0.01
32	Bis(2-ethylhexyl)phthalate	1.177	0.680	42.2#	136	0.00
33	Indeno(1,2,3-cd)pyrene	1.827	1.554	14.9	183#	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	208#	-0.01
35	Benzo(b)fluoranthene	1.449	1.408	2.8	197#	-0.01
36	Benzo(k)fluoranthene	1.432	1.541	-7.6	215#	-0.01
37 C	Benzo(a)pyrene	1.379	1.297	5.9	189#	-0.02
38	Dibenzo(a,h)anthracene	1.376	1.199	12.9	175#	-0.02
39	Benzo(a,h,i)perylene	1.371	1.266	7.7	186#	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN090519\
Data File : BN007536.D
Acq On : 05 Sep 2019 15:34
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 06 02:03:56 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 29 12:18:45 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		