

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN090519\
 Data File : BN007543.D
 Acq On : 05 Sep 2019 21:16
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 06 02:00:30 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 29 12:18:45 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	164#	-0.02
2	1,4-Dioxane	0.665	0.625	6.0	149	0.00
3	n-Nitrosodimethylamine	0.309	0.206	33.3#	106	0.00
4 S	2-Fluorophenol	1.490	1.097	26.4#	128	-0.02
5 S	Phenol-d6	1.953	1.403	28.2#	136	0.00
6	bis(2-Chloroethyl)ether	1.445	1.412	2.3	162#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	184#	-0.01
8 S	Nitrobenzene-d5	0.348	0.267	23.3	148	0.00
9	Naphthalene	1.143	1.130	1.1	177#	-0.01
10	Hexachlorobutadiene	0.234	0.203	13.2	154#	-0.01
11 SURR	2-Methylnaphthalene-d10	0.665	0.669	-0.6	183#	-0.01
12	2-Methylnaphthalene	0.779	0.802	-3.0	184#	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	208#	-0.02
14 S	2,4,6-Tribromophenol	0.209	0.148	29.2#	150	-0.02
15 S	2-Fluorobiphenyl	1.613	1.593	1.2	201#	0.00
16	Acenaphthylene	2.367	1.852	21.8	161#	-0.02
17	Acenaphthene	1.385	1.339	3.3	196#	0.00
18	Fluorene	1.751	1.736	0.9	201#	-0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	215#	0.00
20	4-Bromophenyl-phenylether	0.158	0.061	61.4#	97	0.00
21	Hexachlorobenzene	0.268	0.259	3.4	205#	-0.02
22	Pentachlorophenol	0.083	0.045	45.8#	125	-0.02
23	Phenanthrene	1.203	1.176	2.2	210#	-0.02
24	Anthracene	1.210	1.004	17.0	177#	0.00
25 SURR	Fluoranthene-d10	1.288	1.119	13.1	181#	-0.01
26	Fluoranthene	1.545	1.375	11.0	187#	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	181#	-0.01
28	Pvrene	1.609	1.659	-3.1	183#	-0.01
29 S	Terphenyl-d14	1.047	1.114	-6.4	186#	-0.01
30	Benzo(a)anthracene	1.571	1.340	14.7	152#	-0.01
31	Chrysene	1.516	1.510	0.4	173#	-0.01
32	Bis(2-ethylhexyl)phthalate	1.177	0.620	47.3#	101	-0.01
33	Indeno(1,2,3-cd)pyrene	1.827	1.427	21.9	137	-0.03
34 I	Perylene-d12	1.000	1.000	0.0	151#	-0.02
35	Benzo(b)fluoranthene	1.449	1.439	0.7	146	-0.02
36	Benzo(k)fluoranthene	1.432	1.521	-6.2	154#	-0.02
37 C	Benzo(a)pyrene	1.379	1.247	9.6	133	-0.02
38	Dibenzo(a,h)anthracene	1.376	1.239	10.0	131	-0.03
39	Benzo(a,h,i)perylene	1.371	1.284	6.3	137	-0.03

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN090519\
Data File : BN007543.D
Acq On : 05 Sep 2019 21:16
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 06 02:00:30 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 29 12:18:45 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			