

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090524\
 Data File : BN033760.D
 Acq On : 05 Sep 2024 09:56
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 05 11:27:48 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN090424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 04 18:36:01 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.455	0.444	2.4	101	0.00
3	n-Nitrosodimethylamine	0.534	0.495	7.3	97	0.00
4 S	2-Fluorophenol	1.225	1.079	11.9	92	0.00
5 S	Phenol-d6	1.450	1.257	13.3	93	0.00
6	bis(2-Chloroethyl)ether	1.094	1.054	3.7	99	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
8 S	Nitrobenzene-d5	0.341	0.313	8.2	97	0.00
9	Naphthalene	1.068	1.028	3.7	100	0.00
10	Hexachlorobutadiene	0.222	0.217	2.3	100	-0.01
11 SURR	2-Methylnaphthalene-d10	0.566	0.536	5.3	99	0.00
12	2-Methylnaphthalene	0.655	0.619	5.5	99	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	98	-0.01
14 S	2,4,6-Tribromophenol	0.202	0.164	18.8	87	0.00
15 S	2-Fluorobiphenyl	1.668	1.655	0.8	100	0.00
16	Acenaphthylene	1.729	1.602	7.3	98	0.00
17	Acenaphthene	1.185	1.157	2.4	100	0.00
18	Fluorene	1.462	1.411	3.5	100	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	102	0.00
20	4,6-Dinitro-2-methylphenol	0.063	0.052	17.5	104	0.00
21	4-Bromophenyl-phenylether	0.232	0.219	5.6	99	0.00
22	Hexachlorobenzene	0.266	0.268	-0.8	103	-0.01
23	Atrazine	0.185	0.161	13.0	92	-0.01
24	Pentachlorophenol	0.109	0.093	14.7	96	-0.01
25	Phenanthrene	1.102	1.073	2.6	102	0.00
26	Anthracene	0.958	0.893	6.8	99	-0.01
27 SURR	Fluoranthene-d10	0.987	0.945	4.3	102	0.00
28	Fluoranthene	1.235	1.180	4.5	103	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
30	Pyrene	1.612	1.581	1.9	102	0.00
31 S	Terphenyl-d14	0.853	0.840	1.5	103	0.00
32	Benzo(a)anthracene	1.428	1.375	3.7	104	0.00
33	Chrysene	1.412	1.439	-1.9	109	0.00
34	Bis(2-ethylhexyl)phthalate	0.779	0.659	15.4	91	0.00
35 I	Perylene-d12	1.000	1.000	0.0	109	0.00
36	Indeno(1,2,3-cd)pyrene	1.653	1.640	0.8	110	0.00
37	Benzo(b)fluoranthene	1.470	1.476	-0.4	110	0.00
38	Benzo(k)fluoranthene	1.433	1.422	0.8	110	0.00
39 C	Benzo(a)pyrene	1.229	1.173	4.6	107	0.00
40	Dibenzo(a,h)anthracene	1.325	1.306	1.4	109	0.00
41	Benzo(g,h,i)perylene	1.437	1.449	-0.8	109	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4