

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090524\
 Data File : BN033771.D
 Acq On : 05 Sep 2024 16:50
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 05 17:33:07 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN090424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 04 18:36:01 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	89	0.00
2	1,4-Dioxane	0.400	0.395	1.3	91	0.00
3	n-Nitrosodimethylamine	0.400	0.388	3.0	90	0.00
4 S	2-Fluorophenol	0.400	0.365	8.8	85	0.00
5 S	Phenol-d6	0.400	0.359	10.3	85	0.00
6	bis(2-Chloroethyl)ether	0.400	0.385	3.8	88	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	88	0.00
8 S	Nitrobenzene-d5	0.400	0.362	9.5	85	0.00
9	Naphthalene	0.400	0.385	3.8	89	-0.01
10	Hexachlorobutadiene	0.400	0.395	1.3	89	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.377	5.8	88	0.00
12	2-Methylnaphthalene	0.400	0.380	5.0	88	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	89	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.320	20.0	77	0.00
15 S	2-Fluorobiphenyl	0.400	0.391	2.3	89	0.00
16	Acenaphthylene	0.400	0.365	8.8	86	0.00
17	Acenaphthene	0.400	0.387	3.3	89	0.00
18	Fluorene	0.400	0.387	3.3	90	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	92	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.344	14.0	97	0.00
21	4-Bromophenyl-phenylether	0.400	0.386	3.5	91	0.00
22	Hexachlorobenzene	0.400	0.393	1.8	90	-0.01
23	Atrazine	0.400	0.360	10.0	85	0.00
24	Pentachlorophenol	0.400	0.327	18.3	83	-0.01
25	Phenanthrene	0.400	0.391	2.3	92	0.00
26	Anthracene	0.400	0.371	7.3	89	-0.01
27 SURR	Fluoranthene-d10	0.400	0.384	4.0	91	0.00
28	Fluoranthene	0.400	0.385	3.8	93	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	105	0.00
30	Pyrene	0.400	0.362	9.5	93	0.00
31 S	Terphenyl-d14	0.400	0.361	9.8	93	0.00
32	Benzo(a)anthracene	0.400	0.363	9.3	96	0.00
33	Chrysene	0.400	0.387	3.3	102	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.342	14.5	90	0.00
35 I	Perylene-d12	0.400	0.400	0.0	97	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.392	2.0	97	0.00
37	Benzo(b)fluoranthene	0.400	0.393	1.8	96	0.00
38	Benzo(k)fluoranthene	0.400	0.402	-0.5	99	0.00
39 C	Benzo(a)pyrene	0.400	0.399	0.3	99	0.00
40	Dibenzo(a,h)anthracene	0.400	0.393	1.8	96	0.00
41	Benzo(g,h,i)perylene	0.400	0.395	1.3	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4