



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090524\
Data File : BN033759.D
Acq On : 05 Sep 2024 09:17
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN090424.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Sep 04 18:36:01 2024






AutoFind: Scans 2446, 2447, 2448; Background Corrected with Scan 2439

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	40.8	126318	PASS
68	69	0.00	2	1.7	2254	PASS
69	198	0.00	100	41.7	129271	PASS
70	69	0.00	2	0.6	756	PASS
127	198	10	80	50.2	155427	PASS
197	198	0.00	2	0.2	595	PASS
198	198	100	100	100.0	309717	PASS
199	198	5	9	6.7	20777	PASS
275	198	10	60	26.6	82520	PASS
365	198	1	100	3.7	11363	PASS
441	198	0.01	100	11.0	34200	PASS
442	442	50	100	100.0	192597	PASS
443	442	15	24	21.7	41755	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090524\
Data File : BN033759.D
Acq On : 05 Sep 2024 09:17
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Sep 06 17:05:55 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 31 04:54:27 2024
Response via : Initial Calibration



TIC: BN033759.D\data.ms

(70) Pentachlorophenol (C)

16.563min (-0.023) 22225.18 ng




response 349184

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	64.73
264.00	61.60	61.83
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090524\
Data File : BN033759.D
Acq On : 05 Sep 2024 09:17
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Sep 06 17:05:55 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 31 04:54:27 2024
Response via : Initial Calibration



TIC: BN033759.D\data.ms

(77) Benzidine

19.174min (+ 0.000) 0.00 ng

response 2020794

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.57
183.00	13.20	11.64
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
9/5/2024	BNA_N	<u>BN033759.D</u>
Compound Name	Response	Retention Time
DDT	1070112	20.427
DDD	21723	19.98
DDE	1418	19.468
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
23141	1093253	2.12

Instrument :
BNA_N
ClientSampleId :
DFTPP