

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090622\
 Data File : BN021601.D
 Acq On : 06 Sep 2022 15:08
 Operator : CG/JU
 Sample : PB146659BL
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK659

Quant Time: Sep 06 15:44:16 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Sep 06 14:13:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.063	152	169708	20.000	ng/u1	0.00
20) Naphthalene-d8	10.899	136	768693	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.722	164	490713	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.469	188	985979	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.663	240	819288	20.000	ng/u1	0.00
88) Perylene-d12	24.180	264	694981	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.364	96	28234	6.246	ng/uL	0.00
4) Pyridine-d5	3.793	84	358735	29.120	ng/u1	0.00
7) Phenol-d5	7.216	99	504724	32.791	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.387	67	325858	34.975	ng/u1	0.00
11) 2-Chlorophenol-d4	7.587	132	397277	35.476	ng/u1	0.00
15) 4-Methylphenol-d8	8.775	113	409098	33.874	ng/u1	0.00
21) Nitrobenzene-d5	9.240	128	202047	38.579	ng/u1	0.00
24) 2-Nitrophenol-d4	9.969	143	212659	45.461	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.510	165	375846	36.279	ng/u1	0.00
31) 4-Chloroaniline-d4	11.034	131	608115	34.564	ng/u1	0.00
46) Dimethylphthalate-d6	14.128	166	1274395	37.996	ng/u1	0.00
49) Acenaphthylene-d8	14.416	160	1474267	37.704	ng/u1	0.00
54) 4-Nitrophenol-d4	14.910	143	204609	30.656	ng/u1	0.00
60) Fluorene-d10	15.716	176	1066593	36.547	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.828	200	160274	40.640	ng/u1	0.00
73) Anthracene-d10	17.569	188	1627509	37.457	ng/u1	0.00
81) Pyrene-d10	19.863	212	1789869	45.938	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.015	264	1348067	40.924	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090622\
Data File : BN021601.D
Acq On : 06 Sep 2022 15:08
Operator : CG/JU
Sample : PB146659BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK659

Quant Time: Sep 06 15:44:16 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081922.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Sep 06 14:13:27 2022
Response via : Initial Calibration

