




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090624\
Data File : BN033791.D
Acq On : 06 Sep 2024 21:01
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 07 02:02:12 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 07 02:00:52 2024
Response via : Initial Calibration



TIC: BN033791.D\data.ms




(70) Pentachlorophenol (C)

16.557min (0.000) 18862.89 ng

response 175417

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	63.06
264.00	61.60	62.38
0.00	0.00	0.00

Quant Time: Sep 07 02:02:12 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 07 02:00:52 2024
Response via : Initial Calibration



TIC: BN033791.D\data.ms

(77) Benzidine

19.169min (0.000) 0.00 ng

```
response      1042160
```

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.94
183.00	13.20	11.46
0.00	0.00	0.00