

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090722\
 Data File : BN021630.D
 Acq On : 07 Sep 2022 17:14
 Operator : CG/JU
 Sample : N4497-10
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SVOC-GPC2-BLANK

Quant Time: Sep 08 02:01:54 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 07 01:01:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.063	152	159298	20.000	ng/u1	0.00
20) Naphthalene-d8	10.893	136	685380	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.716	164	467158	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.469	188	1020063	20.000	ng/u1	0.00
79) Chrysene-d12	21.663	240	871252	20.000	ng/u1	0.00
88) Perylene-d12	24.174	264	784434	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.364	96	79	0.019	ng/uL	0.00
4) Pyridine-d5	3.799	84	109	0.009	ng/u1	0.00
7) Phenol-d5	0.000	99	0	0.000	ng/u1	
9) Bis-(2-Chloroethyl)eth...	0.000	67	0	0.000	ng/u1	
11) 2-Chlorophenol-d4	0.000	132	0	0.000	ng/u1	
15) 4-Methylphenol-d8	0.000	113	0	0.000	ng/u1	
21) Nitrobenzene-d5	0.000	128	0	0.000	ng/u1	
24) 2-Nitrophenol-d4	0.000	143	0	0.000	ng/u1	
28) 2,4-Dichlorophenol-d3	0.000	165	0	0.000	ng/u1	
31) 4-Chloroaniline-d4	10.899	131	873	0.056	ng/u1	-0.14
46) Dimethylphthalate-d6	0.000	166	0	0.000	ng/u1	
49) Acenaphthylene-d8	0.000	160	0	0.000	ng/u1	
54) 4-Nitrophenol-d4	0.000	143	0	0.000	ng/u1	
60) Fluorene-d10	15.710	176	148	0.005	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	0.000	200	0	0.000	ng/u1	
73) Anthracene-d10	17.569	188	1875	0.042	ng/u1	0.00
81) Pyrene-d10	19.869	212	223	0.005	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.174	264	784434	21.098	ng/u1	0.16

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090722\
Data File : BN021630.D
Acq On : 07 Sep 2022 17:14
Operator : CG/JU
Sample : N4497-10
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SVOC-GPC2-BLANK

Quant Time: Sep 08 02:01:54 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081922.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Sep 07 01:01:49 2022
Response via : Initial Calibration

