


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090722\
 Data File : BN021619.D
 Acq On : 07 Sep 2022 08:42
 Operator : CG/JU
 Sample : DFTPP250
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP250

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM083022.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Sep 06 13:54:00 2022



AutoFind: Scans 2483, 2484, 2485; Background Corrected with Scan 2476

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.2	115616	PASS
68	69	0.00	2	1.6	1987	PASS
69	198	0.00	100	40.7	123197	PASS
70	69	0.00	2	0.6	719	PASS
127	198	10	80	53.3	161643	PASS
197	198	0.00	2	0.5	1586	PASS
198	198	100	100	100.0	302997	PASS
199	198	5	9	6.7	20357	PASS
275	198	10	60	26.0	78643	PASS
365	198	1	100	3.2	9799	PASS
441	198	0.01	100	12.2	36877	PASS
442	198	50	100	74.9	227008	PASS
443	442	15	24	20.3	45976	PASS

m/z	Abundance
35.90	160.0
39.00	160.0
39.90	1228.0
41.00	306.0
43.00	255.0
43.90	795.0
48.90	471.0
50.90	173.0
55.00	183.0
57.00	442.0
71.00	374.0
73.00	1631.0
73.90	239.0
74.90	212.0
76.90	252.0
83.10	151.0
83.90	348.0
84.90	259.0
86.00	223.0
96.00	256.0
97.10	161.0
133.00	339.0
147.00	777.0
167.00	188.0
190.90	265.0
193.00	289.0
207.00	2322.0
208.00	423.0
209.00	414.0
221.10	476.0
222.90	155.0
266.90	202.0
281.00	599.0
282.00	180.0
340.90	223.0
354.90	190.0

Instrument :
BNA_N
ClientSampleId :
DFTPP250

m/z	Abundance
36.00	256.0
37.00	419.0
38.00	1428.0
39.00	7765.0
39.90	1419.0
41.00	489.0
42.90	288.0
43.90	914.0
45.00	255.0
49.00	1089.0
50.00	29288.0
51.00	116960.0
52.00	6267.0
53.00	272.0
55.00	779.0
56.00	3319.0
57.00	8399.0
58.00	440.0
58.90	162.0
61.00	1558.0
62.00	1688.0
63.00	4653.0
64.00	823.0
64.90	2361.0
65.90	160.0
66.90	174.0
68.00	1859.0
69.00	124464.0
70.00	673.0
70.90	298.0
73.00	1868.0
74.00	12837.0
75.00	20592.0
76.00	6325.0
77.00	147648.0
78.00	10560.0
79.00	8711.0
80.00	7137.0
81.00	9875.0
82.00	2421.0
83.00	2492.0
83.90	503.0
84.90	1685.0
86.00	3051.0
87.00	1493.0
87.90	492.0
89.00	196.0
91.00	2217.0
91.90	2424.0
93.00	16380.0
94.00	1098.0
94.90	316.0
96.00	946.0
96.90	384.0
98.00	11686.0
99.00	11452.0
99.90	1040.0
101.00	5325.0
101.90	321.0
103.00	1655.0
103.90	3205.0
105.00	3336.0
106.00	1135.0
107.00	42560.0
108.00	6748.0
108.90	1277.0
110.00	77920.0
111.00	11953.0
111.90	1499.0
112.90	439.0
114.90	289.0
116.00	2265.0
117.00	38408.0
118.00	3086.0
118.90	455.0
119.90	630.0
120.90	290.0
122.00	2865.0
123.00	3998.0
124.00	1975.0
125.00	1644.0
127.00	153984.0
128.00	11791.0
129.00	56984.0
130.00	4356.0
130.90	968.0
131.90	515.0
132.90	407.0
134.00	1519.0
135.00	4661.0
136.00	1931.0
137.00	2708.0

Instrument :
BNA_N
ClientSampleId :
DFTPP250

137.90	516.0
138.80	296.0
139.90	699.0
141.00	6854.0
142.00	2551.0
143.00	1739.0
144.00	424.0
144.90	325.0
146.00	1431.0
147.00	4194.0
147.90	9478.0
149.00	2302.0
149.90	696.0
151.00	864.0
151.80	558.0
153.00	2396.0
154.00	1557.0
155.00	4153.0
156.00	6075.0
157.00	984.0
158.00	1138.0
159.00	1023.0
159.90	2212.0
161.00	3501.0
161.90	970.0
162.90	355.0
163.80	326.0
164.90	2638.0
166.00	2191.0
167.00	16077.0
168.00	11572.0
169.00	1589.0
169.90	493.0
171.00	635.0
172.00	1335.0
172.90	1788.0
174.00	3183.0
175.00	5510.0
176.00	1652.0
177.00	2306.0
177.90	775.0
179.00	10658.0
180.00	7687.0
181.00	3350.0
182.00	619.0
182.90	315.0
183.90	681.0
185.00	5148.0
186.00	40376.0
187.00	11652.0
188.00	1321.0
188.90	2195.0
190.00	558.0
191.00	1325.0
192.00	3489.0
193.00	3769.0
193.90	882.0
194.90	496.0
196.00	7956.0
197.00	1372.0
198.00	265728.0
199.00	19208.0
200.00	1598.0
201.50	1387.0
203.00	1854.0
204.00	9309.0
205.00	16696.0
206.00	70608.0
207.10	12176.0
208.00	2876.0
209.00	1017.0
210.00	1014.0
211.00	2831.0
212.00	574.0
212.90	236.0
214.90	714.0
216.00	1516.0
217.00	18728.0
218.00	2128.0
219.00	305.0
221.00	10634.0
223.00	3977.0
224.10	38152.0
225.00	9841.0
226.10	1130.0
227.00	16365.0
228.00	2227.0
229.00	3430.0
230.00	452.0
231.00	1423.0
232.00	269.0
232.90	273.0
233.90	897.0
235.10	1104.0
235.90	614.0
237.00	1642.0

Instrument :
BNA_N
ClientSampleId :
DFTPP250

238.00	237.0
239.00	547.0
239.90	514.0
240.90	888.0
242.00	2147.0
243.00	2164.0
244.00	29360.0
245.00	4207.0
246.00	5758.0
246.90	1092.0
247.90	222.0
249.00	1214.0
249.90	199.0
251.00	301.0
252.00	229.0
253.00	547.0
254.00	939.0
255.00	152128.0
256.00	21144.0
257.00	1788.0
258.00	7737.0
259.00	1293.0
260.00	338.0
264.00	301.0
265.00	3827.0
266.00	428.0
267.00	181.0
270.90	400.0
271.90	344.0
273.00	4363.0
273.90	11251.0
275.00	65400.0
276.00	9249.0
277.00	5447.0
278.10	1072.0
278.90	228.0
281.10	450.0
282.10	200.0
283.00	668.0
283.90	363.0
285.00	1063.0
286.10	173.0
288.90	269.0
290.00	221.0
291.00	177.0
291.90	270.0
293.00	1217.0
294.00	364.0
294.90	405.0
296.00	15761.0
297.00	2784.0
298.00	233.0
300.80	208.0
302.00	227.0
303.00	2170.0
304.00	632.0
308.00	236.0
308.90	214.0
310.00	241.0
314.00	777.0
315.00	1800.0
316.00	1323.0
316.90	176.0
321.00	581.0
322.00	323.0
323.10	5090.0
324.00	1289.0
326.00	160.0
326.90	1112.0
328.00	616.0
332.00	623.0
333.10	431.0
334.00	3799.0
335.10	1065.0
341.00	683.0
341.90	299.0
345.90	1179.0
347.00	187.0
352.00	1853.0
353.00	1019.0
354.00	1499.0
355.00	545.0
364.90	7047.0
366.00	1233.0
370.00	213.0
371.10	599.0
372.00	2973.0
373.00	706.0
373.90	158.0
383.00	901.0
384.00	342.0
390.00	462.0
391.00	273.0
392.00	177.0
402.00	1166.0
403.10	1695.0

Instrument :
BNA_N
ClientSampleId :
DFTPP250

m/z	Abundance
35.80	154.0
36.90	554.0
38.00	1516.0
39.10	9302.0
39.90	1272.0
41.00	521.0
43.00	415.0
44.00	739.0
45.00	283.0
49.00	1240.0
50.00	33416.0
51.00	133504.0
52.00	6987.0
53.00	269.0
55.00	896.0
56.00	4276.0
57.00	9219.0
58.00	369.0
58.90	181.0
61.00	1762.0
62.00	2037.0
63.00	5499.0
64.00	750.0
65.00	2903.0
65.90	211.0
67.00	199.0
68.00	2345.0
69.00	139392.0
70.00	898.0
71.10	306.0
73.00	1941.0
74.00	14899.0
75.00	23816.0
76.00	8244.0
77.00	172672.0
78.00	11951.0
79.00	10898.0
80.00	8393.0
81.00	11792.0
82.00	2900.0
83.00	2859.0
83.90	513.0
85.00	1964.0
86.00	3498.0
87.00	1529.0
88.00	581.0
89.00	302.0
91.00	2666.0
92.00	2933.0
93.00	19936.0
94.00	1234.0
94.90	330.0
96.00	944.0
97.10	416.0
98.00	14552.0
99.00	13275.0
99.90	1106.0
101.00	6321.0
101.90	406.0
103.00	2170.0
104.00	4187.0
105.00	3881.0
105.90	1367.0
107.00	52776.0
108.00	8468.0
109.00	1311.0
110.00	92136.0
111.00	14308.0
112.00	1703.0
113.00	643.0
114.80	266.0
116.00	2666.0
117.00	44304.0
117.90	3380.0
118.90	423.0
119.90	686.0
120.90	211.0
122.00	3419.0
123.00	5579.0
124.00	2676.0
125.00	2241.0
126.00	687.0
127.00	186368.0
128.00	13640.0
129.00	70904.0
130.00	5801.0
131.00	1151.0
132.00	552.0
132.90	457.0
134.00	2074.0
135.00	5940.0
136.00	2305.0

Instrument :
BNA_N
ClientSampleId :
DFTPP250

137.00	3367.0
137.80	532.0
138.90	428.0
140.00	870.0
141.00	9108.0
142.00	3029.0
142.90	2081.0
143.90	495.0
145.00	538.0
146.00	1591.0
147.00	5232.0
148.00	11526.0
148.90	2360.0
150.00	689.0
151.10	922.0
151.60	641.0
153.00	2770.0
154.00	2298.0
155.00	5248.0
156.00	7495.0
157.00	1362.0
157.90	1955.0
159.00	1362.0
160.00	2736.0
161.00	4022.0
162.00	1331.0
163.00	464.0
164.00	559.0
165.00	3233.0
166.00	2987.0
167.00	19872.0
168.00	11684.0
169.00	1660.0
170.00	470.0
171.00	941.0
172.00	1458.0
173.00	2058.0
174.00	4083.0
175.00	6845.0
176.00	2001.0
176.90	3158.0
177.90	1082.0
179.00	14197.0
180.00	9627.0
181.00	4424.0
182.00	646.0
182.90	395.0
183.90	1050.0
185.00	6550.0
186.00	50816.0
187.00	15293.0
188.00	1530.0
188.90	3063.0
190.00	479.0
191.00	1665.0
192.00	3831.0
193.00	5148.0
194.00	1067.0
195.00	606.0
196.10	10116.0
196.90	1780.0
198.00	360640.0
199.00	24296.0
200.00	2002.0
201.40	1562.0
203.00	2397.0
204.00	13275.0
205.00	22840.0
206.00	90960.0
207.00	14531.0
208.00	3545.0
209.00	1133.0
210.20	1431.0
211.00	3824.0
211.80	775.0
212.90	267.0
215.00	1174.0
215.90	2069.0
217.00	22608.0
218.00	3359.0
219.10	401.0
221.00	13833.0
223.00	4899.0
224.10	49192.0
225.00	12920.0
226.00	1538.0
227.00	22336.0
228.00	2781.0
229.00	4413.0
229.90	505.0
231.00	2023.0
232.00	364.0
232.90	328.0
233.90	1272.0
235.00	1592.0
236.00	1273.0

Instrument :
BNA_N
ClientSampleId :
DFTPP250

237.00	2086.0
238.00	372.0
239.00	852.0
239.90	717.0
241.00	1027.0
242.00	2469.0
243.00	2794.0
244.00	39232.0
245.10	5389.0
246.00	7746.0
247.00	1495.0
248.00	295.0
249.00	1470.0
249.90	290.0
250.90	370.0
252.00	341.0
253.00	800.0
254.10	1288.0
255.00	213632.0
256.00	28544.0
257.00	2295.0
258.00	11880.0
259.00	1887.0
259.90	324.0
261.00	288.0
264.00	275.0
265.00	4971.0
266.00	779.0
267.00	239.0
271.00	513.0
271.90	506.0
273.00	5307.0
274.00	15809.0
275.00	93312.0
276.00	11583.0
277.00	8624.0
277.90	1559.0
278.90	213.0
280.90	352.0
282.00	303.0
283.00	872.0
283.90	583.0
285.10	1392.0
285.90	229.0
289.10	287.0
289.90	241.0
291.00	204.0
291.90	358.0
293.00	1810.0
294.00	536.0
295.10	664.0
296.00	25816.0
297.00	3689.0
298.00	233.0
301.00	332.0
301.90	482.0
303.00	2874.0
304.00	850.0
308.00	305.0
308.80	225.0
309.90	330.0
313.00	184.0
314.00	1179.0
315.00	2243.0
316.00	1479.0
317.10	264.0
321.00	816.0
322.00	371.0
323.10	7650.0
324.00	1305.0
325.10	154.0
326.00	205.0
327.00	1586.0
328.00	905.0
328.90	198.0
331.90	746.0
332.90	767.0
334.00	5644.0
335.10	1142.0
336.00	180.0
339.90	181.0
341.00	1141.0
342.00	202.0
345.90	1810.0
347.00	306.0
350.90	176.0
352.00	2474.0
353.00	1506.0
354.00	2599.0
355.00	583.0
358.90	229.0
365.00	10201.0
365.90	1653.0
370.00	257.0
371.00	638.0
372.10	4238.0

Instrument :
BNA_N
ClientSampleId :
DFTPP250

373.00	1058.0
374.10	159.0
383.00	1271.0
383.90	425.0
389.90	603.0
391.00	334.0
392.00	371.0
401.00	318.0
402.00	1589.0
403.00	2674.0
404.00	1091.0
405.10	292.0
415.00	205.0
421.00	2314.0
422.00	2374.0
423.00	14676.0
424.00	4475.0
425.00	650.0
441.00	42360.0
442.00	246464.0
442.90	50224.0
444.00	5124.0
445.00	322.0

Instrument :
BNA_N
ClientSampleId :
DFTPP250

m/z	Abundance
37.00	422.0
38.00	1107.0
39.00	6894.0
40.00	1204.0
41.00	437.0
43.00	346.0
44.00	749.0
44.90	247.0
49.00	903.0
50.00	25688.0
51.00	96904.0
52.00	4800.0
53.00	244.0
55.00	558.0
56.00	3077.0
57.00	7269.0
57.90	343.0
61.00	1237.0
62.00	1399.0
63.00	4186.0
64.00	578.0
65.00	1916.0
66.00	166.0
67.10	296.0
68.00	1757.0
69.00	105736.0
70.00	586.0
71.00	498.0
73.00	1698.0
74.00	11335.0
75.00	17464.0
76.00	5764.0
77.00	131968.0
78.00	8820.0
79.00	7847.0
80.00	5621.0
81.00	8811.0
82.00	2138.0
83.00	2082.0
83.90	528.0
85.00	1764.0
86.00	2812.0
87.00	1268.0
87.90	438.0
88.90	235.0
91.00	1950.0
92.00	2328.0
93.00	13991.0
94.00	1136.0
95.00	334.0
96.00	955.0
97.00	445.0
98.00	11614.0
99.00	9510.0
100.00	824.0
101.00	5141.0
101.90	272.0
102.90	1585.0
103.90	3286.0
105.00	3123.0
106.00	884.0
107.00	40160.0
108.00	6269.0
108.90	1079.0
110.00	71352.0
111.00	10898.0
111.90	1316.0
113.00	499.0
114.90	372.0
116.00	2084.0
117.00	34408.0
118.00	2681.0
119.00	335.0
119.90	528.0
121.00	248.0
122.00	2496.0
123.00	4062.0
124.00	1772.0
125.00	1822.0
126.00	485.0
127.00	144576.0
128.00	10392.0
129.00	53720.0
130.00	4624.0
130.90	842.0
132.00	532.0
132.90	396.0
133.90	1763.0
135.00	4351.0
136.00	1690.0
137.00	2547.0
138.00	474.0

Instrument :
BNA_N
ClientSampleId :
DFTPP250

138.90	239.0
139.90	660.0
140.90	7019.0
142.00	2186.0
142.90	1693.0
143.90	399.0
145.00	552.0
146.00	1281.0
147.00	4417.0
148.00	8796.0
149.00	1832.0
150.00	529.0
151.00	861.0
151.60	565.0
152.90	2420.0
154.00	1728.0
155.00	3911.0
156.00	5862.0
157.00	1241.0
157.90	1278.0
158.90	865.0
160.00	2401.0
161.00	3305.0
161.90	979.0
163.00	335.0
163.90	345.0
165.00	2601.0
166.00	2243.0
167.00	16193.0
168.00	9034.0
169.00	1486.0
170.10	520.0
171.00	485.0
172.00	1399.0
173.00	1621.0
174.00	3097.0
175.00	5728.0
176.00	1509.0
176.90	2678.0
177.90	1024.0
179.00	10912.0
180.00	7990.0
181.00	3150.0
182.00	553.0
183.00	328.0
184.00	1033.0
185.00	4530.0
186.00	39944.0
187.00	11884.0
188.00	1262.0
189.00	2673.0
190.00	326.0
190.90	1317.0
192.00	3750.0
193.00	3718.0
194.00	995.0
195.00	578.0
196.00	8313.0
196.80	1605.0
197.90	282624.0
199.00	17568.0
200.00	1682.0
201.50	1253.0
203.00	2091.0
204.00	10803.0
205.00	17280.0
206.10	75784.0
207.00	11962.0
208.00	2980.0
209.00	888.0
210.20	1269.0
211.00	2977.0
213.00	218.0
215.00	908.0
216.00	1819.0
217.00	20280.0
217.90	2679.0
218.90	284.0
221.10	12347.0
221.80	2829.0
223.00	4596.0
224.00	40792.0
225.00	11471.0
226.00	1356.0
227.00	17616.0
228.00	2256.0
229.00	3677.0
230.00	415.0
231.00	1722.0
232.00	306.0
232.90	265.0
234.00	999.0
235.00	1511.0
236.00	869.0
237.00	1461.0
238.10	239.0

Instrument :
BNA_N
ClientSampleId :
DFTPP250

239.00	610.0
239.90	656.0
241.00	885.0
242.00	2097.0
243.00	2499.0
244.00	35968.0
245.00	4229.0
246.00	6792.0
247.00	1339.0
247.90	354.0
249.00	1044.0
250.00	236.0
250.90	348.0
252.00	352.0
253.00	775.0
254.00	1109.0
255.00	176128.0
256.00	25152.0
257.00	2109.0
258.00	10798.0
259.00	1519.0
259.90	263.0
260.90	269.0
264.10	334.0
265.00	3940.0
266.00	497.0
270.90	447.0
272.00	466.0
273.00	5098.0
274.00	13944.0
275.00	77216.0
276.00	10930.0
277.00	6725.0
278.00	1083.0
279.00	268.0
281.00	465.0
281.90	270.0
283.00	786.0
284.00	475.0
285.00	1292.0
285.90	175.0
289.00	198.0
290.00	282.0
290.90	262.0
292.00	370.0
293.00	1479.0
294.00	383.0
295.00	398.0
296.00	21680.0
296.90	3564.0
298.00	217.0
301.00	365.0
302.00	333.0
303.00	2737.0
304.00	896.0
307.90	322.0
308.90	287.0
310.10	313.0
313.00	280.0
314.00	1128.0
314.90	3033.0
316.00	1270.0
316.90	350.0
320.90	977.0
322.00	560.0
323.00	7556.0
324.00	1579.0
325.90	250.0
327.00	1440.0
328.00	758.0
332.00	814.0
333.10	748.0
334.00	5396.0
334.90	1118.0
336.00	177.0
340.90	871.0
342.00	297.0
346.00	1855.0
346.90	402.0
352.00	2559.0
352.90	1572.0
354.00	2757.0
355.10	702.0
359.00	190.0
365.00	12150.0
365.90	1517.0
370.00	228.0
371.00	691.0
372.00	4221.0
373.00	1112.0
383.00	1169.0
384.00	384.0
389.90	502.0
391.00	423.0
392.00	198.0
401.00	255.0

Instrument :
BNA_N
ClientSampleId :
DFTPP250

402.00	1705.0
403.00	2307.0
403.90	1025.0
405.10	205.0
421.00	2111.0
422.10	1971.0
423.00	14619.0
424.00	3574.0
425.00	460.0
441.00	42816.0
442.00	281536.0
443.00	55448.0
444.10	4480.0
445.10	362.0

Instrument :
BNA_N
ClientSampleId :
DFTPP250

404.00	722.0
405.00	194.0
421.00	1245.0
422.10	1152.0
422.90	7326.0
424.00	3154.0
425.00	497.0
441.00	25456.0
442.00	153024.0
443.10	32256.0
444.00	2851.0
444.90	167.0

Instrument :
BNA_N
ClientSampleId :
DFTPP250