

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN090819\
 Data File : BN007559.D
 Acq On : 08 Sep 2019 20:24
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 09 04:46:46 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 29 12:18:45 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	312#	-0.02
2	1,4-Dioxane	0.665	0.658	1.1	297#	0.00
3	n-Nitrosodimethylamine	0.309	0.234	24.3	228#	0.00
4 S	2-Fluorophenol	1.490	1.171	21.4	258#	-0.02
5 S	Phenol-d6	1.953	1.470	24.7	269#	0.00
6	bis(2-Chloroethyl)ether	1.445	1.530	-5.9	333#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	310#	-0.03
8 S	Nitrobenzene-d5	0.348	0.323	7.2	301#	-0.01
9	Naphthalene	1.143	1.133	0.9	300#	-0.01
10	Hexachlorobutadiene	0.234	0.203	13.2	259#	-0.01
11 SURR	2-Methylnaphthalene-d10	0.665	0.611	8.1	281#	-0.01
12	2-Methylnaphthalene	0.779	0.734	5.8	284#	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	302#	-0.02
14 S	2,4,6-Tribromophenol	0.209	0.148	29.2#	218#	-0.02
15 S	2-Fluorobiphenyl	1.613	1.578	2.2	289#	0.00
16	Acenaphthylene	2.367	2.033	14.1	256#	-0.02
17	Acenaphthene	1.385	1.351	2.5	288#	0.00
18	Fluorene	1.751	1.661	5.1	280#	-0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	287#	-0.02
20	4-Bromophenyl-phenylether	0.158	0.056	64.6#	118	0.00
21	Hexachlorobenzene	0.268	0.258	3.7	273#	-0.02
22	Pentachlorophenol	0.083	0.049	41.0#	178#	-0.02
23	Phenanthrene	1.203	1.222	-1.6	291#	-0.02
24	Anthracene	1.210	1.066	11.9	251#	0.00
25 SURR	Fluoranthene-d10	1.288	1.134	12.0	244#	-0.02
26	Fluoranthene	1.545	1.394	9.8	252#	-0.02
27 I	Chrysene-d12	1.000	1.000	0.0	276#	-0.01
28	Pvrene	1.609	1.489	7.5	249#	-0.01
29 S	Terphenyl-d14	1.047	0.943	9.9	240#	-0.02
30	Benzo(a)anthracene	1.571	1.343	14.5	232#	-0.01
31	Chrysene	1.516	1.473	2.8	257#	-0.01
32	Bis(2-ethylhexyl)phthalate	1.177	0.633	46.2#	157#	-0.02
33	Indeno(1,2,3-cd)pyrene	1.827	1.651	9.6	240#	-0.04
34 I	Perylene-d12	1.000	1.000	0.0	266#	-0.02
35	Benzo(b)fluoranthene	1.449	1.354	6.6	243#	-0.02
36	Benzo(k)fluoranthene	1.432	1.525	-6.5	272#	-0.02
37 C	Benzo(a)pyrene	1.379	1.266	8.2	237#	-0.03
38	Dibenzo(a,h)anthracene	1.376	1.251	9.1	234#	-0.04
39	Benzo(a,h,i)perylene	1.371	1.253	8.6	236#	-0.04

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN090819\
Data File : BN007559.D
Acq On : 08 Sep 2019 20:24
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 09 04:46:46 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 29 12:18:45 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			