

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090822\
 Data File : BN021638.D
 Acq On : 08 Sep 2022 10:23
 Operator : CG/JU
 Sample : SST0.140
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST0.1244

Quant Time: Sep 08 13:40:29 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 08 13:38:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.067	152	6438	0.400	ng/ul	0.00
4) Naphthalene-d8	10.894	136	20555	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.721	164	11801	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.472	188	24573	0.400	ng/ul	0.00
17) Chrysene-d12	21.659	240	17055	0.400	ng/ul	0.00
23) Perylene-d12	24.173	264	12583	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	0.000	96	0d	0.000	ng/ul	
6) 2-Methylnaphthalene-d10	12.483	152	3208	0.103	ng/ul	0.00
18) Fluoranthene-d10	19.497	212	6178	0.118	ng/ul	0.00
Target Compounds						
					Qvalue	
5) Naphthalene	10.944	128	5900	0.091	ng/ul	99
7) 2-Methylnaphthalene	12.555	142	3796	0.093	ng/ul	99
8) 1-Methylnaphthalene	12.775	142	3844	0.101	ng/ul	99
10) Acenaphthylene	14.444	152	4995	0.082	ng/ul	100
11) Acenaphthene	14.781	153	4256	0.091	ng/ul	98
12) Fluorene	15.771	166	4693	0.090	ng/ul	99
15) Phenanthrene	17.515	178	7771	0.085	ng/ul	99
16) Anthracene	17.607	178	6272	0.086	ng/ul	99
19) Fluoranthene	19.529	202	7662	0.095	ng/ul	98
20) Pyrene	19.892	202	7802	0.096	ng/ul	95
21) Benzo(a)anthracene	21.644	228	5643	0.095	ng/ul	100
22) Chrysene	21.697	228	6301	0.083	ng/ul	99
24) Benzo(b)fluoranthene	23.398	252	5740	0.106	ng/ul	87
25) Benzo(k)fluoranthene	23.451	252	5560	0.096	ng/ul#	87
26) Benzo(a)pyrene	24.062	252	4500	0.088	ng/ul#	78
27) Indeno(1,2,3-cd)pyrene	26.827	276	5047	0.084	ng/ul#	92
28) Dibenzo(a,h)anthracene	26.854	278	4217	0.086	ng/ul#	93
29) Benzo(g,h,i)perylene	27.645	276	4666	0.079	ng/ul	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090822\
Data File : BN021638.D
Acq On : 08 Sep 2022 10:23
Operator : CG/JU
Sample : SST0.140
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.1244

Quant Time: Sep 08 13:40:29 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 08 13:38:06 2022
Response via : Initial Calibration

