

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN090920\
 Data File : BN011762.D
 Acq On : 09 Sep 2020 09:54
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 09 10:55:57 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082720.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 09 10:54:46 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
2	1,4-Dioxane	0.643	0.591	8.1	105	0.00
3	n-Nitrosodimethylamine	0.743	0.668	10.1	94	0.00
4 S	2-Fluorophenol	1.130	1.030	8.8	101	0.00
5 S	Phenol-d6	1.535	1.357	11.6	100	0.00
6	bis(2-Chloroethyl)ether	1.249	1.193	4.5	107	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	105	0.00
8 S	Nitrobenzene-d5	0.387	0.362	6.5	108	0.00
9	Naphthalene	1.144	1.042	8.9	104	0.00
10	Hexachlorobutadiene	0.251	0.230	8.4	106	0.00
11 SURR	2-Methylnaphthalene-d10	0.721	0.647	10.3	105	0.00
12	2-Methylnaphthalene	0.814	0.734	9.8	105	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	105	0.00
14 S	2,4,6-Tribromophenol	0.194	0.163	16.0	100	0.00
15 S	2-Fluorobiphenyl	1.691	1.554	8.1	106	0.00
16	Acenaphthylene	1.948	1.753	10.0	106	0.00
17	Acenaphthene	1.345	1.228	8.7	107	0.00
18	Fluorene	1.724	1.566	9.2	106	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	106	0.00
20	4,6-Dinitro-2-methylphenol	0.062	0.054	12.9	108	0.00
21	4-Bromophenyl-phenylether	0.234	0.213	9.0	107	0.00
22	Hexachlorobenzene	0.289	0.275	4.8	111	0.00
23	Atrazine	0.206	0.123	40.3#	70	0.00
24	Pentachlorophenol	0.115	0.090	21.7	91	0.00
25	Phenanthrene	1.292	1.163	10.0	107	0.00
26	Anthracene	1.148	1.005	12.5	103	0.00
27 SURR	Fluoranthene-d10	1.284	1.059	17.5	97	0.00
28	Fluoranthene	1.532	1.288	15.9	100	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	96	0.00
30	Pyrene	1.391	1.332	4.2	98	0.00
31 S	Terphenyl-d14	0.993	0.942	5.1	97	0.00
32	Benzo(a)anthracene	1.428	1.230	13.9	90	0.00
33	Chrysene	1.473	1.315	10.7	91	0.00
34	Bis(2-ethylhexyl)phthalate	0.581	0.470	19.1	86	0.00
35	Indeno(1,2,3-cd)pyrene	1.921	1.521	20.8	79	0.00
36 I	Perylene-d12	1.000	1.000	0.0	81	0.00
37	Benzo(b)fluoranthene	1.534	1.449	5.5	90	0.00
38	Benzo(k)fluoranthene	1.568	1.414	9.8	82	0.00
39 C	Benzo(a)pyrene	1.371	1.217	11.2	84	0.00
40	Dibenzo(a,h)anthracene	1.548	1.340	13.4	79	0.00
41	Benzo(g,h,i)perylene	1.516	1.382	8.8	85	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN090920\
Data File : BN011762.D
Acq On : 09 Sep 2020 09:54
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 09 10:55:57 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082720.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 09 10:54:46 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0