

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN090920\
 Data File : BN011762.D
 Acq On : 09 Sep 2020 09:54
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 09 10:55:57 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082720.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 09 10:54:46 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	104	0.00
2	1,4-Dioxane	0.400	0.368	8.0	105	0.00
3	n-Nitrosodimethylamine	0.400	0.360	10.0	94	0.00
4 S	2-Fluorophenol	0.400	0.365	8.8	101	0.00
5 S	Phenol-d6	0.400	0.353	11.8	100	0.00
6	bis(2-Chloroethyl)ether	0.400	0.382	4.5	107	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	105	0.00
8 S	Nitrobenzene-d5	0.400	0.373	6.8	108	0.00
9	Naphthalene	0.400	0.364	9.0	104	0.00
10	Hexachlorobutadiene	0.400	0.367	8.3	106	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.359	10.3	105	0.00
12	2-Methylnaphthalene	0.400	0.361	9.8	105	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	105	0.00
14 S	2,4,6-Tribromophenol	0.400	0.336	16.0	100	0.00
15 S	2-Fluorobiphenyl	0.400	0.368	8.0	106	0.00
16	Acenaphthylene	0.400	0.360	10.0	106	0.00
17	Acenaphthene	0.400	0.365	8.8	107	0.00
18	Fluorene	0.400	0.363	9.3	106	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	106	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.349	12.8	108	0.00
21	4-Bromophenyl-phenylether	0.400	0.364	9.0	107	0.00
22	Hexachlorobenzene	0.400	0.380	5.0	111	0.00
23	Atrazine	0.400	0.239	40.3#	70	0.00
24	Pentachlorophenol	0.400	0.313	21.8	91	0.00
25	Phenanthrene	0.400	0.360	10.0	107	0.00
26	Anthracene	0.400	0.350	12.5	103	0.00
27 SURR	Fluoranthene-d10	0.400	0.330	17.5	97	0.00
28	Fluoranthene	0.400	0.336	16.0	100	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	96	0.00
30	Pyrene	0.400	0.383	4.3	98	0.00
31 S	Terphenyl-d14	0.400	0.380	5.0	97	0.00
32	Benzo(a)anthracene	0.400	0.345	13.8	90	0.00
33	Chrysene	0.400	0.357	10.8	91	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.323	19.3	86	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.317	20.8	79	0.00
36 I	Perylene-d12	0.400	0.400	0.0	81	0.00
37	Benzo(b)fluoranthene	0.400	0.378	5.5	90	0.00
38	Benzo(k)fluoranthene	0.400	0.361	9.8	82	0.00
39 C	Benzo(a)pyrene	0.400	0.355	11.3	84	0.00
40	Dibenzo(a,h)anthracene	0.400	0.346	13.5	79	0.00
41	Benzo(g,h,i)perylene	0.400	0.365	8.8	85	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN090920\
Data File : BN011762.D
Acq On : 09 Sep 2020 09:54
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 09 10:55:57 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082720.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 09 10:54:46 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0