

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN090920\
 Data File : BN011775.D
 Acq On : 09 Sep 2020 19:33
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 09 20:03:21 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082720.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 09 10:54:46 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
2	1,4-Dioxane	0.643	0.596	7.3	90	0.00
3	n-Nitrosodimethylamine	0.743	0.764	-2.8	91	0.00
4 S	2-Fluorophenol	1.130	1.074	5.0	89	0.00
5 S	Phenol-d6	1.535	1.418	7.6	89	0.00
6	bis(2-Chloroethyl)ether	1.249	1.176	5.8	90	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	88	0.00
8 S	Nitrobenzene-d5	0.387	0.377	2.6	94	0.00
9	Naphthalene	1.144	1.049	8.3	88	0.00
10	Hexachlorobutadiene	0.251	0.231	8.0	89	0.00
11 SURR	2-Methylnaphthalene-d10	0.721	0.643	10.8	87	0.00
12	2-Methylnaphthalene	0.814	0.734	9.8	88	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	87	0.00
14 S	2,4,6-Tribromophenol	0.194	0.176	9.3	90	-0.01
15 S	2-Fluorobiphenyl	1.691	1.573	7.0	89	0.00
16	Acenaphthylene	1.948	1.735	10.9	88	-0.01
17	Acenaphthene	1.345	1.226	8.8	89	0.00
18	Fluorene	1.724	1.543	10.5	87	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	86	0.00
20	4,6-Dinitro-2-methylphenol	0.062	0.060	3.2	97	0.00
21	4-Bromophenyl-phenylether	0.234	0.213	9.0	86	0.00
22	Hexachlorobenzene	0.289	0.273	5.5	89	0.00
23	Atrazine	0.206	0.126	38.8#	59	0.00
24	Pentachlorophenol	0.115	0.097	15.7	79	0.00
25	Phenanthrene	1.292	1.190	7.9	88	0.00
26	Anthracene	1.148	1.013	11.8	85	0.00
27 SURR	Fluoranthene-d10	1.284	1.087	15.3	81	0.00
28	Fluoranthene	1.532	1.324	13.6	83	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	81	0.00
30	Pyrene	1.391	1.318	5.2	82	0.00
31 S	Terphenyl-d14	0.993	0.920	7.4	80	0.00
32	Benzo(a)anthracene	1.428	1.272	10.9	79	0.00
33	Chrysene	1.473	1.334	9.4	78	0.00
34	Bis(2-ethylhexyl)phthalate	0.581	0.500	13.9	77	-0.01
35	Indeno(1,2,3-cd)pyrene	1.921	1.604	16.5	71	0.00
36 I	Perylene-d12	1.000	1.000	0.0	72	0.00
37	Benzo(b)fluoranthene	1.534	1.418	7.6	78	0.00
38	Benzo(k)fluoranthene	1.568	1.420	9.4	73	0.00
39 C	Benzo(a)pyrene	1.371	1.240	9.6	76	0.00
40	Dibenzo(a,h)anthracene	1.548	1.346	13.0	71	-0.01
41	Benzo(g,h,i)perylene	1.516	1.353	10.8	73	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN090920\
Data File : BN011775.D
Acq On : 09 Sep 2020 19:33
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 09 20:03:21 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082720.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 09 10:54:46 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0