

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN090920\
 Data File : BN011781.D
 Acq On : 10 Sep 2020 04:48
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 10 05:19:30 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082720.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 09 10:54:46 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
2	1,4-Dioxane	0.643	0.618	3.9	101	0.00
3	n-Nitrosodimethylamine	0.743	0.782	-5.2	101	0.00
4 S	2-Fluorophenol	1.130	1.018	9.9	91	0.00
5 S	Phenol-d6	1.535	1.345	12.4	91	0.00
6	bis(2-Chloroethyl)ether	1.249	1.129	9.6	93	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	92	0.00
8 S	Nitrobenzene-d5	0.387	0.387	0.0	100	0.00
9	Naphthalene	1.144	1.052	8.0	92	-0.01
10	Hexachlorobutadiene	0.251	0.246	2.0	98	0.00
11 SURR	2-Methylnaphthalene-d10	0.721	0.666	7.6	94	0.00
12	2-Methylnaphthalene	0.814	0.755	7.2	94	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	94	0.00
14 S	2,4,6-Tribromophenol	0.194	0.174	10.3	95	-0.01
15 S	2-Fluorobiphenyl	1.691	1.590	6.0	97	0.00
16	Acenaphthylene	1.948	1.739	10.7	95	-0.01
17	Acenaphthene	1.345	1.245	7.4	97	0.00
18	Fluorene	1.724	1.590	7.8	97	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
20	4,6-Dinitro-2-methylphenol	0.062	0.063	-1.6	112	-0.01
21	4-Bromophenyl-phenylether	0.234	0.219	6.4	98	0.00
22	Hexachlorobenzene	0.289	0.274	5.2	99	0.00
23	Atrazine	0.206	0.128	37.9#	66	0.00
24	Pentachlorophenol	0.115	0.094	18.3	85	0.00
25	Phenanthrene	1.292	1.168	9.6	96	0.00
26	Anthracene	1.148	1.015	11.6	93	0.00
27 SURR	Fluoranthene-d10	1.284	1.066	17.0	88	0.00
28	Fluoranthene	1.532	1.297	15.3	90	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	86	-0.01
30	Pyrene	1.391	1.330	4.4	88	0.00
31 S	Terphenyl-d14	0.993	0.945	4.8	87	0.00
32	Benzo(a)anthracene	1.428	1.243	13.0	81	0.00
33	Chrysene	1.473	1.333	9.5	83	-0.01
34	Bis(2-ethylhexyl)phthalate	0.581	0.531	8.6	87	-0.01
35	Indeno(1,2,3-cd)pyrene	1.921	1.617	15.8	75	-0.02
36 I	Perylene-d12	1.000	1.000	0.0	75	-0.01
37	Benzo(b)fluoranthene	1.534	1.407	8.3	81	-0.01
38	Benzo(k)fluoranthene	1.568	1.428	8.9	77	-0.01
39 C	Benzo(a)pyrene	1.371	1.246	9.1	80	-0.01
40	Dibenzo(a,h)anthracene	1.548	1.371	11.4	75	-0.01
41	Benzo(g,h,i)perylene	1.516	1.379	9.0	78	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN090920\
Data File : BN011781.D
Acq On : 10 Sep 2020 04:48
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 10 05:19:30 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082720.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 09 10:54:46 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0