

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090924\
 Data File : BN033824.D
 Acq On : 09 Sep 2024 19:38
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Sep 10 03:18:15 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN090424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 04 18:36:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.509	152	6030	0.400	ng	0.00
7) Naphthalene-d8	10.261	136	15254	0.400	ng	-0.01
13) Acenaphthene-d10	14.146	164	7009	0.400	ng	-0.01
19) Phenanthrene-d10	16.892	188	13494	0.400	ng	#-0.01
29) Chrysene-d12	21.103	240	10486	0.400	ng	0.00
35) Perylene-d12	23.259	264	10868	0.400	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.147	112	6087	0.330	ng	0.00
5) Phenol-d6	6.700	99	7136	0.327	ng	0.00
8) Nitrobenzene-d5	8.638	82	4885	0.376	ng	-0.01
11) 2-Methylnaphthalene-d10	11.862	152	8017	0.372	ng	-0.01
14) 2,4,6-Tribromophenol	15.651	330	1142	0.323	ng	0.00
15) 2-Fluorobiphenyl	12.756	172	11454	0.392	ng	-0.01
27) Fluoranthene-d10	18.938	212	12457	0.374	ng	0.00
31) Terphenyl-d14	19.555	244	7979	0.357	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.168	88	2623	0.382	ng	98
3) n-Nitrosodimethylamine	3.457	42	2967	0.368	ng	99
6) bis(2-Chloroethyl)ether	6.945	93	6510	0.395	ng	100
9) Naphthalene	10.314	128	16177	0.397	ng	99
10) Hexachlorobutadiene	10.613	225	3342	0.394	ng	# 99
12) 2-Methylnaphthalene	11.937	142	9734	0.390	ng	99
16) Acenaphthylene	13.857	152	11504	0.380	ng	100
17) Acenaphthene	14.210	154	8218	0.396	ng	100
18) Fluorene	15.204	166	9982	0.390	ng	99
20) 4,6-Dinitro-2-methylph...	15.290	198	567	0.266	ng	# 77
21) 4-Bromophenyl-phenylether	16.098	248	3100	0.397	ng	# 80
22) Hexachlorobenzene	16.209	284	3714	0.414	ng	99
23) Atrazine	16.383	200	2276	0.365	ng	# 92
24) Pentachlorophenol	16.557	266	1221	0.333	ng	98
25) Phenanthrene	16.942	178	14587	0.392	ng	100
26) Anthracene	17.028	178	12428	0.384	ng	99
28) Fluoranthene	18.970	202	16258	0.390	ng	100
30) Pyrene	19.332	202	16697	0.395	ng	100
32) Benzo(a)anthracene	21.094	228	14651	0.391	ng	99
33) Chrysene	21.139	228	14975	0.404	ng	100
34) Bis(2-ethylhexyl)phtha...	21.050	149	7882	0.386	ng	99
36) Indeno(1,2,3-cd)pyrene	25.390	276	17791	0.396	ng	99
37) Benzo(b)fluoranthene	22.621	252	15730	0.394	ng	98
38) Benzo(k)fluoranthene	22.662	252	15789	0.405	ng	100
39) Benzo(a)pyrene	23.162	252	13201	0.395	ng	97
40) Dibenzo(a,h)anthracene	25.405	278	14176	0.394	ng	98
41) Benzo(g,h,i)perylene	26.033	276	15291	0.392	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090924\
Data File : BN033824.D
Acq On : 09 Sep 2024 19:38
Operator : MA/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Quant Time: Sep 10 03:18:15 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN090424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 04 18:36:01 2024
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.509 min Scan# 61
 Delta R.T. -0.007 min
 Lab File: BN033824.D
 Acq: 09 Sep 2024 19:38

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC



Tgt Ion:152 Resp: 6030




Ion	Ratio	Lower	Upper
152	100		
150	152.6	123.3	184.9
115	62.5	51.0	76.4




#2
 1,4-Dioxane
 Concen: 0.382 ng
 RT: 3.168 min Scan# 26
 Delta R.T. 0.000 min
 Lab File: BN033824.D
 Acq: 09 Sep 2024 19:38

Tgt Ion: 88 Resp: 2623

Ion	Ratio	Lower	Upper
88	100		
43	32.8	29.0	43.4
58	80.7	64.1	96.1









#3
 n-Nitrosodimethylamine
 Concen: 0.368 ng
 RT: 3.457 min Scan# 60
 Delta R.T. -0.007 min
 Lab File: BN033824.D
 Acq: 09 Sep 2024 19:38

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion: 42 Resp: 2967




Ion	Ratio	Lower	Upper
42	100		
74	118.0	93.5	140.3
44	7.9	7.4	11.0




#4
 2-Fluorophenol
 Concen: 0.330 ng
 RT: 5.147 min Scan# 300
 Delta R.T. -0.007 min
 Lab File: BN033824.D
 Acq: 09 Sep 2024 19:38

Tgt Ion: 112 Resp: 6087

Ion	Ratio	Lower	Upper
112	100		
64	60.8	47.6	71.4
63	32.1	25.9	38.9









#5
Phenol-d6
Concen: 0.327 ng
RT: 6.700 min Scan# 516
Delta R.T. -0.007 min
Lab File: BN033824.D
Acq: 09 Sep 2024 19:38




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 99 Resp: 7136
Ion Ratio Lower Upper
99 100
42 21.1 18.0 27.0
71 35.4 28.7 43.1



#6
bis(2-Chloroethyl)ether
Concen: 0.395 ng
RT: 6.945 min Scan# 549
Delta R.T. -0.015 min
Lab File: BN033824.D
Acq: 09 Sep 2024 19:38

Tgt Ion: 93 Resp: 6510
Ion Ratio Lower Upper
93 100
63 80.8 64.4 96.6
95 33.4 26.5 39.7









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.261 min Scan# 91
Delta R.T. -0.011 min
Lab File: BN033824.D
Acq: 09 Sep 2024 19:38




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	136	Resp:	15254
Ion	Ratio	Lower	Upper
136	100		
137	11.4	9.0	13.4
54	10.7	8.1	12.1
68	7.3	5.8	8.8



#8
Nitrobenzene-d5
Concen: 0.376 ng
RT: 8.638 min Scan# 769
Delta R.T. -0.011 min
Lab File: BN033824.D
Acq: 09 Sep 2024 19:38

Tgt Ion:	82	Resp:	4885
Ion	Ratio	Lower	Upper
82	100		
128	37.7	31.3	46.9
54	61.2	46.9	70.3









#9
Naphthalene
Concen: 0.397 ng
RT: 10.314 min Scan# 91
Delta R.T. -0.011 min
Lab File: BN033824.D
Acq: 09 Sep 2024 19:38




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	128	Resp:	16177
Ion Ratio	Lower	Upper	
128	100		
129	11.5	8.9	13.3
127	13.6	10.8	16.2



#10
Hexachlorobutadiene
Concen: 0.394 ng
RT: 10.613 min Scan# 954
Delta R.T. -0.011 min
Lab File: BN033824.D
Acq: 09 Sep 2024 19:38

Tgt Ion:	225	Resp:	3342
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	64.9	51.2	76.8









#11
2-Methylnaphthalene-d10
Concen: 0.372 ng
RT: 11.862 min Scan# 1131
Delta R.T. -0.012 min
Lab File: BN033824.D
Acq: 09 Sep 2024 19:38




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:152 Resp: 8017
Ion Ratio Lower Upper
152 100
151 20.7 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.390 ng
RT: 11.937 min Scan# 1151
Delta R.T. -0.012 min
Lab File: BN033824.D
Acq: 09 Sep 2024 19:38

Tgt Ion:142 Resp: 9734
Ion Ratio Lower Upper
142 100
141 89.5 70.9 106.3
115 39.2 31.0 46.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.146 min Scan# 1473
Delta R.T. -0.011 min
Lab File: BN033824.D
Acq: 09 Sep 2024 19:38




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:164 Resp: 7009
Ion Ratio Lower Upper
164 100
162 103.8 82.0 123.0
160 49.9 39.1 58.7



#14
2,4,6-Tribromophenol
Concen: 0.323 ng
RT: 15.651 min Scan# 1611
Delta R.T. -0.000 min
Lab File: BN033824.D
Acq: 09 Sep 2024 19:38

Tgt Ion:330 Resp: 1142
Ion Ratio Lower Upper
330 100
332 94.2 74.2 111.4
141 43.4 33.9 50.9









#15
2-Fluorobiphenyl
Concen: 0.392 ng
RT: 12.756 min Scan# 11
Delta R.T. -0.011 min
Lab File: BN033824.D
Acq: 09 Sep 2024 19:38




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	172	Resp:	11454
Ion Ratio	Lower	Upper	
172	100		
171	36.1	28.8	43.2
170	24.8	19.7	29.5



#16
Acenaphthylene
Concen: 0.380 ng
RT: 13.857 min Scan# 1445
Delta R.T. -0.011 min
Lab File: BN033824.D
Acq: 09 Sep 2024 19:38

Tgt Ion:	152	Resp:	11504
Ion Ratio	Lower	Upper	
152	100		
151	19.6	15.6	23.4
153	13.1	10.3	15.5





#17

Acenaphthene

Concen: 0.396 ng

RT: 14.210 min Scan# 1478

Delta R.T. -0.000 min

Lab File: BN033824.D

Acq: 09 Sep 2024 19:38

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:154 Resp: 8218

Ion Ratio Lower Upper

154 100

153 114.9 91.4 137.2

152 57.2 45.8 68.6



#18

Fluorene

Concen: 0.390 ng

RT: 15.204 min Scan# 1571

Delta R.T. -0.000 min

Lab File: BN033824.D

Acq: 09 Sep 2024 19:38




Tgt Ion:166 Resp: 9982


Ion Ratio Lower Upper

166 100

165 99.4 78.8 118.2

167 13.7 10.6 16.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 16.892 min Scan# 1711

Delta R.T. -0.013 min

Lab File: BN033824.D

Acq: 09 Sep 2024 19:38

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:188 Resp: 13494

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 14.7 7.9 11.9#



#20

4,6-Dinitro-2-methylphenol

Concen: 0.266 ng

RT: 15.290 min Scan# 1579

Delta R.T. -0.000 min

Lab File: BN033824.D

Acq: 09 Sep 2024 19:38




Tgt Ion:198 Resp: 567


Ion Ratio Lower Upper

198 100

51 91.9 92.8 139.2#

105 54.0 59.4 89.0#









#21
4-Bromophenyl-phenylether
Concen: 0.397 ng
RT: 16.098 min Scan# 1647
Delta R.T. -0.013 min
Lab File: BN033824.D
Acq: 09 Sep 2024 19:38




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:248 Resp: 3100
Ion Ratio Lower Upper
248 100
250 91.3 79.5 119.3
141 85.3 44.9 67.3#



#22
Hexachlorobenzene
Concen: 0.414 ng
RT: 16.209 min Scan# 1656
Delta R.T. -0.013 min
Lab File: BN033824.D
Acq: 09 Sep 2024 19:38

Tgt Ion:284 Resp: 3714
Ion Ratio Lower Upper
284 100
142 40.2 32.8 49.2
249 32.6 26.2 39.2





#23

Atrazine

Concen: 0.365 ng

RT: 16.383 min Scan# 1670

Delta R.T. -0.013 min

Lab File: BN033824.D

Acq: 09 Sep 2024 19:38

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:200 Resp: 2276

Ion Ratio Lower Upper

200 100

173 30.4 18.2 27.4#

215 46.0 39.0 58.4



#24

Pentachlorophenol

Concen: 0.333 ng

RT: 16.557 min Scan# 1684

Delta R.T. -0.013 min

Lab File: BN033824.D

Acq: 09 Sep 2024 19:38




Tgt Ion:266 Resp: 1221


Ion Ratio Lower Upper

266 100

264 63.1 49.6 74.4

268 63.0 52.2 78.4









#25
Phenanthrene
Concen: 0.392 ng
RT: 16.942 min Scan# 1715
Delta R.T. -0.000 min
Lab File: BN033824.D
Acq: 09 Sep 2024 19:38




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:178 Resp: 14587
Ion Ratio Lower Upper
178 100
176 19.1 15.4 23.0
179 15.4 12.2 18.4



#26
Anthracene
Concen: 0.384 ng
RT: 17.028 min Scan# 1722
Delta R.T. -0.013 min
Lab File: BN033824.D
Acq: 09 Sep 2024 19:38

Tgt Ion:178 Resp: 12428
Ion Ratio Lower Upper
178 100
176 18.9 14.8 22.2
179 15.2 12.4 18.6









#27
Fluoranthene-d10
Concen: 0.374 ng
RT: 18.938 min Scan# 1934
Delta R.T. -0.009 min
Lab File: BN033824.D
Acq: 09 Sep 2024 19:38




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:212 Resp: 12457
Ion Ratio Lower Upper
212 100
106 14.1 11.5 17.3
104 8.2 6.6 9.8



#28
Fluoranthene
Concen: 0.390 ng
RT: 18.970 min Scan# 1941
Delta R.T. -0.005 min
Lab File: BN033824.D
Acq: 09 Sep 2024 19:38

Tgt Ion:202 Resp: 16258
Ion Ratio Lower Upper
202 100
101 11.7 9.2 13.8
203 17.3 13.8 20.6





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.103 min Scan# 21

Delta R.T. -0.009 min

Lab File: BN033824.D

Acq: 09 Sep 2024 19:38

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:240 Resp: 10486

Ion Ratio Lower Upper

240 100

120 20.1 13.7 20.5

236 29.1 23.4 35.2



#30

Pyrene

Concen: 0.395 ng

RT: 19.332 min Scan# 2019

Delta R.T. -0.010 min

Lab File: BN033824.D

Acq: 09 Sep 2024 19:38




Tgt Ion:202 Resp: 16697


Ion Ratio Lower Upper

202 100

200 21.0 16.8 25.2

203 18.1 14.4 21.6









#31
Terphenyl-d14
Concen: 0.357 ng
RT: 19.555 min Scan# 2067
Delta R.T. -0.005 min
Lab File: BN033824.D
Acq: 09 Sep 2024 19:38




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:244 Resp: 7979
Ion Ratio Lower Upper
244 100
212 9.9 7.9 11.9
122 15.8 13.2 19.8



#32
Benzo(a)anthracene
Concen: 0.391 ng
RT: 21.094 min Scan# 2389
Delta R.T. -0.009 min
Lab File: BN033824.D
Acq: 09 Sep 2024 19:38

Tgt Ion:228 Resp: 14651
Ion Ratio Lower Upper
228 100
226 26.8 21.2 31.8
229 20.5 15.9 23.9









#33
 Chrysene
 Concen: 0.404 ng
 RT: 21.139 min Scan# 21
 Delta R.T. -0.009 min
 Lab File: BN033824.D
 Acq: 09 Sep 2024 19:38




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion:	228	Resp:	14975
Ion Ratio	Lower	Upper	
228	100		
226	31.4	25.0	37.6
229	19.5	15.6	23.4



#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.386 ng
 RT: 21.050 min Scan# 2384
 Delta R.T. -0.009 min
 Lab File: BN033824.D
 Acq: 09 Sep 2024 19:38

Tgt Ion:	149	Resp:	7882
Ion Ratio	Lower	Upper	
149	100		
167	26.6	21.0	31.4
279	3.0	2.4	3.6









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.259 min Scan# 21
Delta R.T. -0.012 min
Lab File: BN033824.D
Acq: 09 Sep 2024 19:38




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:264 Resp: 10868
Ion Ratio Lower Upper
264 100
260 26.4 21.1 31.7
265 81.5 72.0 108.0



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.396 ng
RT: 25.390 min Scan# 3555
Delta R.T. -0.012 min
Lab File: BN033824.D
Acq: 09 Sep 2024 19:38


Tgt Ion:276 Resp: 17791
Ion Ratio Lower Upper
276 100
138 27.9 23.0 34.6
277 24.7 20.2 30.4








#37
Benzo(b)fluoranthene
Concen: 0.394 ng
RT: 22.621 min Scan# 2608
Delta R.T. -0.009 min
Lab File: BN033824.D
Acq: 09 Sep 2024 19:38

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC







Tgt Ion:252 Resp: 15730
Ion Ratio Lower Upper
252 100
253 26.8 21.8 32.6
125 17.4 15.4 23.0



#38
Benzo(k)fluoranthene
Concen: 0.405 ng
RT: 22.662 min Scan# 2622
Delta R.T. -0.012 min
Lab File: BN033824.D
Acq: 09 Sep 2024 19:38

Tgt Ion:252 Resp: 15789
Ion Ratio Lower Upper
252 100
253 27.5 21.8 32.8
125 19.5 15.4 23.2





#39

Benzo(a)pyrene

Concen: 0.395 ng

RT: 23.162 min Scan# 21

Delta R.T. -0.015 min

Lab File: BN033824.D


Acq: 09 Sep 2024 19:38

Instrument :



BNA_N

ClientSampleId :

SSTDCCC0.4EC



Tgt Ion:	252	Resp:	13201
Ion Ratio	Lower	Upper	
252	100		
253	29.9	24.0	36.0
125	21.1	19.1	28.7



#40

Dibenzo(a,h)anthracene

Concen: 0.394 ng




RT: 25.405 min Scan# 3560


Delta R.T. -0.012 min

Lab File: BN033824.D

Acq: 09 Sep 2024 19:38


Tgt Ion:	278	Resp:	14176
Ion Ratio	Lower	Upper	
278	100		
139	21.0	16.7	25.1
279	27.8	23.4	35.0





#41
Benzo(g,h,i)perylene
Concen: 0.392 ng
RT: 26.033 min Scan# 31
Delta R.T. -0.015 min
Lab File: BN033824.D
Acq: 09 Sep 2024 19:38

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:276 Resp: 15291
Ion Ratio Lower Upper
276 100
277 25.3 20.3 30.5
138 25.8 20.9 31.3

