

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090924\
 Data File : BN033833.D
 Acq On : 10 Sep 2024 01:07
 Operator : MA/JU
 Sample : P3886-04
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BR-01-190-090624

Quant Time: Sep 10 03:20:33 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN090424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 04 18:36:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.509	152	4607	0.400	ng	0.00
7) Naphthalene-d8	10.261	136	11637	0.400	ng	-0.01
13) Acenaphthene-d10	14.146	164	5507	0.400	ng	-0.01
19) Phenanthrene-d10	16.892	188	10854	0.400	ng	#-0.01
29) Chrysene-d12	21.112	240	8445	0.400	ng	# 0.00
35) Perylene-d12	23.262	264	9146	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.154	112	1884	0.134	ng	0.00
5) Phenol-d6	6.700	99	1350	0.081	ng	0.00
8) Nitrobenzene-d5	8.638	82	3582	0.362	ng	-0.01
11) 2-Methylnaphthalene-d10	11.862	152	5496	0.334	ng	-0.01
14) 2,4,6-Tribromophenol	15.638	330	950	0.342	ng	-0.01
15) 2-Fluorobiphenyl	12.756	172	7753	0.338	ng	-0.01
27) Fluoranthene-d10	18.938	212	11673	0.436	ng	0.00
31) Terphenyl-d14	19.555	244	8359	0.464	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.161	88	33309	6.355	ng	99
9) Naphthalene	10.314	128	910	0.029	ng	# 76
25) Phenanthrene	16.942	178	721	0.024	ng	# 93
34) Bis(2-ethylhexyl)phtha...	21.059	149	1642	0.100	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090924\
Data File : BN033833.D
Acq On : 10 Sep 2024 01:07
Operator : MA/JU
Sample : P3886-04
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BR-01-190-090624

Quant Time: Sep 10 03:20:33 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN090424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 04 18:36:01 2024
Response via : Initial Calibration

