




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090924\
Data File : BN033825.D
Acq On : 09 Sep 2024 20:15
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 10 04:22:46 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 10 04:22:28 2024
Response via : Initial Calibration



TIC: BN033825.D\data.ms

(70) Pentachlorophenol (C)




16.557min (0.000) 22357.28 ng

response 412844

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	64.57
264.00	61.60	60.13
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN090924\
Data File : BN033825.D
Acq On : 09 Sep 2024 20:15
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 10 04:22:46 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 10 04:22:28 2024
Response via : Initial Calibration



TIC: BN033825.D\data.ms

(77) Benzidine

19.169min (0.000) 0.00 ng

response 2477118

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.76
183.00	13.20	12.02
0.00	0.00	0.00