

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091022\
 Data File : BN021690.D
 Acq On : 10 Sep 2022 09:44
 Operator : CG/JU
 Sample : N4508-01
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW-05-3-13-090122

Quant Time: Sep 12 01:07:22 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 09 23:19:14 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 09/12/2022
 Supervised By :Jagrut Upadhyay 09/12/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.061	152	5038	0.400	ng	0.00
7) Naphthalene-d8	10.887	136	16459	0.400	ng	-0.01
13) Acenaphthene-d10	14.707	164	9839	0.400	ng	-0.01
19) Phenanthrene-d10	17.459	188	20457	0.400	ng	-0.01
29) Chrysene-d12	21.646	240	13779	0.400	ng	# 0.00
35) Perylene-d12	24.162	264	10278	0.400	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.591	112	1834	0.186	ng	0.00
5) Phenol-d6	7.216	99	1420	0.113	ng	0.00
8) Nitrobenzene-d5	9.232	82	4194m	0.377	ng	0.00
11) 2-Methylnaphthalene-d10	12.474	152	12422	0.335	ng	0.00
14) 2,4,6-Tribromophenol	16.208	330	1410	0.436	ng	0.00
15) 2-Fluorobiphenyl	13.339	172	15813	0.368	ng	0.00
27) Fluoranthene-d10	19.493	212	21273	0.404	ng	0.00
31) Terphenyl-d14	20.079	244	14581	0.471	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.548	149	3182	0.165	ng	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091022\
Data File : BN021690.D
Acq On : 10 Sep 2022 09:44
Operator : CG/JU
Sample : N4508-01
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW-05-3-13-090122

Quant Time: Sep 12 01:07:22 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 09 23:19:14 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/12/2022
Supervised By :Jagrut Upadhyay 09/12/2022

