

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091022\
 Data File : BN021710.D
 Acq On : 10 Sep 2022 22:36
 Operator : CG/JU
 Sample : N4558-05
 Misc :
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB-RW5-GW-203-205

Quant Time: Sep 12 01:14:37 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 09 23:19:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.062	152	5761	0.400	ng	0.00
7) Naphthalene-d8	10.887	136	19006	0.400	ng	-0.01
13) Acenaphthene-d10	14.707	164	11010	0.400	ng	-0.01
19) Phenanthrene-d10	17.459	188	21840	0.400	ng	-0.01
29) Chrysene-d12	21.646	240	12819	0.400	ng	# 0.00
35) Perylene-d12	24.162	264	10537	0.400	ng	#-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.592	112	1614	0.143	ng	0.00
5) Phenol-d6	7.217	99	1902	0.132	ng	0.00
8) Nitrobenzene-d5	9.233	82	4444	0.346	ng	0.00
11) 2-Methylnaphthalene-d10	12.474	152	14101	0.329	ng	0.00
14) 2,4,6-Tribromophenol	16.208	330	1148	0.318	ng	0.00
15) 2-Fluorobiphenyl	13.339	172	16013	0.333	ng	0.00
27) Fluoranthene-d10	19.493	212	21201	0.378	ng	0.00
31) Terphenyl-d14	20.079	244	14883	0.517	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.403	88	5719	0.794	ng	95
34) Bis(2-ethylhexyl)phtha...	21.539	149	6145	0.342	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091022\
Data File : BN021710.D
Acq On : 10 Sep 2022 22:36
Operator : CG/JU
Sample : N4558-05
Misc :
ALS Vial : 49 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-VPB-RW5-GW-203-205

Quant Time: Sep 12 01:14:37 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 09 23:19:14 2022
Response via : Initial Calibration

