

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN091222\  
 Data File : BN021715.D  
 Acq On : 12 Sep 2022 10:46  
 Operator : CG/JU  
 Sample : SSTDCCC0.4  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 SSTDCCC0.4

Quant Time: Sep 13 02:33:25 2022  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\8270-SIM-BN082922.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Sep 13 02:32:27 2022  
 Response via : Initial Calibration


| Compound                      | R.T.   | QIon | Response | Conc  | Units | Dev(Min) |
|-------------------------------|--------|------|----------|-------|-------|----------|
| -----                         |        |      |          |       |       |          |
| Internal Standards            |        |      |          |       |       |          |
| 1) 1,4-Dichlorobenzene-d4     | 8.062  | 152  | 5281     | 0.400 | ng    | 0.00     |
| 7) Naphthalene-d8             | 10.887 | 136  | 17434    | 0.400 | ng    | 0.00     |
| 13) Acenaphthene-d10          | 14.707 | 164  | 10644    | 0.400 | ng    | 0.00     |
| 19) Phenanthrene-d10          | 17.459 | 188  | 20765    | 0.400 | ng    | 0.00     |
| 29) Chrysene-d12              | 21.655 | 240  | 12263    | 0.400 | ng    | 0.00     |
| 35) Perylene-d12              | 24.167 | 264  | 10939    | 0.400 | ng    | # 0.00   |
| System Monitoring Compounds   |        |      |          |       |       |          |
| 4) 2-Fluorophenol             | 5.584  | 112  | 5772     | 0.558 | ng    | 0.00     |
| 5) Phenol-d6                  | 7.209  | 99   | 7154     | 0.543 | ng    | 0.00     |
| 8) Nitrobenzene-d5            | 9.233  | 82   | 5379     | 0.457 | ng    | 0.00     |
| 11) 2-Methylnaphthalene-d10   | 12.475 | 152  | 16399    | 0.417 | ng    | 0.00     |
| 14) 2,4,6-Tribromophenol      | 16.208 | 330  | 1343     | 0.384 | ng    | 0.00     |
| 15) 2-Fluorobiphenyl          | 13.339 | 172  | 16733    | 0.360 | ng    | 0.00     |
| 27) Fluoranthene-d10          | 19.493 | 212  | 18620    | 0.349 | ng    | 0.00     |
| 31) Terphenyl-d14             | 20.088 | 244  | 11604    | 0.421 | ng    | 0.00     |
| Target Compounds              |        |      |          |       |       |          |
|                               |        |      |          |       |       | Qvalue   |
| 2) 1,4-Dioxane                | 3.403  | 88   | 2595     | 0.393 | ng    | # 89     |
| 3) n-Nitrosodimethylamine     | 3.743  | 42   | 2646     | 0.445 | ng    | 99       |
| 6) bis(2-Chloroethyl)ether    | 7.477  | 93   | 6612     | 0.399 | ng    | 99       |
| 9) Naphthalene                | 10.941 | 128  | 19598    | 0.379 | ng    | 99       |
| 10) Hexachlorobutadiene       | 11.218 | 225  | 3223     | 0.361 | ng    | # 99     |
| 12) 2-Methylnaphthalene       | 12.180 | 142  | 4038     | 0.596 | ng    | # 96     |
| 16) Acenaphthylene            | 14.429 | 152  | 18805    | 0.376 | ng    | 100      |
| 17) Acenaphthene              | 14.771 | 154  | 13132    | 0.373 | ng    | 99       |
| 18) Fluorene                  | 15.755 | 166  | 16707    | 0.385 | ng    | 99       |
| 21) 4-Bromophenyl-phenylether | 16.649 | 248  | 4923     | 0.379 | ng    | # 93     |
| 22) Hexachlorobenzene         | 16.762 | 284  | 5913     | 0.372 | ng    | 99       |
| 23) Atrazine                  | 16.916 | 200  | 3450     | 0.455 | ng    | 96       |
| 24) Pentachlorophenol         | 17.111 | 266  | 1394     | 0.495 | ng    | 99       |
| 25) Phenanthrene              | 17.500 | 178  | 26380    | 0.368 | ng    | 100      |
| 26) Anthracene                | 17.593 | 178  | 21090    | 0.370 | ng    | 99       |
| 28) Fluoranthene              | 19.522 | 202  | 24791    | 0.333 | ng    | 99       |
| 30) Pyrene                    | 19.885 | 202  | 26740    | 0.409 | ng    | 97       |
| 32) Benzo(a)anthracene        | 21.637 | 228  | 17404    | 0.408 | ng    | 100      |
| 33) Chrysene                  | 21.691 | 228  | 18755    | 0.369 | ng    | 100      |
| 34) Bis(2-ethylhexyl)phtha... | 21.548 | 149  | 10506    | 0.611 | ng    | 99       |
| 36) Indeno(1,2,3-cd)pyrene    | 26.816 | 276  | 16339    | 0.331 | ng    | 98       |
| 37) Benzo(b)fluoranthene      | 23.393 | 252  | 16433    | 0.305 | ng    | 97       |
| 38) Benzo(k)fluoranthene      | 23.442 | 252  | 16292    | 0.300 | ng    | 95       |
| 39) Benzo(a)pyrene            | 24.053 | 252  | 13420    | 0.357 | ng    | 96       |
| 40) Dibenzo(a,h)anthracene    | 26.837 | 278  | 12424    | 0.312 | ng    | 96       |
| 41) Benzo(g,h,i)perylene      | 27.629 | 276  | 14975    | 0.343 | ng    | 98       |
| -----                         |        |      |          |       |       |          |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN091222\  
Data File : BN021715.D  
Acq On : 12 Sep 2022 10:46  
Operator : CG/JU  
Sample : SSTDCCC0.4  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
SSTDCCC0.4

Quant Time: Sep 13 02:33:25 2022  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\8270-SIM-BN082922.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Sep 13 02:32:27 2022  
Response via : Initial Calibration









#1  
1,4-Dichlorobenzene-d4  
Concen: 0.400 ng  
RT: 8.062 min Scan# 6  
Delta R.T. -0.000 min  
Lab File: BN021715.D  
Acq: 12 Sep 2022 10:46




Instrument :  
BNA\_N  
ClientSampleId :  
SSTDCCC0.4


Tgt Ion:152 Resp: 5281  
Ion Ratio Lower Upper  
152 100  
150 151.6 121.4 182.0  
115 61.1 48.6 72.8



#2  
1,4-Dioxane  
Concen: 0.393 ng  
RT: 3.403 min Scan# 17  
Delta R.T. 0.000 min  
Lab File: BN021715.D  
Acq: 12 Sep 2022 10:46

Tgt Ion: 88 Resp: 2595  
Ion Ratio Lower Upper  
88 100  
43 27.1 35.8 53.8#  
58 71.6 57.1 85.7









#3  
 n-Nitrosodimethylamine  
 Concen: 0.445 ng  
 RT: 3.743 min Scan# 64  
 Delta R.T. 0.000 min  
 Lab File: BN021715.D  
 Acq: 12 Sep 2022 10:46




Instrument :  
 BNA\_N  
 ClientSampleId :  
 SSTDCCC0.4


Tgt Ion: 42 Resp: 2646  
 Ion Ratio Lower Upper  
 42 100  
 74 140.2 113.0 169.6  
 44 10.5 10.5 15.7



#4  
 2-Fluorophenol  
 Concen: 0.558 ng  
 RT: 5.584 min Scan# 319  
 Delta R.T. 0.000 min  
 Lab File: BN021715.D  
 Acq: 12 Sep 2022 10:46

Tgt Ion: 112 Resp: 5772  
 Ion Ratio Lower Upper  
 112 100  
 64 56.8 43.5 65.3  
 63 29.5 22.6 34.0









#5  
Phenol-d6  
Concen: 0.543 ng  
RT: 7.209 min Scan# 544  
Delta R.T. 0.000 min  
Lab File: BN021715.D  
Acq: 12 Sep 2022 10:46




Instrument :  
BNA\_N  
ClientSampleId :  
SSTDCCC0.4


|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 99    | Resp: | 7154  |
| Ion      | Ratio | Lower | Upper |
| 99       | 100   |       |       |
| 42       | 18.2  | 15.5  | 23.3  |
| 71       | 31.6  | 24.2  | 36.2  |



#6  
bis(2-Chloroethyl)ether  
Concen: 0.399 ng  
RT: 7.477 min Scan# 581  
Delta R.T. 0.000 min  
Lab File: BN021715.D  
Acq: 12 Sep 2022 10:46

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 93    | Resp: | 6612  |
| Ion      | Ratio | Lower | Upper |
| 93       | 100   |       |       |
| 63       | 73.8  | 57.8  | 86.8  |
| 95       | 32.1  | 25.9  | 38.9  |









#7  
Naphthalene-d8  
Concen: 0.400 ng  
RT: 10.887 min Scan# 91  
Delta R.T. 0.000 min  
Lab File: BN021715.D  
Acq: 12 Sep 2022 10:46




Instrument :  
BNA\_N  
ClientSampleId :  
SSTDCCC0.4


|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 136   | Resp: | 17434 |
| Ion Ratio | Lower | Upper |       |
| 136       | 100   |       |       |
| 137       | 11.3  | 9.3   | 13.9  |
| 54        | 8.3   | 6.9   | 10.3  |
| 68        | 7.2   | 6.1   | 9.1   |



#8  
Nitrobenzene-d5  
Concen: 0.457 ng  
RT: 9.233 min Scan# 783  
Delta R.T. 0.000 min  
Lab File: BN021715.D  
Acq: 12 Sep 2022 10:46

|           |       |       |      |
|-----------|-------|-------|------|
| Tgt Ion:  | 82    | Resp: | 5379 |
| Ion Ratio | Lower | Upper |      |
| 82        | 100   |       |      |
| 128       | 41.7  | 30.9  | 46.3 |
| 54        | 48.4  | 42.7  | 64.1 |









#9  
Naphthalene  
Concen: 0.379 ng  
RT: 10.941 min Scan# 941  
Delta R.T. 0.000 min  
Lab File: BN021715.D  
Acq: 12 Sep 2022 10:46




Instrument :  
BNA\_N  
ClientSampleId :  
SSTDCCC0.4


Tgt Ion:128 Resp: 19598  
Ion Ratio Lower Upper  
128 100  
129 11.5 9.3 13.9  
127 13.2 11.0 16.4



#10  
Hexachlorobutadiene  
Concen: 0.361 ng  
RT: 11.218 min Scan# 969  
Delta R.T. 0.000 min  
Lab File: BN021715.D  
Acq: 12 Sep 2022 10:46

Tgt Ion:225 Resp: 3223  
Ion Ratio Lower Upper  
225 100  
223 0.0 0.0 0.0  
227 63.4 51.2 76.8









#11  
2-Methylnaphthalene-d10  
Concen: 0.417 ng  
RT: 12.475 min Scan# 1174  
Delta R.T. 0.000 min  
Lab File: BN021715.D  
Acq: 12 Sep 2022 10:46




Instrument :  
BNA\_N  
ClientSampleId :  
SSTDCCC0.4


Tgt Ion:152 Resp: 16399  
Ion Ratio Lower Upper  
152 100  
151 19.2 14.5 21.7



#12  
2-Methylnaphthalene  
Concen: 0.596 ng  
RT: 12.180 min Scan# 1174  
Delta R.T. 0.000 min  
Lab File: BN021715.D  
Acq: 12 Sep 2022 10:46

Tgt Ion:142 Resp: 4038  
Ion Ratio Lower Upper  
142 100  
141 14.0 12.2 18.2  
115 7.4 7.9 11.9#









#13  
Acenaphthene-d10  
Concen: 0.400 ng  
RT: 14.707 min Scan# 1466  
Delta R.T. 0.000 min  
Lab File: BN021715.D  
Acq: 12 Sep 2022 10:46




Instrument :  
BNA\_N  
ClientSampleId :  
SSTDCCC0.4


Tgt Ion:164 Resp: 10644  
Ion Ratio Lower Upper  
164 100  
162 106.1 82.2 123.4  
160 50.6 39.5 59.3



#14  
2,4,6-Tribromophenol  
Concen: 0.384 ng  
RT: 16.208 min Scan# 1608  
Delta R.T. 0.000 min  
Lab File: BN021715.D  
Acq: 12 Sep 2022 10:46

Tgt Ion:330 Resp: 1343  
Ion Ratio Lower Upper  
330 100  
332 98.8 79.3 118.9  
141 46.5 37.1 55.7









#15  
2-Fluorobiphenyl  
Concen: 0.360 ng  
RT: 13.339 min Scan# 11  
Delta R.T. 0.000 min  
Lab File: BN021715.D  
Acq: 12 Sep 2022 10:46




Instrument :  
BNA\_N  
ClientSampleId :  
SSTDCCC0.4


Tgt Ion:172 Resp: 16733  
Ion Ratio Lower Upper  
172 100  
171 34.9 27.8 41.6  
170 23.5 18.6 28.0



#16  
Acenaphthylene  
Concen: 0.376 ng  
RT: 14.429 min Scan# 1440  
Delta R.T. 0.000 min  
Lab File: BN021715.D  
Acq: 12 Sep 2022 10:46

Tgt Ion:152 Resp: 18805  
Ion Ratio Lower Upper  
152 100  
151 19.4 15.4 23.2  
153 12.9 10.3 15.5









#17  
Acenaphthene  
Concen: 0.373 ng  
RT: 14.771 min Scan# 1472  
Delta R.T. 0.000 min  
Lab File: BN021715.D  
Acq: 12 Sep 2022 10:46




Instrument :  
BNA\_N  
ClientSampleId :  
SSTDCCC0.4


Tgt Ion:154 Resp: 13132  
Ion Ratio Lower Upper  
154 100  
153 114.0 89.5 134.3  
152 57.3 45.5 68.3



#18  
Fluorene  
Concen: 0.385 ng  
RT: 15.755 min Scan# 1564  
Delta R.T. 0.000 min  
Lab File: BN021715.D  
Acq: 12 Sep 2022 10:46

Tgt Ion:166 Resp: 16707  
Ion Ratio Lower Upper  
166 100  
165 97.4 78.7 118.1  
167 13.5 10.6 16.0









#19  
Phenanthrene-d10  
Concen: 0.400 ng  
RT: 17.459 min Scan# 1730  
Delta R.T. 0.000 min  
Lab File: BN021715.D  
Acq: 12 Sep 2022 10:46




Instrument :  
BNA\_N  
ClientSampleId :  
SSTDCCC0.4


Tgt Ion:188 Resp: 20765  
Ion Ratio Lower Upper  
188 100  
94 0.0 0.0 0.0  
80 13.5 9.5 14.3



#21  
4-Bromophenyl-phenylether  
Concen: 0.379 ng  
RT: 16.649 min Scan# 1651  
Delta R.T. 0.000 min  
Lab File: BN021715.D  
Acq: 12 Sep 2022 10:46

Tgt Ion:248 Resp: 4923  
Ion Ratio Lower Upper  
248 100  
250 99.6 81.8 122.6  
141 60.1 38.5 57.7#









#22  
Hexachlorobenzene  
Concen: 0.372 ng  
RT: 16.762 min Scan# 1662  
Delta R.T. 0.000 min  
Lab File: BN021715.D  
Acq: 12 Sep 2022 10:46




Instrument :  
BNA\_N  
ClientSampleId :  
SSTDCCC0.4


Tgt Ion:284 Resp: 5913  
Ion Ratio Lower Upper  
284 100  
142 41.3 32.3 48.5  
249 31.6 25.4 38.2



#23  
Atrazine  
Concen: 0.455 ng  
RT: 16.916 min Scan# 1677  
Delta R.T. 0.000 min  
Lab File: BN021715.D  
Acq: 12 Sep 2022 10:46

Tgt Ion:200 Resp: 3450  
Ion Ratio Lower Upper  
200 100  
173 28.3 24.3 36.5  
215 45.4 38.4 57.6





#24

Pentachlorophenol

Concen: 0.495 ng

RT: 17.111 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN021715.D

Acq: 12 Sep 2022 10:46

Instrument :

BNA\_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:266 Resp: 1394

Ion Ratio Lower Upper

266 100

264 62.3 49.0 73.6

268 64.1 50.9 76.3



#25

Phenanthrene

Concen: 0.368 ng

RT: 17.500 min Scan# 1734

Delta R.T. 0.000 min

Lab File: BN021715.D

Acq: 12 Sep 2022 10:46




Tgt Ion:178 Resp: 26380


Ion Ratio Lower Upper

178 100

176 19.0 15.4 23.2

179 15.2 12.2 18.4









#26  
 Anthracene  
 Concen: 0.370 ng  
 RT: 17.593 min Scan# 11  
 Delta R.T. 0.000 min  
 Lab File: BN021715.D  
 Acq: 12 Sep 2022 10:46




Instrument :  
 BNA\_N  
 ClientSampleId :  
 SSTDCCC0.4


|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 178   | Resp: | 21090 |
| Ion Ratio | Lower | Upper |       |
| 178       | 100   |       |       |
| 176       | 18.6  | 14.6  | 22.0  |
| 179       | 15.1  | 12.2  | 18.2  |



#27  
 Fluoranthene-d10  
 Concen: 0.349 ng  
 RT: 19.493 min Scan# 2135  
 Delta R.T. 0.000 min  
 Lab File: BN021715.D  
 Acq: 12 Sep 2022 10:46

|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 212   | Resp: | 18620 |
| Ion Ratio | Lower | Upper |       |
| 212       | 100   |       |       |
| 106       | 17.5  | 14.2  | 21.2  |
| 104       | 9.8   | 7.8   | 11.6  |









#28  
Fluoranthene  
Concen: 0.333 ng  
RT: 19.522 min Scan# 2142  
Delta R.T. 0.000 min  
Lab File: BN021715.D  
Acq: 12 Sep 2022 10:46




Instrument :  
BNA\_N  
ClientSampleId :  
SSTDCCC0.4


Tgt Ion:202 Resp: 24791  
Ion Ratio Lower Upper  
202 100  
101 13.2 10.9 16.3  
203 16.9 13.8 20.6



#29  
Chrysene-d12  
Concen: 0.400 ng  
RT: 21.655 min Scan# 2441  
Delta R.T. 0.000 min  
Lab File: BN021715.D  
Acq: 12 Sep 2022 10:46

Tgt Ion:240 Resp: 12263  
Ion Ratio Lower Upper  
240 100  
120 14.7 10.2 15.4  
236 27.4 22.1 33.1









#30  
Pyrene  
Concen: 0.409 ng  
RT: 19.885 min Scan# 21  
Delta R.T. 0.000 min  
Lab File: BN021715.D  
Acq: 12 Sep 2022 10:46




Instrument :  
BNA\_N  
ClientSampleId :  
SSTDCCC0.4


Tgt Ion:202 Resp: 26740  
Ion Ratio Lower Upper  
202 100  
200 18.9 13.9 20.9  
203 16.1 11.9 17.9



#31  
Terphenyl-d14  
Concen: 0.421 ng  
RT: 20.088 min Scan# 2266  
Delta R.T. 0.000 min  
Lab File: BN021715.D  
Acq: 12 Sep 2022 10:46

Tgt Ion:244 Resp: 11604  
Ion Ratio Lower Upper  
244 100  
212 8.1 6.5 9.7  
122 12.1 9.1 13.7









#32  
Benzo(a)anthracene  
Concen: 0.408 ng  
RT: 21.637 min Scan# 2439  
Delta R.T. 0.000 min  
Lab File: BN021715.D  
Acq: 12 Sep 2022 10:46




Instrument :  
BNA\_N  
ClientSampleId :  
SSTDCCC0.4


Tgt Ion:228 Resp: 17404  
Ion Ratio Lower Upper  
228 100  
226 27.7 22.0 33.0  
229 19.7 15.9 23.9



#33  
Chrysene  
Concen: 0.369 ng  
RT: 21.691 min Scan# 2445  
Delta R.T. 0.000 min  
Lab File: BN021715.D  
Acq: 12 Sep 2022 10:46

Tgt Ion:228 Resp: 18755  
Ion Ratio Lower Upper  
228 100  
226 30.2 24.2 36.2  
229 19.7 15.9 23.9









#34  
 Bis(2-ethylhexyl)phthalate  
 Concen: 0.611 ng  
 RT: 21.548 min Scan# 2431  
 Delta R.T. 0.000 min  
 Lab File: BN021715.D  
 Acq: 12 Sep 2022 10:46




Instrument :  
 BNA\_N  
 ClientSampleId :  
 SSTDCCC0.4


Tgt Ion:149 Resp: 10506  
 Ion Ratio Lower Upper  
 149 100  
 167 26.4 21.4 32.0  
 279 2.6 2.2 3.2



#35  
 Perylene-d12  
 Concen: 0.400 ng  
 RT: 24.167 min Scan# 3127  
 Delta R.T. 0.000 min  
 Lab File: BN021715.D  
 Acq: 12 Sep 2022 10:46

Tgt Ion:264 Resp: 10939  
 Ion Ratio Lower Upper  
 264 100  
 260 24.5 21.0 31.6  
 265 49.8 32.2 48.2#









#36  
Indeno(1,2,3-cd)pyrene  
Concen: 0.331 ng  
RT: 26.816 min Scan# 4044  
Delta R.T. 0.000 min  
Lab File: BN021715.D  
Acq: 12 Sep 2022 10:46




Instrument :  
BNA\_N  
ClientSampleId :  
SSTDCCC0.4


Tgt Ion: 276 Resp: 16339  
Ion Ratio Lower Upper  
276 100  
138 32.1 27.4 41.2  
277 25.0 19.9 29.9



#37  
Benzo(b)fluoranthene  
Concen: 0.305 ng  
RT: 23.393 min Scan# 2862  
Delta R.T. 0.000 min  
Lab File: BN021715.D  
Acq: 12 Sep 2022 10:46

Tgt Ion: 252 Resp: 16433  
Ion Ratio Lower Upper  
252 100  
253 26.7 19.8 29.6  
125 17.8 13.7 20.5









#38  
Benzo(k)fluoranthene  
Concen: 0.300 ng  
RT: 23.442 min Scan# 21  
Delta R.T. 0.000 min  
Lab File: BN021715.D  
Acq: 12 Sep 2022 10:46




Instrument :  
BNA\_N  
ClientSampleId :  
SSTDCCC0.4


Tgt Ion:252 Resp: 16292  
Ion Ratio Lower Upper  
252 100  
253 26.9 19.5 29.3  
125 17.4 12.3 18.5



#39  
Benzo(a)pyrene  
Concen: 0.357 ng  
RT: 24.053 min Scan# 3088  
Delta R.T. 0.000 min  
Lab File: BN021715.D  
Acq: 12 Sep 2022 10:46

Tgt Ion:252 Resp: 13420  
Ion Ratio Lower Upper  
252 100  
253 29.2 21.1 31.7  
125 21.9 16.6 25.0









#40  
Dibenzo(a,h)anthracene  
Concen: 0.312 ng  
RT: 26.837 min Scan# 4053  
Delta R.T. 0.000 min  
Lab File: BN021715.D  
Acq: 12 Sep 2022 10:46

Instrument :  
BNA\_N  
ClientSampleId :  
SSTDCCC0.4

Tgt Ion:278 Resp: 12424  
Ion Ratio Lower Upper  
278 100  
139 25.0 21.8 32.6  
279 27.6 20.4 30.6



#41  
Benzo(g,h,i)perylene  
Concen: 0.343 ng  
RT: 27.629 min Scan# 4311  
Delta R.T. 0.000 min  
Lab File: BN021715.D  
Acq: 12 Sep 2022 10:46

Tgt Ion:276 Resp: 14975  
Ion Ratio Lower Upper  
276 100  
277 25.2 19.7 29.5  
138 30.6 23.7 35.5

