

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091322\
 Data File : BN021725.D
 Acq On : 13 Sep 2022 12:30
 Operator : CG/JU
 Sample : PB147570BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB147570BL

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 09/14/2022
 Supervised By : Jagrut Upadhyay 09/14/2022

Quant Time: Sep 14 01:00:53 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 13 02:32:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.062	152	4838	0.400	ng	0.00
7) Naphthalene-d8	10.888	136	14909	0.400	ng	0.00
13) Acenaphthene-d10	14.718	164	8222	0.400	ng	0.01
19) Phenanthrene-d10	17.470	188	17190	0.400	ng	0.01
29) Chrysene-d12	21.655	240	10100	0.400	ng	# 0.00
35) Perylene-d12	24.173	264	8083	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.592	112	5082	0.537	ng	0.00
5) Phenol-d6	7.217	99	5948	0.493	ng	0.00
8) Nitrobenzene-d5	9.233	82	4263m	0.423	ng	0.00
11) 2-Methylnaphthalene-d10	12.482	152	13704	0.408	ng	0.00
14) 2,4,6-Tribromophenol	16.219	330	825	0.306	ng	0.01
15) 2-Fluorobiphenyl	13.349	172	12419	0.346	ng	0.01
27) Fluoranthene-d10	19.497	212	16932	0.383	ng	0.00
31) Terphenyl-d14	20.088	244	10285	0.453	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091322\
Data File : BN021725.D
Acq On : 13 Sep 2022 12:30
Operator : CG/JU
Sample : PB147570BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB147570BL

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 09/14/2022
Supervised By : Jagrut Upadhyay 09/14/2022

Quant Time: Sep 14 01:00:53 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 13 02:32:27 2022
Response via : Initial Calibration

