

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091322\
 Data File : BN021724.D
 Acq On : 13 Sep 2022 11:47
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 14 01:00:26 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 13 02:32:27 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
2	1,4-Dioxane	0.500	0.474	5.2	88	0.00
3	n-Nitrosodimethylamine	0.451	0.487	-8.0	102	0.00
4 S	2-Fluorophenol	0.783	1.079	-37.8#	130	0.00
5 S	Phenol-d6	0.998	1.357	-36.0#	132	0.00
6	bis(2-Chloroethyl)ether	1.256	1.239	1.4	91	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	94	0.00
8 S	Nitrobenzene-d5	0.270	0.305	-13.0	120	0.00
9	Naphthalene	1.185	1.131	4.6	97	0.00
10	Hexachlorobutadiene	0.205	0.182	11.2	89	0.00
11 SURR	2-Methylnaphthalene-d10	0.902	1.005	-11.4	119	0.00
12	2-Methylnaphthalene	0.150	0.228	-52.0#	172#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	108	0.00
14 S	2,4,6-Tribromophenol	0.131	0.124	5.3	133	0.00
15 S	2-Fluorobiphenyl	1.749	1.575	9.9	104	0.00
16	Acenaphthylene	1.878	1.770	5.8	119	0.00
17	Acenaphthene	1.321	1.237	6.4	113	0.00
18	Fluorene	1.630	1.565	4.0	116	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	106	0.00
20	4,6-Dinitro-2-methylphenol	0.000	0.000	0.0	0#	-15.71#
21	4-Bromophenyl-phenylether	0.250	0.233	6.8	111	0.00
22	Hexachlorobenzene	0.306	0.282	7.8	107	0.00
23	Atrazine	0.146	0.161	-10.3	148	0.00
24	Pentachlorophenol	0.065	0.072	-10.8	157#	0.01
25	Phenanthrene	1.381	1.267	8.3	108	0.00
26	Anthracene	1.097	1.007	8.2	115	0.00
27 SURR	Fluoranthene-d10	1.028	0.892	13.2	107	0.00
28	Fluoranthene	1.435	1.191	17.0	102	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	88	0.00
30	Pyrene	2.132	2.253	-5.7	98	0.00
31 S	Terphenyl-d14	0.899	0.971	-8.0	98	0.00
32	Benzo(a)anthracene	1.393	1.430	-2.7	105	0.00
33	Chrysene	1.659	1.534	7.5	89	0.00
34	Bis(2-ethylhexyl)phthalate	0.561	0.882	-57.2#	160#	0.00
35 I	Perylene-d12	1.000	1.000	0.0	103	0.00
36	Indeno(1,2,3-cd)pyrene	1.806	1.577	12.7	106	0.00
37	Benzo(b)fluoranthene	1.972	1.704	13.6	100	0.00
38	Benzo(k)fluoranthene	1.984	1.645	17.1	97	0.00
39 C	Benzo(a)pyrene	1.376	1.277	7.2	114	0.00
40	Dibenzo(a,h)anthracene	1.454	1.187	18.4	99	0.00
41	Benzo(g,h,i)perylene	1.595	1.513	5.1	113	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4