



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091324\
Data File : BN033888.D
Acq On : 13 Sep 2024 08:18
Operator : JU/RC
Sample : DFTPP356
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP356

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN091124.MA.M
Title : SVOA CALIBRATION
Last Update : Wed Sep 11 01:58:59 2024



AutoFind: Scans 2443, 2444, 2445; Background Corrected with Scan 2438

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.8	123932	PASS
68	69	0.00	2	1.9	2463	PASS
69	198	0.00	100	43.0	127389	PASS
70	69	0.00	2	0.6	747	PASS
127	198	10	80	51.4	152435	PASS
197	198	0.00	2	0.2	564	PASS
198	198	100	100	100.0	296576	PASS
199	198	5	9	6.9	20365	PASS
275	198	10	60	26.5	78501	PASS
365	198	1	100	3.5	10472	PASS
441	198	0.01	100	11.0	32580	PASS
442	198	50	100	70.6	209291	PASS
443	442	15	24	20.4	42723	PASS

m/z	Abundance
39.90	1052.0
41.00	188.0
43.00	250.0
44.00	3806.0
45.10	163.0
48.90	383.0
51.00	255.0
55.00	166.0
73.00	438.0
75.00	186.0
77.00	181.0
83.90	257.0
86.00	182.0
95.90	260.0
132.90	237.0
191.10	152.0
193.00	328.0
207.00	1977.0
207.90	474.0
209.00	306.0
281.00	494.0

Instrument :
BNA_N
ClientSampleId :
DFTPP356

m/z	Abundance
37.00	467.0
38.00	1096.0
39.00	6989.0
40.00	1204.0
41.00	437.0
44.00	3473.0
44.90	372.0
47.00	154.0
49.00	1230.0
50.00	24168.0
51.00	93648.0
52.00	5299.0
52.90	260.0
55.00	735.0
55.90	2913.0
57.00	6666.0
57.90	246.0
58.90	176.0
61.00	1468.0
62.00	1472.0
63.00	4119.0
64.00	541.0
65.00	1879.0
66.00	158.0
66.80	229.0
67.10	196.0
68.00	2023.0
69.00	95640.0
69.90	450.0
73.00	1419.0
74.00	10172.0
75.00	15185.0
76.00	5375.0
77.00	103920.0
78.00	7358.0
79.00	7111.0
80.00	5914.0
81.00	8057.0
81.90	1995.0
82.90	1537.0
83.90	438.0
84.90	1491.0
86.00	2387.0
87.00	1054.0
87.90	374.0
88.90	284.0
91.00	2020.0
92.00	2188.0
93.00	12808.0
94.00	874.0
94.90	326.0
95.90	805.0
96.90	370.0
98.00	9811.0
99.00	9434.0
100.00	652.0
101.00	4319.0
102.00	379.0
102.90	1270.0
103.90	2308.0
105.00	2883.0
106.00	875.0
107.00	31072.0
108.00	5229.0
109.10	1057.0
110.00	56200.0
111.00	8879.0
111.90	880.0
112.90	259.0
114.90	263.0
116.00	2123.0
117.00	28440.0
118.00	2035.0
118.90	273.0
120.00	370.0
121.00	218.0
121.90	1833.0
123.00	3319.0
123.90	1583.0
125.00	1392.0
127.00	106136.0
128.00	8899.0
129.00	41760.0
130.00	3663.0
130.90	749.0
132.00	435.0
132.80	456.0
133.90	1433.0
134.90	3273.0
136.00	1568.0
137.00	2225.0
138.00	430.0

Instrument :
BNA_N
ClientSampleId :
DFTPP356

138.90	202.0
140.00	601.0
141.00	5194.0
142.00	1682.0
143.00	1343.0
143.90	477.0
144.90	317.0
145.90	961.0
147.00	2917.0
148.00	6344.0
148.90	1327.0
150.00	563.0
151.00	809.0
151.90	541.0
153.00	1689.0
153.90	1297.0
155.00	3243.0
156.00	4315.0
157.10	751.0
157.90	1099.0
158.90	759.0
159.90	1229.0
161.00	2274.0
162.00	691.0
163.00	318.0
164.00	317.0
165.00	1966.0
166.00	1900.0
167.00	10364.0
168.00	7846.0
169.00	1143.0
169.90	537.0
170.90	444.0
171.90	957.0
172.90	1321.0
173.90	2154.0
175.00	4017.0
175.90	989.0
177.00	2281.0
178.00	654.0
179.00	8439.0
180.00	5782.0
181.00	2459.0
181.90	388.0
182.60	228.0
184.00	609.0
185.00	3320.0
186.00	27552.0
187.00	8271.0
188.00	1079.0
189.00	1833.0
190.10	354.0
190.90	1263.0
192.00	2884.0
193.00	3014.0
193.90	566.0
195.00	262.0
196.00	5159.0
197.10	1303.0
197.90	188352.0
199.00	14079.0
199.90	1026.0
200.90	359.0
201.50	812.0
203.00	1706.0
204.00	7161.0
205.00	11995.0
206.00	51488.0
207.00	9257.0
208.00	1965.0
209.00	1098.0
210.10	821.0
210.90	1942.0
212.10	163.0
215.00	699.0
215.90	1140.0
217.00	14114.0
218.00	2074.0
221.00	6346.0
221.80	1778.0
222.90	2620.0
224.00	28040.0
225.00	7164.0
226.10	968.0
227.00	12203.0
228.00	2050.0
229.00	2377.0
229.90	389.0
231.10	1026.0
232.00	156.0
232.90	298.0
233.90	806.0
234.90	905.0
236.10	701.0
237.00	1642.0
238.90	470.0

Instrument :
BNA_N
ClientSampleId :
DFTPP356

239.90	386.0
240.90	754.0
242.00	1455.0
243.00	1469.0
244.00	18264.0
245.00	2359.0
245.90	3723.0
247.00	949.0
247.90	245.0
248.90	868.0
249.70	168.0
252.10	198.0
252.90	379.0
254.00	1002.0
255.00	105968.0
256.00	15826.0
257.00	1296.0
258.00	6491.0
258.90	1301.0
261.00	201.0
265.00	2506.0
265.90	527.0
266.90	161.0
269.90	160.0
271.00	344.0
273.00	2820.0
274.00	7893.0
275.00	46480.0
276.00	6794.0
277.00	4105.0
277.90	948.0
279.10	164.0
281.10	472.0
281.90	323.0
283.00	694.0
284.10	451.0
285.00	956.0
288.90	202.0
290.90	168.0
291.80	216.0
293.00	911.0
293.80	296.0
296.00	12419.0
297.00	1928.0
298.00	152.0
300.80	201.0
302.00	207.0
303.00	1394.0
304.10	420.0
308.00	211.0
308.90	173.0
310.00	160.0
313.90	624.0
314.90	1093.0
316.00	883.0
317.00	192.0
321.00	308.0
321.90	269.0
323.00	4106.0
323.90	681.0
324.90	232.0
326.90	492.0
327.90	413.0
331.80	233.0
332.90	367.0
334.00	2736.0
335.00	808.0
341.00	471.0
341.90	161.0
346.00	808.0
346.90	236.0
352.00	1400.0
353.00	869.0
354.00	1145.0
355.00	249.0
364.90	5400.0
365.90	607.0
371.00	279.0
372.00	2021.0
373.00	527.0
382.90	474.0
383.80	225.0
390.00	246.0
391.00	209.0
400.90	219.0
402.10	639.0
402.80	1033.0
403.90	366.0
405.10	163.0
421.00	793.0
422.00	908.0
423.00	6143.0
424.00	2243.0
425.00	236.0
441.00	13908.0
442.00	98464.0

Instrument :
BNA_N
ClientSampleId :
DFTPP356

443.0020096.0

444.001565.0

444.90178.0

Instrument :
BNA_N
ClientSampleId :
DFTPP356

m/z	Abundance
35.90	159.0
37.00	905.0
38.00	1796.0
39.00	10713.0
40.00	1520.0
41.00	485.0
43.10	255.0
44.00	3443.0
44.90	597.0
47.00	168.0
49.10	1376.0
50.00	36320.0
51.00	146112.0
52.00	8351.0
53.00	451.0
55.10	671.0
56.00	4690.0
57.00	10014.0
58.00	383.0
61.00	1914.0
62.00	2077.0
63.00	5917.0
64.00	1063.0
65.00	3141.0
66.00	207.0
66.90	276.0
68.00	3048.0
68.90	149248.0
70.00	955.0
72.90	1676.0
74.00	15069.0
75.00	23608.0
76.00	7545.0
77.00	161088.0
78.00	11029.0
79.00	11649.0
80.00	9392.0
81.00	11627.0
82.00	2758.0
83.00	2820.0
83.90	496.0
85.00	1810.0
86.00	3820.0
86.90	1917.0
87.90	619.0
89.00	298.0
91.00	2741.0
92.00	3201.0
93.00	20232.0
94.00	1268.0
94.90	390.0
96.00	880.0
97.00	309.0
98.00	15591.0
99.00	12081.0
99.90	1259.0
101.00	6513.0
101.90	202.0
102.90	2136.0
103.90	4821.0
105.00	4234.0
106.10	1331.0
107.00	53472.0
108.00	8696.0
110.00	92456.0
111.00	14517.0
112.00	1951.0
113.00	611.0
114.90	361.0
116.00	3183.0
117.00	45104.0
118.00	3479.0
119.00	334.0
120.00	794.0
122.00	3188.0
123.00	5220.0
124.00	2782.0
125.00	2137.0
126.10	1059.0
127.00	180416.0
128.00	14506.0
129.00	72384.0
130.00	6416.0
130.90	1395.0
132.10	652.0
133.10	355.0
133.90	2175.0
135.00	5961.0
136.00	2272.0
136.90	3382.0
138.00	634.0
139.00	406.0

Instrument :
BNA_N
ClientSampleId :
DFTPP356

139.90	902.0
141.00	9018.0
142.00	3073.0
143.00	2346.0
143.90	574.0
145.00	418.0
145.90	1655.0
147.00	5650.0
148.00	10586.0
149.00	2417.0
150.00	649.0
151.00	1207.0
151.80	701.0
153.00	2976.0
154.00	2335.0
155.00	4728.0
156.00	7282.0
157.10	1270.0
158.00	1563.0
158.90	1256.0
160.00	3062.0
161.00	4687.0
161.90	1520.0
163.00	402.0
163.90	382.0
165.00	3548.0
166.00	2672.0
167.00	17672.0
168.00	9918.0
169.00	1478.0
170.00	625.0
171.00	670.0
171.90	1837.0
173.00	2138.0
174.00	4038.0
175.00	7397.0
175.90	1997.0
176.90	3226.0
177.90	1121.0
178.90	13837.0
180.00	9768.0
181.00	4224.0
182.00	698.0
182.70	282.0
183.90	1255.0
185.00	6834.0
186.00	49120.0
187.00	14822.0
188.00	1687.0
188.90	3377.0
190.00	558.0
191.00	1858.0
191.90	4257.0
193.00	5342.0
193.90	1051.0
195.10	412.0
196.00	9462.0
196.70	1691.0
198.00	352576.0
199.00	22952.0
200.00	1716.0
201.50	1354.0
203.00	2555.0
204.00	12956.0
205.00	22616.0
206.00	94184.0
207.00	14868.0
208.00	3795.0
209.00	1422.0
210.00	1504.0
211.00	3576.0
212.00	754.0
212.90	265.0
215.00	979.0
216.00	2009.0
216.90	25048.0
218.00	3232.0
219.00	368.0
220.30	636.0
221.00	12115.0
221.90	3557.0
223.00	5746.0
224.00	49080.0
225.00	13432.0
226.00	1782.0
227.00	23216.0
228.00	3092.0
229.00	4956.0
229.90	667.0
231.00	1887.0
232.10	430.0
233.10	370.0
234.00	1387.0
235.00	1506.0
236.00	1090.0
236.90	2524.0

Instrument :
BNA_N
ClientSampleId :
DFTPP356

237.90	204.0
238.90	833.0
240.00	703.0
241.00	1099.0
242.00	2674.0
243.00	2982.0
244.00	37112.0
245.00	5115.0
246.00	8449.0
246.90	1532.0
247.90	428.0
249.00	1336.0
250.00	230.0
251.00	372.0
251.90	418.0
252.90	966.0
254.10	1419.0
255.00	205056.0
256.00	28808.0
257.00	2228.0
258.00	12073.0
259.00	2036.0
260.10	326.0
260.90	353.0
264.00	386.0
265.00	4563.0
265.90	875.0
266.90	244.0
269.80	241.0
271.10	655.0
272.00	695.0
273.00	5812.0
274.00	15216.0
275.00	84600.0
276.00	11744.0
277.00	7243.0
277.90	1506.0
279.00	293.0
281.00	523.0
282.00	269.0
283.00	937.0
284.00	796.0
285.00	1357.0
286.00	362.0
288.90	418.0
290.10	223.0
290.90	189.0
292.00	309.0
293.00	1891.0
293.90	543.0
295.00	626.0
296.00	26456.0
297.00	3451.0
298.10	335.0
301.00	305.0
302.10	408.0
303.00	2988.0
304.00	797.0
308.00	452.0
309.00	369.0
310.00	473.0
313.20	179.0
314.00	1305.0
314.90	3188.0
316.00	1458.0
316.90	317.0
321.00	793.0
322.00	353.0
323.00	7024.0
324.00	1636.0
326.00	231.0
326.90	1380.0
327.90	598.0
329.00	154.0
332.00	550.0
333.00	697.0
334.00	5394.0
334.90	1151.0
336.00	219.0
341.00	1173.0
342.00	293.0
345.90	1548.0
346.80	402.0
352.00	2420.0
353.00	1921.0
354.00	2237.0
355.00	530.0
359.00	202.0
365.00	12583.0
365.90	1833.0
369.90	227.0
370.90	620.0
372.00	4165.0
373.00	1306.0
382.90	1192.0
384.00	409.0

Instrument :
BNA_N
ClientSampleId :
DFTPP356

385.10	158.0
389.80	561.0
391.00	387.0
392.00	235.0
400.90	355.0
402.00	1617.0
403.00	2443.0
403.90	1096.0
405.00	250.0
414.90	197.0
421.00	2017.0
422.00	2096.0
423.00	14162.0
424.00	4247.0
424.90	527.0
441.00	35296.0
442.00	207040.0
443.00	46752.0
444.00	4229.0
445.10	370.0

Instrument :
BNA_N
ClientSampleId :
DFTPP356

m/z	Abundance
37.00	544.0
38.00	1839.0
39.00	9383.0
40.00	1392.0
41.00	394.0
41.90	165.0
42.90	223.0
44.00	3289.0
44.90	310.0
49.00	1207.0
50.00	34024.0
51.00	132800.0
52.00	6114.0
52.90	415.0
55.00	752.0
56.00	4158.0
57.00	9138.0
58.00	376.0
61.00	1965.0
62.00	1739.0
63.00	5093.0
64.00	966.0
65.00	2650.0
66.00	324.0
67.00	245.0
68.00	2319.0
69.00	137280.0
70.00	837.0
71.00	198.0
73.10	1793.0
74.00	15153.0
75.00	21880.0
76.00	6773.0
77.00	150976.0
78.00	9650.0
79.00	9918.0
80.00	7874.0
81.00	11476.0
82.00	2673.0
83.00	2536.0
83.90	443.0
85.00	1796.0
86.00	3332.0
87.00	1513.0
88.00	531.0
88.80	387.0
90.00	156.0
90.90	2457.0
92.00	3086.0
93.00	17760.0
94.00	1206.0
95.10	288.0
96.00	947.0
97.00	426.0
98.00	14997.0
99.00	12086.0
99.90	1011.0
101.00	5937.0
101.90	390.0
103.00	2229.0
104.00	4684.0
104.90	3840.0
106.00	1367.0
107.00	49592.0
108.00	7433.0
109.10	1353.0
110.00	86184.0
111.00	14250.0
111.90	1597.0
113.00	554.0
114.80	269.0
116.00	2827.0
117.00	41264.0
118.00	3101.0
119.00	368.0
120.00	720.0
121.00	169.0
121.90	3131.0
122.90	5249.0
124.00	2259.0
125.00	2044.0
127.00	170752.0
128.00	13443.0
129.00	66472.0
130.00	5296.0
131.00	1245.0
132.00	685.0
132.90	488.0
134.00	1973.0
135.00	6269.0
136.00	2574.0
137.00	2809.0

Instrument :
BNA_N
ClientSampleId :
DFTPP356

138.00	521.0
139.00	389.0
139.90	869.0
141.00	9287.0
142.00	2977.0
142.90	2024.0
143.90	603.0
145.10	463.0
145.90	1588.0
147.00	5126.0
148.00	10866.0
148.90	2187.0
150.00	598.0
151.00	1297.0
151.80	596.0
152.10	574.0
152.90	2561.0
153.90	2030.0
155.00	4928.0
156.00	7667.0
157.00	1326.0
157.90	2054.0
158.90	1306.0
160.00	2885.0
161.00	3998.0
162.00	1287.0
162.90	372.0
164.00	592.0
165.00	3748.0
166.00	2955.0
167.00	17712.0
168.00	9046.0
169.00	1664.0
170.00	681.0
171.00	891.0
172.00	1773.0
173.00	2136.0
174.00	4423.0
175.00	6713.0
176.00	2021.0
177.00	3751.0
177.90	1410.0
179.00	13572.0
180.00	10088.0
181.00	4046.0
182.00	703.0
182.90	314.0
184.00	1207.0
185.00	6844.0
186.00	51272.0
187.00	14514.0
188.00	1814.0
189.00	3528.0
189.90	561.0
191.00	1889.0
192.00	4316.0
193.00	5006.0
194.10	931.0
195.00	556.0
196.00	9098.0
198.00	348800.0
199.00	24064.0
199.90	1908.0
201.50	1356.0
203.00	2825.0
204.00	13208.0
205.00	23384.0
206.00	97144.0
207.00	14216.0
208.00	3501.0
209.00	1526.0
210.20	1918.0
211.00	4254.0
211.90	826.0
212.90	436.0
215.00	1246.0
216.00	2199.0
216.90	26472.0
218.00	2973.0
219.10	277.0
221.00	12947.0
221.90	3202.0
223.00	5783.0
224.00	51888.0
225.00	13711.0
226.10	1556.0
227.00	26360.0
228.00	3522.0
229.00	4893.0
230.00	592.0
231.00	2018.0
232.00	436.0
232.90	467.0
234.00	1467.0
235.00	1895.0
236.10	1352.0

Instrument :
BNA_N
ClientSampleId :
DFTPP356

237.00	2248.0
238.00	404.0
239.00	1029.0
239.90	643.0
241.00	1283.0
242.00	2751.0
243.00	3366.0
244.00	41520.0
245.10	5184.0
246.00	9162.0
247.00	1747.0
247.90	427.0
249.00	1478.0
249.90	276.0
250.90	478.0
252.00	322.0
252.90	894.0
255.00	226368.0
256.00	31304.0
257.00	2661.0
258.00	13920.0
259.00	2330.0
259.90	420.0
260.90	383.0
263.90	298.0
265.00	5540.0
265.90	1015.0
267.10	163.0
270.10	239.0
271.00	625.0
271.90	695.0
273.00	6701.0
274.00	17984.0
275.00	104424.0
276.00	12997.0
276.90	9501.0
278.00	1848.0
279.10	308.0
281.00	622.0
282.00	354.0
283.00	1028.0
284.00	695.0
285.00	1732.0
285.90	339.0
288.90	325.0
289.90	417.0
290.80	174.0
292.00	445.0
292.90	2139.0
294.00	564.0
296.00	32304.0
297.00	4565.0
297.90	260.0
301.00	377.0
302.00	744.0
303.00	3624.0
303.90	815.0
307.90	475.0
308.90	301.0
310.00	452.0
313.00	240.0
314.00	1511.0
315.00	3684.0
316.00	1960.0
317.00	192.0
319.90	159.0
321.00	1079.0
322.00	502.0
323.00	9448.0
324.00	1841.0
324.90	264.0
326.00	159.0
327.00	1854.0
327.90	979.0
328.80	201.0
332.00	901.0
332.90	831.0
334.00	6721.0
335.00	1465.0
335.90	272.0
338.90	202.0
340.00	230.0
341.00	1305.0
342.00	331.0
343.00	156.0
345.80	2014.0
346.90	433.0
350.90	160.0
352.00	3409.0
352.90	2225.0
354.00	3024.0
355.00	620.0
356.00	153.0
358.90	217.0
364.90	13434.0
366.00	2084.0

Instrument :
BNA_N
ClientSampleId :
DFTPP356

366.90	162.0
369.90	331.0
371.00	957.0
372.00	5938.0
372.90	1161.0
376.90	169.0
383.10	1413.0
383.90	355.0
384.80	201.0
390.00	709.0
391.00	571.0
392.00	346.0
401.00	460.0
401.90	2210.0
403.00	3372.0
403.90	1124.0
405.00	219.0
414.80	179.0
421.00	2564.0
422.00	2596.0
423.10	17656.0
424.00	4658.0
425.00	586.0
441.00	48536.0
442.00	322368.0
443.00	61320.0
444.00	5409.0
444.90	255.0

Instrument :
BNA_N
ClientSampleId :
DFTPP356