

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091422\
 Data File : BN021759.D
 Acq On : 14 Sep 2022 23:08
 Operator : CG/JU
 Sample : PB147574BL
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB147574BL

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 09/15/2022
 Supervised By :Jagrut Upadhyay 09/15/2022

Quant Time: Sep 15 02:13:41 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 14 01:55:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.055	152	4607	0.400	ng	0.00
7) Naphthalene-d8	10.888	136	13514	0.400	ng	0.00
13) Acenaphthene-d10	14.707	164	7277	0.400	ng	0.00
19) Phenanthrene-d10	17.456	188	16437	0.400	ng	0.00
29) Chrysene-d12	21.652	240	11664	0.400	ng	0.00
35) Perylene-d12	24.156	264	9623	0.400	ng	#-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.585	112	4429	0.367	ng	0.00
5) Phenol-d6	7.217	99	5318	0.348	ng	0.00
8) Nitrobenzene-d5	9.233	82	3694m	0.331	ng	0.00
11) 2-Methylnaphthalene-d10	12.475	152	11446	0.336	ng	0.00
14) 2,4,6-Tribromophenol	16.209	330	637	0.192	ng	0.00
15) 2-Fluorobiphenyl	13.339	172	10610	0.357	ng	0.00
27) Fluoranthene-d10	19.490	212	16143	0.365	ng	0.00
31) Terphenyl-d14	20.076	244	10654	0.400	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091422\
Data File : BN021759.D
Acq On : 14 Sep 2022 23:08
Operator : CG/JU
Sample : PB147574BL
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB147574BL

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/15/2022
Supervised By :Jagrut Upadhyay 09/15/2022

Quant Time: Sep 15 02:13:41 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 14 01:55:04 2022
Response via : Initial Calibration

