

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091422\
 Data File : BN021736.D
 Acq On : 13 Sep 2022 21:51
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN091422

Quant Time: Sep 14 01:57:54 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 14 01:55:04 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
2	1,4-Dioxane	0.517	0.507	1.9	91	0.00
3	n-Nitrosodimethylamine	0.525	0.490	6.7	87	0.00
4 S	2-Fluorophenol	1.046	0.953	8.9	83	0.00
5 S	Phenol-d6	1.327	1.194	10.0	84	0.00
6	bis(2-Chloroethyl)ether	1.316	1.251	4.9	88	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	88	0.00
8 S	Nitrobenzene-d5	0.331	0.297	10.3	85	0.00
9	Naphthalene	1.181	1.130	4.3	89	0.00
10	Hexachlorobutadiene	0.197	0.191	3.0	88	0.00
11 SURR	2-Methylnaphthalene-d10	1.009	0.916	9.2	82	0.00
12	2-Methylnaphthalene	0.254	0.211	16.9	83	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	89	0.00
14 S	2,4,6-Tribromophenol	0.183	0.154	15.8	83	-0.01
15 S	2-Fluorobiphenyl	1.633	1.573	3.7	89	0.00
16	Acenaphthylene	1.910	1.703	10.8	88	0.00
17	Acenaphthene	1.305	1.243	4.8	90	0.00
18	Fluorene	1.735	1.659	4.4	89	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	87	0.00
20	4,6-Dinitro-2-methylphenol	0.000	0.000	0.0	0#	-15.71#
21	4-Bromophenyl-phenylether	0.246	0.231	6.1	88	0.00
22	Hexachlorobenzene	0.277	0.269	2.9	92	0.00
23	Atrazine	0.204	0.175	14.2	84	0.00
24	Pentachlorophenol	0.104	0.074	28.8#	75	0.00
25	Phenanthrene	1.341	1.275	4.9	89	0.00
26	Anthracene	1.136	1.023	9.9	87	0.00
27 SURR	Fluoranthene-d10	1.076	0.994	7.6	87	0.00
28	Fluoranthene	1.447	1.342	7.3	87	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	86	0.00
30	Pyrene	1.986	1.937	2.5	86	0.00
31 S	Terphenyl-d14	0.914	0.909	0.5	88	0.00
32	Benzo(a)anthracene	1.534	1.393	9.2	84	0.00
33	Chrysene	1.595	1.562	2.1	88	0.00
34	Bis(2-ethylhexyl)phthalate	1.025	1.144	-11.6	106	0.00
35 I	Perylene-d12	1.000	1.000	0.0	83	0.00
36	Indeno(1,2,3-cd)pyrene	1.900	1.840	3.2	85	-0.01
37	Benzo(b)fluoranthene	1.817	1.748	3.8	85	0.00
38	Benzo(k)fluoranthene	1.776	1.751	1.4	86	0.00
39 C	Benzo(a)pyrene	1.448	1.356	6.4	84	0.00
40	Dibenzo(a,h)anthracene	1.504	1.465	2.6	86	-0.01
41	Benzo(g,h,i)perylene	1.548	1.502	3.0	86	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
ClientSampleId :
ICVBN091422