

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091520\
 Data File : BN011802.D
 Acq On : 15 Sep 2020 15:41
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN091520

Quant Time: Sep 15 16:52:06 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 15 14:55:25 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	74	0.00
2	1,4-Dioxane	0.562	0.525	6.6	78	0.00
3	n-Nitrosodimethylamine	0.864	0.855	1.0	73	0.00
4 S	2-Fluorophenol	1.107	1.067	3.6	73	0.00
5 S	Phenol-d6	1.423	1.356	4.7	72	0.00
6	bis(2-Chloroethyl)ether	1.162	1.129	2.8	74	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	73	0.00
8 S	Nitrobenzene-d5	0.441	0.424	3.9	72	0.01
9	Naphthalene	1.145	1.107	3.3	74	0.00
10	Hexachlorobutadiene	0.277	0.270	2.5	74	0.00
11 SURR	2-Methylnaphthalene-d10	0.752	0.723	3.9	74	0.00
12	2-Methylnaphthalene	0.821	0.781	4.9	74	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	73	0.00
14 S	2,4,6-Tribromophenol	0.219	0.203	7.3	72	0.00
15 S	2-Fluorobiphenyl	1.792	1.737	3.1	75	0.01
16	Acenaphthylene	1.995	1.832	8.2	72	0.00
17	Acenaphthene	1.411	1.345	4.7	73	0.00
18	Fluorene	1.777	1.667	6.2	73	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	75	0.01
20	4,6-Dinitro-2-methylphenol	0.090	0.102	-13.3	75	0.00
21	4-Bromophenyl-phenylether	0.253	0.233	7.9	73	0.00
22	Hexachlorobenzene	0.297	0.278	6.4	75	0.00
23	Atrazine	0.233	0.234	-0.4	74	0.00
24	Pentachlorophenol	0.125	0.147	-17.6	74	0.00
25	Phenanthrene	1.303	1.232	5.4	74	0.00
26	Anthracene	1.170	1.129	3.5	74	0.00
27 SURR	Fluoranthene-d10	1.304	1.198	8.1	74	0.00
28	Fluoranthene	1.623	1.929	-18.9	73	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	73	0.00
30	Pyrene	1.418	1.396	1.6	73	0.00
31 S	Terphenyl-d14	1.123	1.679	-49.5#	74	0.00
32	Benzo(a)anthracene	1.459	1.665	-14.1	74	0.00
33	Chrysene	1.528	1.781	-16.6	72	0.00
34	Bis(2-ethylhexyl)phthalate	0.654	0.717	-9.6	75	0.00
35	Indeno(1,2,3-cd)pyrene	2.007	2.315	-15.3	68	0.00
36 I	Perylene-d12	1.000	1.000	0.0	71	0.00
37	Benzo(b)fluoranthene	1.649	1.929	-17.0	73	0.00
38	Benzo(k)fluoranthene	1.653	1.861	-12.6	70	0.00
39 C	Benzo(a)pyrene	1.447	1.611	-11.3	70	0.00
40	Dibenzo(a,h)anthracene	1.673	1.924	-15.0	68	0.00
41	Benzo(g,h,i)perylene	1.671	1.918	-14.8	68	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091520\
Data File : BN011802.D
Acq On : 15 Sep 2020 15:41
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN091520

Quant Time: Sep 15 16:52:06 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 15 14:55:25 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0