

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091520\
 Data File : BN011810.D
 Acq On : 15 Sep 2020 21:05
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 16 02:14:22 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 15 14:55:25 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	119	0.00
2	1,4-Dioxane	0.562	0.533	5.2	128	0.00
3	n-Nitrosodimethylamine	0.864	0.733	15.2	101	0.00
4 S	2-Fluorophenol	1.107	1.229	-11.0	135	0.00
5 S	Phenol-d6	1.423	1.612	-13.3	138	0.00
6	bis(2-Chloroethyl)ether	1.162	1.281	-10.2	135	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	127	0.00
8 S	Nitrobenzene-d5	0.441	0.401	9.1	119	0.00
9	Naphthalene	1.145	1.108	3.2	129	0.00
10	Hexachlorobutadiene	0.277	0.243	12.3	115	0.00
11 SURR	2-Methylnaphthalene-d10	0.752	0.701	6.8	126	0.00
12	2-Methylnaphthalene	0.821	0.761	7.3	125	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	121	0.00
14 S	2,4,6-Tribromophenol	0.219	0.198	9.6	115	0.00
15 S	2-Fluorobiphenyl	1.792	1.669	6.9	119	0.00
16	Acenaphthylene	1.995	1.789	10.3	115	0.00
17	Acenaphthene	1.411	1.303	7.7	117	0.00
18	Fluorene	1.777	1.599	10.0	116	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	119	0.00
20	4,6-Dinitro-2-methylphenol	0.090	0.081	10.0	94	0.00
21	4-Bromophenyl-phenylether	0.253	0.218	13.8	109	0.00
22	Hexachlorobenzene	0.297	0.277	6.7	118	0.00
23	Atrazine	0.233	0.221	5.2	110	0.00
24	Pentachlorophenol	0.125	0.147	-17.6	118	0.00
25	Phenanthrene	1.303	1.209	7.2	116	0.00
26	Anthracene	1.170	1.103	5.7	115	0.00
27 SURR	Fluoranthene-d10	1.304	1.179	9.6	116	0.00
28	Fluoranthene	1.623	1.883	-16.0	113	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	113	-0.01
30	Pyrene	1.418	1.395	1.6	114	0.00
31 S	Terphenyl-d14	1.123	1.674	-49.1#	115	0.00
32	Benzo(a)anthracene	1.459	1.616	-10.8	112	0.00
33	Chrysene	1.528	1.750	-14.5	111	0.00
34	Bis(2-ethylhexyl)phthalate	0.654	0.688	-5.2	111	-0.01
35	Indeno(1,2,3-cd)pyrene	2.007	2.226	-10.9	102	0.00
36 I	Perylene-d12	1.000	1.000	0.0	108	0.00
37	Benzo(b)fluoranthene	1.649	1.860	-12.8	107	0.00
38	Benzo(k)fluoranthene	1.653	1.889	-14.3	108	0.00
39 C	Benzo(a)pyrene	1.447	1.636	-13.1	108	0.00
40	Dibenzo(a,h)anthracene	1.673	1.887	-12.8	101	-0.01
41	Benzo(g,h,i)perylene	1.671	1.863	-11.5	101	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091520\
Data File : BN011810.D
Acq On : 15 Sep 2020 21:05
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 16 02:14:22 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 15 14:55:25 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0