

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN091522\  
 Data File : BN021768.D  
 Acq On : 15 Sep 2022 14:48  
 Operator : CG/JU  
 Sample : N4558-11  
 Misc :  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 BP-VPB-RW5-GW-58-60

Quant Time: Sep 16 06:23:45 2022  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\8270-SIM-BN091422.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Sep 14 01:55:04 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc  | Units | Dev(Min)  |
|-------------------------------|--------|------|----------|-------|-------|-----------|
| Internal Standards            |        |      |          |       |       |           |
| 1) 1,4-Dichlorobenzene-d4     | 8.054  | 152  | 4689     | 0.400 | ng    | 0.00      |
| 7) Naphthalene-d8             | 10.888 | 136  | 14943    | 0.400 | ng    | 0.00      |
| 13) Acenaphthene-d10          | 14.707 | 164  | 8946     | 0.400 | ng    | 0.00      |
| 19) Phenanthrene-d10          | 17.456 | 188  | 20433    | 0.400 | ng    | 0.00      |
| 29) Chrysene-d12              | 21.656 | 240  | 15271    | 0.400 | ng    | # 0.00    |
| 35) Perylene-d12              | 24.165 | 264  | 12602    | 0.400 | ng    | # 0.00    |
| System Monitoring Compounds   |        |      |          |       |       |           |
| 4) 2-Fluorophenol             | 5.584  | 112  | 3084     | 0.251 | ng    | 0.00      |
| 5) Phenol-d6                  | 7.209  | 99   | 4078     | 0.262 | ng    | 0.00      |
| 8) Nitrobenzene-d5            | 9.233  | 82   | 2734     | 0.221 | ng    | 0.00      |
| 11) 2-Methylnaphthalene-d10   | 12.471 | 152  | 9112     | 0.242 | ng    | 0.00      |
| 14) 2,4,6-Tribromophenol      | 16.209 | 330  | 556      | 0.136 | ng    | 0.00      |
| 15) 2-Fluorobiphenyl          | 13.339 | 172  | 7958     | 0.218 | ng    | 0.00      |
| 27) Fluoranthene-d10          | 19.494 | 212  | 12491    | 0.227 | ng    | 0.00      |
| 31) Terphenyl-d14             | 20.080 | 244  | 8048     | 0.231 | ng    | 0.00      |
| Target Compounds              |        |      |          |       |       |           |
| 34) Bis(2-ethylhexyl)phtha... | 21.549 | 149  | 20950    | 0.535 | ng    | Qvalue 99 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN091522\  
Data File : BN021768.D  
Acq On : 15 Sep 2022 14:48  
Operator : CG/JU  
Sample : N4558-11  
Misc :  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
BP-VPB-RW5-GW-58-60

Quant Time: Sep 16 06:23:45 2022  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\8270-SIM-BN091422.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Sep 14 01:55:04 2022  
Response via : Initial Calibration

