

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091522\
 Data File : BN021779.D
 Acq On : 15 Sep 2022 22:57
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 16 06:27:26 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 14 01:55:04 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	0.00
2	1,4-Dioxane	0.517	0.501	3.1	86	0.00
3	n-Nitrosodimethylamine	0.525	0.512	2.5	87	0.00
4 S	2-Fluorophenol	1.046	1.034	1.1	86	0.00
5 S	Phenol-d6	1.327	1.322	0.4	89	0.00
6	bis(2-Chloroethyl)ether	1.316	1.255	4.6	84	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	85	0.00
8 S	Nitrobenzene-d5	0.331	0.307	7.3	85	0.00
9	Naphthalene	1.181	1.138	3.6	86	-0.01
10	Hexachlorobutadiene	0.197	0.192	2.5	85	0.00
11 SURR	2-Methylnaphthalene-d10	1.009	0.985	2.4	85	0.00
12	2-Methylnaphthalene	0.254	0.237	6.7	90	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	82	0.00
14 S	2,4,6-Tribromophenol	0.183	0.149	18.6	74	-0.01
15 S	2-Fluorobiphenyl	1.633	1.593	2.4	83	0.00
16	Acenaphthylene	1.910	1.715	10.2	81	0.00
17	Acenaphthene	1.305	1.237	5.2	82	0.00
18	Fluorene	1.735	1.640	5.5	81	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	79	0.00
20	4,6-Dinitro-2-methylphenol	0.000	0.000	0.0	0#	-15.71#
21	4-Bromophenyl-phenylether	0.246	0.229	6.9	79	0.00
22	Hexachlorobenzene	0.277	0.272	1.8	83	0.00
23	Atrazine	0.204	0.169	17.2	73	0.00
24	Pentachlorophenol	0.104	0.065	37.5#	59	0.00
25	Phenanthrene	1.341	1.251	6.7	78	0.00
26	Anthracene	1.136	1.005	11.5	77	0.00
27 SURR	Fluoranthene-d10	1.076	0.965	10.3	76	0.00
28	Fluoranthene	1.447	1.311	9.4	77	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	74	-0.01
30	Pyrene	1.986	1.971	0.8	75	0.00
31 S	Terphenyl-d14	0.914	0.902	1.3	75	0.00
32	Benzo(a)anthracene	1.534	1.396	9.0	72	-0.01
33	Chrysene	1.595	1.588	0.4	77	-0.01
34	Bis(2-ethylhexyl)phthalate	1.025	0.824	19.6	66	-0.01
35 I	Perylene-d12	1.000	1.000	0.0	69	-0.02
36	Indeno(1,2,3-cd)pyrene	1.900	1.745	8.2	67	-0.02
37	Benzo(b)fluoranthene	1.817	1.788	1.6	72	-0.02
38	Benzo(k)fluoranthene	1.776	1.749	1.5	71	-0.01
39 C	Benzo(a)pyrene	1.448	1.351	6.7	70	-0.02
40	Dibenzo(a,h)anthracene	1.504	1.363	9.4	66	-0.02
41	Benzo(g,h,i)perylene	1.548	1.433	7.4	68	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4