



Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN091522\  
 Data File : BN021765.D  
 Acq On : 15 Sep 2022 09:28  
 Operator : CG/JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\8270-SIM-BN091422.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Wed Sep 14 01:55:04 2022






AutoFind: Scans 2482, 2483, 2484; Background Corrected with Scan 2474

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 34.4         | 128507     | PASS                |
| 68             | 69              | 0.00            | 2               | 1.7          | 2472       | PASS                |
| 69             | 198             | 0.00            | 100             | 37.9         | 141619     | PASS                |
| 70             | 69              | 0.00            | 2               | 0.6          | 783        | PASS                |
| 127            | 198             | 10              | 80              | 50.2         | 187341     | PASS                |
| 197            | 198             | 0.00            | 2               | 0.5          | 1887       | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 373227     | PASS                |
| 199            | 198             | 5               | 9               | 6.5          | 24360      | PASS                |
| 275            | 198             | 10              | 60              | 25.9         | 96843      | PASS                |
| 365            | 198             | 1               | 100             | 4.3          | 16028      | PASS                |
| 441            | 198             | 0.01            | 100             | 14.0         | 52237      | PASS                |
| 442            | 442             | 50              | 100             | 100.0        | 328811     | PASS                |
| 443            | 442             | 15              | 24              | 18.7         | 61416      | PASS                |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN091522\  
 Data File : BN021765.D  
 Acq On : 15 Sep 2022 09:28  
 Operator : CG/JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 DFTPP

Quant Time: Sep 16 07:52:20 2022  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270E-Tune.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Sep 16 07:52:15 2022  
 Response via : Initial Calibration



TIC: BN021765.D\data.ms

**(70) Pentachlorophenol (C)**

17.104min ( 0.000) 378205.02 ng




response 513474

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 62.20  | 55.34  |
| 264.00 | 61.60  | 56.81  |
| 0.00   | 0.00   | 0.00   |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN091522\  
Data File : BN021765.D  
Acq On : 15 Sep 2022 09:28  
Operator : CG/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP

Quant Time: Sep 16 07:52:20 2022  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270E-Tune.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Sep 16 07:52:15 2022  
Response via : Initial Calibration



TIC: BN021765.D\data.ms

(77) Benzidine

19.710min ( 0.000) 364647.42 ng

response 2663973

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 15.50  | 14.11  |
| 183.00 | 13.20  | 12.06  |
| 0.00   | 0.00   | 0.00   |

# DDT Breakdown

| Date          | Instrument Name  | DFTPP Data File    |
|---------------|------------------|--------------------|
| 9/15/2022     | BNA_N            | <u>BN021765.D</u>  |
|               |                  |                    |
|               |                  |                    |
| Compound Name | Response         | Retention Time     |
|               |                  |                    |
| DDT           | 1355431          | 20.939             |
| DDD           | 46523            | 20.498             |
| DDE           | 3276             | 19.986             |
|               |                  |                    |
|               |                  |                    |
| SUM(DDD+DDE)  | SUM(DDT+DDD+DDE) | % Breakdown Of DDT |
|               |                  |                    |
| 49799         | 1405230          | 3.54               |
|               |                  |                    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP