

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091620\
 Data File : BN011817.D
 Acq On : 16 Sep 2020 17:30
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 16 17:59:14 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 16 17:28:08 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	135	0.00
2	1,4-Dioxane	0.562	0.583	-3.7	158#	0.00
3	n-Nitrosodimethylamine	0.864	0.694	19.7	109	0.00
4 S	2-Fluorophenol	1.107	1.281	-15.7	159#	0.00
5 S	Phenol-d6	1.423	1.692	-18.9	164#	0.00
6	bis(2-Chloroethyl)ether	1.162	1.263	-8.7	151#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	146	0.00
8 S	Nitrobenzene-d5	0.441	0.417	5.4	143	0.01
9	Naphthalene	1.145	1.084	5.3	146	0.00
10	Hexachlorobutadiene	0.277	0.236	14.8	130	0.00
11 SURR	2-Methylnaphthalene-d10	0.752	0.682	9.3	141	0.00
12	2-Methylnaphthalene	0.821	0.751	8.5	143	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	140	0.00
14 S	2,4,6-Tribromophenol	0.219	0.201	8.2	135	0.00
15 S	2-Fluorobiphenyl	1.792	1.696	5.4	140	0.00
16	Acenaphthylene	1.995	1.796	10.0	134	0.00
17	Acenaphthene	1.411	1.305	7.5	135	0.00
18	Fluorene	1.777	1.597	10.1	134	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	134	0.00
20	4,6-Dinitro-2-methylphenol	0.090	0.079	12.2	103	0.00
21	4-Bromophenyl-phenylether	0.253	0.220	13.0	124	0.00
22	Hexachlorobenzene	0.297	0.281	5.4	136	0.00
23	Atrazine	0.233	0.225	3.4	127	0.00
24	Pentachlorophenol	0.125	0.152	-21.6	138	0.00
25	Phenanthrene	1.303	1.224	6.1	133	0.00
26	Anthracene	1.170	1.118	4.4	131	0.00
27 SURR	Fluoranthene-d10	1.304	1.175	9.9	131	0.00
28	Fluoranthene	1.623	1.913	-17.9	130	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	122	0.00
30	Pyrene	1.418	1.487	-4.9	131	0.00
31 S	Terphenyl-d14	1.123	1.751	-55.9#	129	0.00
32	Benzo(a)anthracene	1.459	1.666	-14.2	124	0.00
33	Chrysene	1.528	1.778	-16.4	121	0.00
34	Bis(2-ethylhexyl)phthalate	0.654	0.728	-11.3	127	0.00
35	Indeno(1,2,3-cd)pyrene	2.007	2.194	-9.3	108	0.00
36 I	Perylene-d12	1.000	1.000	0.0	115	0.00
37	Benzo(b)fluoranthene	1.649	1.887	-14.4	116	0.00
38	Benzo(k)fluoranthene	1.653	1.862	-12.6	113	0.00
39 C	Benzo(a)pyrene	1.447	1.653	-14.2	116	0.00
40	Dibenzo(a,h)anthracene	1.673	1.866	-11.5	106	0.00
41	Benzo(g,h,i)perylene	1.671	1.857	-11.1	107	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN091620\
Data File : BN011817.D
Acq On : 16 Sep 2020 17:30
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 16 17:59:14 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 16 17:28:08 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0