

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091622\
 Data File : BN021795.D
 Acq On : 16 Sep 2022 19:01
 Operator : CG/JU
 Sample : PB147637BS
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020339

Quant Time: Sep 17 00:24:46 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 17 00:17:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.052	152	158686	20.000	ng/u1	0.00
20) Naphthalene-d8	10.881	136	711175	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.710	164	437197	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.457	188	836242	20.000	ng/u1	0.00
79) Chrysene-d12	21.651	240	584903	20.000	ng/u1	0.00
88) Perylene-d12	24.151	264	509277	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.358	96	3227	0.764	ng/uL	0.00
4) Pyridine-d5	3.799	84	8182	0.710	ng/u1	0.01
7) Phenol-d5	7.205	99	5474	0.380	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.375	67	3770	0.433	ng/u1	0.00
11) 2-Chlorophenol-d4	7.575	132	4855	0.464	ng/u1	0.00
15) 4-Methylphenol-d8	8.769	113	4432	0.392	ng/u1	0.00
21) Nitrobenzene-d5	9.228	128	2232	0.461	ng/u1	0.00
24) 2-Nitrophenol-d4	9.951	143	2210	0.511	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.522	165	20360	2.124	ng/u1	0.03
31) 4-Chloroaniline-d4	11.022	131	8730	0.536	ng/u1	0.00
46) Dimethylphthalate-d6	14.116	166	14285	0.478	ng/u1	0.00
49) Acenaphthylene-d8	14.398	160	15985	0.459	ng/u1	0.00
54) 4-Nitrophenol-d4	14.904	143	3011	0.506	ng/u1	0.00
60) Fluorene-d10	15.698	176	13090	0.503	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.822	200	4158	1.243	ng/u1	0.01
73) Anthracene-d10	17.557	188	20374	0.553	ng/u1	0.00
81) Pyrene-d10	19.851	212	20484	0.736	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.986	264	13546	0.561	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.393	88	49793	11.439	ng/uL	94
5) Pyridine	3.811	79	333556	28.779	ng/u1#	83
6) Benzaldehyde	7.181	77	309831	49.546	ng/u1	91
8) Phenol	7.234	94	483470	32.508	ng/u1	95
10) Bis(2-Chloroethyl)ether	7.463	93	379552	32.007	ng/u1	97
12) 2-Chlorophenol	7.605	128	383028	34.097	ng/u1	97
13) 2-Methylphenol	8.499	108	370867	32.820	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.587	45	499110	33.368	ng/u1	96
16) Acetophenone	8.887	105	566263	31.705	ng/u1	95
17) N-Nitroso-di-n-propyla...	8.869	70	299115	33.634	ng/u1	93
18) 4-Methylphenol	8.834	108	402928	32.460	ng/u1	99
19) Hexachloroethane	9.140	117	154239	34.448	ng/u1	88
22) Nitrobenzene	9.269	77	440682	36.907	ng/u1	97
23) Isophorone	9.799	82	863717	35.129	ng/u1	99
25) 2-Nitrophenol	9.987	139	220296	43.455	ng/u1	95
26) 2,4-Dimethylphenol	10.046	107	425896	33.926	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.287	93	525185	34.109	ng/u1	98
29) 2,4-Dichlorophenol	10.528	162	367094	36.554	ng/u1	96
30) Naphthalene	10.934	128	1214885	32.808	ng/u1	100
32) 4-Chloroaniline	11.046	127	495552	29.377	ng/u1	98
33) Hexachlorobutadiene	11.210	225	205711	35.557	ng/u1	97
34) Caprolactam	11.834	113	128130	33.356	ng/u1	84
35) 4-Chloro-3-methylphenol	12.169	107	410922	36.307	ng/u1	98
36) 2-Methylnaphthalene	12.540	142	820193	32.742	ng/u1	97


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091622\
 Data File : BN021795.D
 Acq On : 16 Sep 2022 19:01
 Operator : CG/JU
 Sample : PB147637BS
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020339

Quant Time: Sep 17 00:24:46 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 17 00:17:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.757	142	831829	33.042	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.904	216	386887	34.510	ng/ul	97
40) Hexachlorocyclopentadiene	12.881	237	199564	29.900	ng/ul	96
41) 2,4,6-Trichlorophenol	13.145	196	278452	40.030	ng/ul	99
42) 2,4,5-Trichlorophenol	13.216	196	308573	40.430	ng/ul	97
43) 1,1'-Biphenyl	13.545	154	1039187	32.525	ng/ul	99
44) 2-Chloronaphthalene	13.587	162	822422	33.623	ng/ul	98
45) 2-Nitroaniline	13.798	65	266145	43.382	ng/ul	89
47) Dimethylphthalate	14.163	163	1045223	34.196	ng/ul	99
48) 2,6-Dinitrotoluene	14.287	165	229508	43.767	ng/ul	97
50) Acenaphthylene	14.434	152	1285203	31.957	ng/ul	99
51) 3-Nitroaniline	14.616	138	243998	41.767	ng/ul	99
52) Acenaphthene	14.775	153	883615	34.004	ng/ul	98
53) 2,4-Dinitrophenol	14.822	184	118185	46.038	ng/ul	98
55) 4-Nitrophenol	14.922	109	155425	32.607	ng/ul	95
56) Dibenzofuran	15.104	168	1178391	31.562	ng/ul	99
57) 2,4-Dinitrotoluene	15.075	165	320569	42.092	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	15.334	232	245623	41.928	ng/ul	98
59) Diethylphthalate	15.522	149	1112367	36.087	ng/ul	98
61) Fluorene	15.757	166	951917	32.328	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.745	204	455495	33.557	ng/ul	96
63) 4-Nitroaniline	15.781	138	252224	43.022	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.834	198	159352	44.970	ng/ul	99
67) N-Nitrosodiphenylamine	15.963	169	850540	36.289	ng/ul	98
68) 4-Bromophenyl-phenylether	16.639	248	289415	38.914	ng/ul	96
69) Hexachlorobenzene	16.757	284	322237	36.398	ng/ul#	92
70) Atrazine	16.916	200	309223	37.284	ng/ul	96
71) Pentachlorophenol	17.104	266	205666	41.726	ng/ul	96
72) Phenanthrene	17.504	178	1520735	34.428	ng/ul	99
74) Anthracene	17.592	178	1513824	34.419	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.510	216	410786	38.595	ng/ul	97
76) Pentachlorobenzene	15.022	250	375366	37.022	ng/ul	87
77) Carbazole	17.869	167	1429091	34.383	ng/ul	99
78) Di-n-butylphthalate	18.422	149	1987465	42.412	ng/ul	100
80) Fluoranthene	19.516	202	1630708	48.603	ng/ul	95
82) Pyrene	19.880	202	1624213	46.447	ng/ul#	91
83) Butylbenzylphthalate	20.769	149	832807	62.922	ng/ul	93
84) 3,3'-Dichlorobenzidine	21.563	252	435117	39.016	ng/ul#	94
85) Benzo(a)anthracene	21.633	228	1307529	37.127	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.545	149	1227203	60.558	ng/ul	98
87) Chrysene	21.686	228	1234911	36.217	ng/ul	97
89) Di-n-octyl phthalate	22.504	149	1941132	65.069	ng/ul	100
90) Benzo(b)fluoranthene	23.386	252	1112328	35.900	ng/ul#	97
91) Benzo(k)fluoranthene	23.386	252	1112328	40.133	ng/ul	97
93) Benzo(a)pyrene	24.045	252	961733	32.442	ng/ul	95
94) Indeno(1,2,3-cd)pyrene	26.804	276	1126710	30.959	ng/ul	92
95) Dibenzo(a,h)anthracene	26.821	278	1017973	33.144	ng/ul	99
96) Benzo(g,h,i)perylene	27.621	276	987834	31.105	ng/ul#	91





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 11.439 ng/uL
 RT: 3.393 min Scan# 18
 Delta R.T. -0.000 min
 Lab File: BN021795.D
 Acq: 16 Sep 2022 19:01




Instrument :
 BNA_N
 ClientSampleId :
 SST020339


Tgt Ion: 88 Resp: 49793
 Ion Ratio Lower Upper
 88 100
 43 27.1 22.4 33.6
 58 69.3 61.0 91.4



#5
 Pyridine
 Concen: 28.779 ng/uL
 RT: 3.811 min Scan# 89
 Delta R.T. -0.000 min
 Lab File: BN021795.D
 Acq: 16 Sep 2022 19:01

Tgt Ion: 79 Resp: 333556
 Ion Ratio Lower Upper
 79 100
 52 64.2 65.0 97.4#
 51 32.4 31.6 47.4









#6
Benzaldehyde
Concen: 49.546 ng/ul
RT: 7.181 min Scan# 661
Delta R.T. 0.006 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01




Instrument :
BNA_N
ClientSampleId :
SSTD020339


Tgt Ion: 77 Resp: 309831
Ion Ratio Lower Upper
77 100
105 101.9 74.7 112.1
106 96.8 70.2 105.4



#8
Phenol
Concen: 32.508 ng/ul
RT: 7.234 min Scan# 671
Delta R.T. 0.006 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01

Tgt Ion: 94 Resp: 483470
Ion Ratio Lower Upper
94 100
65 27.4 23.6 35.4
66 36.0 31.2 46.8









#10
Bis(2-Chloroethyl)ether
Concen: 32.007 ng/ul
RT: 7.463 min Scan# 710
Delta R.T. -0.000 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01




Instrument :
BNA_N
ClientSampleId :
SSTD020339


Tgt Ion: 93 Resp: 379552
Ion Ratio Lower Upper
93 100
63 75.7 63.4 95.0
95 33.2 26.2 39.2



#12
2-Chlorophenol
Concen: 34.097 ng/ul
RT: 7.605 min Scan# 734
Delta R.T. -0.000 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01

Tgt Ion:128 Resp: 383028
Ion Ratio Lower Upper
128 100
64 45.7 38.6 57.8
130 32.9 25.0 37.4





#13

2-Methylphenol

Concen: 32.820 ng/ul

RT: 8.499 min Scan# 885

Delta R.T. 0.006 min

Lab File: BN021795.D

Acq: 16 Sep 2022 19:01

Instrument :

BNA_N

ClientSampleId :





SSTD020339

Tgt Ion:108 Resp: 370867

Ion Ratio Lower Upper

108 100

107 86.9 71.5 107.3



#14

2,2'-oxybis(1-Chloropropane)

Concen: 33.368 ng/ul

RT: 8.587 min Scan# 901

Delta R.T. 0.006 min

Lab File: BN021795.D

Acq: 16 Sep 2022 19:01




Tgt Ion: 45 Resp: 499110


Ion Ratio Lower Upper

45 100

77 15.5 10.9 16.3

79 12.2 8.6 13.0









#16
 Acetophenone
 Concen: 31.705 ng/ul
 RT: 8.887 min Scan# 91
 Delta R.T. 0.006 min
 Lab File: BN021795.D
 Acq: 16 Sep 2022 19:01




Instrument :
 BNA_N
 ClientSampleId :
 SST020339


Tgt Ion:105 Resp: 566263
 Ion Ratio Lower Upper
 105 100
 77 75.6 63.5 95.3
 51 24.8 23.5 35.3



#17
 N-Nitroso-di-n-propylamine
 Concen: 33.634 ng/ul
 RT: 8.869 min Scan# 949
 Delta R.T. 0.006 min
 Lab File: BN021795.D
 Acq: 16 Sep 2022 19:01

Tgt Ion: 70 Resp: 299115
 Ion Ratio Lower Upper
 70 100
 42 52.9 47.8 71.6
 101 10.0 7.8 11.6
 130 21.1 15.8 23.8





#18

4-Methylphenol

Concen: 32.460 ng/ul

RT: 8.834 min Scan# 942

Delta R.T. 0.006 min

Lab File: BN021795.D

Acq: 16 Sep 2022 19:01

Instrument :

BNA_N

ClientSampleId :





SSTD020339

Tgt Ion:108 Resp: 402928

Ion Ratio Lower Upper

108 100

107 111.2 88.0 132.0



#19

Hexachloroethane

Concen: 34.448 ng/ul

RT: 9.140 min Scan# 995

Delta R.T. -0.000 min

Lab File: BN021795.D

Acq: 16 Sep 2022 19:01




Tgt Ion:117 Resp: 154239


Ion Ratio Lower Upper

117 100

201 90.5 83.2 124.8

199 55.9 50.2 75.4









#22
Nitrobenzene
Concen: 36.907 ng/ul
RT: 9.269 min Scan# 1107
Delta R.T. -0.000 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01




Instrument :
BNA_N
ClientSampleId :
SSTD020339


Tgt Ion: 77 Resp: 440682
Ion Ratio Lower Upper
77 100
123 44.9 34.5 51.7
65 12.9 11.1 16.7



#23
Isophorone
Concen: 35.129 ng/ul
RT: 9.799 min Scan# 1107
Delta R.T. 0.006 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01

Tgt Ion: 82 Resp: 863717
Ion Ratio Lower Upper
82 100
95 7.1 5.8 8.8
138 17.7 13.8 20.8





#25

2-Nitrophenol

Concen: 43.455 ng/ul

RT: 9.987 min Scan# 1139

Delta R.T. -0.000 min

Lab File: BN021795.D

Acq: 16 Sep 2022 19:01

Instrument :

BN021795.D

ClientSampleId :

SSTD020339





Tgt Ion:139 Resp: 220296

Ion Ratio Lower Upper

139 100

65 50.2 43.7 65.5

109 31.0 25.8 38.6



#26

2,4-Dimethylphenol

Concen: 33.926 ng/ul

RT: 10.046 min Scan# 1149

Delta R.T. 0.006 min

Lab File: BN021795.D

Acq: 16 Sep 2022 19:01




Tgt Ion:107 Resp: 425896


Ion Ratio Lower Upper

107 100

121 47.6 38.3 57.5

122 82.7 66.3 99.5









#27
Bis(2-Chloroethoxy)methane
Concen: 34.109 ng/ul
RT: 10.287 min Scan# 1189
Delta R.T. 0.006 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01




Instrument :
BNA_N
ClientSampleId :
SSTD020339


Tgt Ion: 93 Resp: 525185
Ion Ratio Lower Upper
93 100
95 33.0 25.4 38.0
123 11.4 8.3 12.5



#29
2,4-Dichlorophenol
Concen: 36.554 ng/ul
RT: 10.528 min Scan# 1231
Delta R.T. 0.006 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01

Tgt Ion: 162 Resp: 367094
Ion Ratio Lower Upper
162 100
164 62.2 52.9 79.3
98 36.1 29.8 44.8









#30
Naphthalene
Concen: 32.808 ng/ul
RT: 10.934 min Scan# 1300
Delta R.T. 0.006 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01




Instrument :
BNA_N
ClientSampleId :
SSTD020339


Tgt Ion:128 Resp: 1214885
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 13.0 10.6 16.0



#32
4-Chloroaniline
Concen: 29.377 ng/ul
RT: 11.046 min Scan# 1319
Delta R.T. 0.006 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01

Tgt Ion:127 Resp: 495552
Ion Ratio Lower Upper
127 100
129 31.3 26.1 39.1









#33
Hexachlorobutadiene
Concen: 35.557 ng/ul
RT: 11.210 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01




Instrument :
BNA_N
ClientSampleId :
SSTD020339


Tgt Ion:225 Resp: 205711
Ion Ratio Lower Upper
225 100
223 65.8 51.8 77.6
227 66.9 50.6 76.0



#34
Caprolactam
Concen: 33.356 ng/ul
RT: 11.834 min Scan# 1453
Delta R.T. 0.018 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01

Tgt Ion:113 Resp: 128130
Ion Ratio Lower Upper
113 100
55 155.9 145.5 218.3
56 117.6 106.5 159.7









#35
4-Chloro-3-methylphenol
Concen: 36.307 ng/ul
RT: 12.169 min Scan# 1509
Delta R.T. 0.006 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01




Instrument :
BNA_N
ClientSampleId :
SSTD020339


Tgt Ion:107 Resp: 410922
Ion Ratio Lower Upper
107 100
144 25.1 20.6 31.0
142 80.0 62.2 93.4



#36
2-Methylnaphthalene
Concen: 32.742 ng/ul
RT: 12.540 min Scan# 1573
Delta R.T. -0.000 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01

Tgt Ion:142 Resp: 820193
Ion Ratio Lower Upper
142 100
141 91.4 70.6 105.8









#37
1-Methylnaphthalene
Concen: 33.042 ng/ul
RT: 12.757 min Scan# 1610
Delta R.T. -0.000 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01




Instrument :
BNA_N
ClientSampleId :
SSTD020339


Tgt Ion:142 Resp: 831829
Ion Ratio Lower Upper
142 100
141 88.3 70.9 106.3
116 3.7 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 34.510 ng/ul
RT: 12.904 min Scan# 1635
Delta R.T. 0.006 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01

Tgt Ion:216 Resp: 386887
Ion Ratio Lower Upper
216 100
214 75.7 63.4 95.0
179 21.6 17.8 26.8
108 20.1 15.1 22.7









#40
Hexachlorocyclopentadiene
Concen: 29.900 ng/ul
RT: 12.881 min Scan# 1631
Delta R.T. 0.006 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01




Instrument :
BNA_N
ClientSampleId :
SSTD020339


Tgt Ion:237 Resp: 199564
Ion Ratio Lower Upper
237 100
235 64.5 48.8 73.2
272 12.0 9.3 13.9



#41
2,4,6-Trichlorophenol
Concen: 40.030 ng/ul
RT: 13.145 min Scan# 1676
Delta R.T. 0.006 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01

Tgt Ion:196 Resp: 278452
Ion Ratio Lower Upper
196 100
198 98.9 79.8 119.8
200 29.7 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 40.430 ng/ul
RT: 13.216 min Scan# 1
Delta R.T. 0.006 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01




Instrument :
BNA_N
ClientSampleId :
SSTD020339


Tgt Ion:196 Resp: 308573
Ion Ratio Lower Upper
196 100
198 93.8 77.6 116.4
200 31.0 24.4 36.6



#43
1,1'-Biphenyl
Concen: 32.525 ng/ul
RT: 13.545 min Scan# 1744
Delta R.T. 0.006 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01

Tgt Ion:154 Resp: 1039187
Ion Ratio Lower Upper
154 100
153 38.9 31.5 47.3
76 13.8 10.5 15.7









#44
2-Chloronaphthalene
Concen: 33.623 ng/ul
RT: 13.587 min Scan# 1751
Delta R.T. -0.000 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01




Instrument :
BNA_N
ClientSampleId :
SSTD020339


Tgt Ion:162 Resp: 822422
Ion Ratio Lower Upper
162 100
164 31.8 25.8 38.6
127 41.0 31.4 47.2



#45
2-Nitroaniline
Concen: 43.382 ng/ul
RT: 13.798 min Scan# 1787
Delta R.T. 0.006 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01

Tgt Ion: 65 Resp: 266145
Ion Ratio Lower Upper
65 100
92 71.0 54.1 81.1
138 112.2 77.0 115.4









#47
Dimethylphthalate
Concen: 34.196 ng/ul
RT: 14.163 min Scan# 1849
Delta R.T. -0.000 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01




Instrument :
BNA_N
ClientSampleId :
SSTD020339


Tgt Ion:163 Resp: 1045223
Ion Ratio Lower Upper
163 100
194 4.1 3.4 5.0
164 10.3 7.7 11.5



#48
2,6-Dinitrotoluene
Concen: 43.767 ng/ul
RT: 14.287 min Scan# 1870
Delta R.T. -0.000 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01

Tgt Ion:165 Resp: 229508
Ion Ratio Lower Upper
165 100
89 60.5 50.6 75.8
121 21.7 17.9 26.9









#50
Acenaphthylene
Concen: 31.957 ng/ul
RT: 14.434 min Scan# 1894
Delta R.T. 0.006 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01




Instrument :
BNA_N
ClientSampleId :
SSTD020339


Tgt Ion:152 Resp: 1285203
Ion Ratio Lower Upper
152 100
151 19.8 16.0 24.0
153 12.9 10.7 16.1



#51
3-Nitroaniline
Concen: 41.767 ng/ul
RT: 14.616 min Scan# 1926
Delta R.T. -0.000 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01

Tgt Ion:138 Resp: 243998
Ion Ratio Lower Upper
138 100
108 10.9 8.6 12.8
92 117.6 92.9 139.3









#52
Acenaphthene
Concen: 34.004 ng/ul
RT: 14.775 min Scan# 1953
Delta R.T. 0.006 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01




Instrument :
BNA_N
ClientSampleId :
SSTD020339


Tgt Ion:153 Resp: 883615
Ion Ratio Lower Upper
153 100
152 48.1 38.4 57.6
154 89.6 69.7 104.5



#53
2,4-Dinitrophenol
Concen: 46.038 ng/ul
RT: 14.822 min Scan# 1961
Delta R.T. 0.006 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01

Tgt Ion:184 Resp: 118185
Ion Ratio Lower Upper
184 100
63 67.6 55.9 83.9
154 66.0 53.1 79.7









#55
4-Nitrophenol
Concen: 32.607 ng/ul
RT: 14.922 min Scan# 1977
Delta R.T. 0.006 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01




Instrument :
BNA_N
ClientSampleId :
SSTD020339


Tgt Ion:109 Resp: 155425
Ion Ratio Lower Upper
109 100
139 130.2 100.1 150.1
65 125.0 104.6 157.0



#56
Dibenzofuran
Concen: 31.562 ng/ul
RT: 15.104 min Scan# 2009
Delta R.T. -0.000 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01

Tgt Ion:168 Resp: 1178391
Ion Ratio Lower Upper
168 100
139 39.6 31.1 46.7









#57
2,4-Dinitrotoluene
Concen: 42.092 ng/ul
RT: 15.075 min Scan# 2004
Delta R.T. 0.006 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01




Instrument :
BNA_N
ClientSampleId :
SSTD020339


Tgt Ion:165 Resp: 320569
Ion Ratio Lower Upper
165 100
63 41.8 36.3 54.5
182 2.7 2.3 3.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 41.928 ng/ul
RT: 15.334 min Scan# 2048
Delta R.T. 0.006 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01

Tgt Ion:232 Resp: 245623
Ion Ratio Lower Upper
232 100
131 49.2 38.5 57.7
130 2.7 2.2 3.2
166 33.6 25.8 38.8









#59
Diethylphthalate
Concen: 36.087 ng/ul
RT: 15.522 min Scan# 2079
Delta R.T. 0.006 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01




Instrument :
BNA_N
ClientSampleId :
SSTD020339


Tgt Ion:149 Resp: 1112367
Ion Ratio Lower Upper
149 100
177 21.3 17.8 26.8
150 12.0 10.5 15.7



#61
Fluorene
Concen: 32.328 ng/ul
RT: 15.757 min Scan# 2120
Delta R.T. 0.006 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01

Tgt Ion:166 Resp: 951917
Ion Ratio Lower Upper
166 100
165 97.0 77.7 116.5
167 13.0 11.1 16.7









#62
4-Chlorophenyl-phenylether
Concen: 33.557 ng/ul
RT: 15.745 min Scan# 2118
Delta R.T. -0.000 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01




Instrument :
BNA_N
ClientSampleId :
SSTD020339


Tgt Ion:204 Resp: 455495
Ion Ratio Lower Upper
204 100
206 33.3 26.7 40.1
141 67.4 50.2 75.2



#63
4-Nitroaniline
Concen: 43.022 ng/ul
RT: 15.781 min Scan# 2124
Delta R.T. 0.006 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01

Tgt Ion:138 Resp: 252224
Ion Ratio Lower Upper
138 100
92 53.2 41.0 61.6
108 81.2 65.3 97.9









#66
4,6-Dinitro-2-methylphenol
Concen: 44.970 ng/ul
RT: 15.834 min Scan# 2133
Delta R.T. 0.006 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01




Instrument :
BNA_N
ClientSampleId :
SSTD020339


Tgt Ion:198 Resp: 159352
Ion Ratio Lower Upper
198 100
182 3.6 3.0 4.4
77 25.8 20.1 30.1



#67
N-Nitrosodiphenylamine
Concen: 36.289 ng/ul
RT: 15.963 min Scan# 2155
Delta R.T. 0.006 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01

Tgt Ion:169 Resp: 850540
Ion Ratio Lower Upper
169 100
168 64.0 52.4 78.6
167 35.2 27.5 41.3









#68
4-Bromophenyl-phenylether
Concen: 38.914 ng/ul
RT: 16.639 min Scan# 2170
Delta R.T. -0.000 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01




Instrument :
BNA_N
ClientSampleId :
SSTD020339


Tgt Ion:248 Resp: 289415
Ion Ratio Lower Upper
248 100
250 94.9 77.8 116.8
141 83.5 62.6 93.8



#69
Hexachlorobenzene
Concen: 36.398 ng/ul
RT: 16.757 min Scan# 2290
Delta R.T. -0.000 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01

Tgt Ion:284 Resp: 322237
Ion Ratio Lower Upper
284 100
142 43.7 27.8 41.8#
249 31.9 25.7 38.5





#70

Atrazine

Concen: 37.284 ng/ul

RT: 16.916 min Scan# 21

Delta R.T. 0.006 min

Lab File: BN021795.D

Acq: 16 Sep 2022 19:01

Instrument :

BNA_N

ClientSampleId :

SSTD020339





Tgt Ion:200 Resp: 309223

Ion Ratio Lower Upper

200 100

173 29.8 20.5 30.7

215 47.7 38.9 58.3



#71

Pentachlorophenol

Concen: 41.726 ng/ul

RT: 17.104 min Scan# 2349

Delta R.T. 0.006 min

Lab File: BN021795.D

Acq: 16 Sep 2022 19:01




Tgt Ion:266 Resp: 205666


Ion Ratio Lower Upper

266 100

264 65.5 50.2 75.2

268 58.4 49.3 73.9









#72
Phenanthrene
Concen: 34.428 ng/ul
RT: 17.504 min Scan# 2416
Delta R.T. 0.006 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01




Instrument :
BNA_N
ClientSampleId :
SSTD020339


Tgt Ion:178 Resp: 1520735
Ion Ratio Lower Upper
178 100
179 15.4 12.2 18.4
176 18.1 15.2 22.8



#74
Anthracene
Concen: 34.419 ng/ul
RT: 17.592 min Scan# 2432
Delta R.T. -0.000 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01

Tgt Ion:178 Resp: 1513824
Ion Ratio Lower Upper
178 100
179 15.7 13.0 19.4
176 19.3 15.5 23.3





#75

1,2,3,4-Tetrachlorobenzene

Concen: 38.595 ng/uL

RT: 13.510 min Scan# 1738

Delta R.T. 0.006 min

Lab File: BN021795.D

Acq: 16 Sep 2022 19:01

Instrument :

BNA_N

ClientSampleId :

SSTD020339





Tgt Ion:216 Resp: 410786

Ion Ratio Lower Upper

216 100

214 81.4 63.2 94.8

218 50.4 37.9 56.9



#76

Pentachlorobenzene

Concen: 37.022 ng/uL

RT: 15.022 min Scan# 1995

Delta R.T. -0.000 min

Lab File: BN021795.D

Acq: 16 Sep 2022 19:01




Tgt Ion:250 Resp: 375366


Ion Ratio Lower Upper

250 100

248 55.6 53.0 79.4

252 55.9 53.4 80.2









#77
 Carbazole
 Concen: 34.383 ng/ul
 RT: 17.869 min Scan# 2478
 Delta R.T. 0.006 min
 Lab File: BN021795.D
 Acq: 16 Sep 2022 19:01




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020339


Tgt Ion:167 Resp: 1429091
 Ion Ratio Lower Upper
 167 100
 166 20.4 16.1 24.1
 139 13.3 10.2 15.2



#78
 Di-n-butylphthalate
 Concen: 42.412 ng/ul
 RT: 18.422 min Scan# 2573
 Delta R.T. 0.006 min
 Lab File: BN021795.D
 Acq: 16 Sep 2022 19:01

Tgt Ion:149 Resp: 1987465
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.1 10.7
 104 5.0 4.1 6.1









#80
Fluoranthene
Concen: 48.603 ng/u1
RT: 19.516 min Scan# 21
Delta R.T. 0.006 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01




Instrument :
BNA_N
ClientSampleId :
SSTD020339


Tgt Ion:202 Resp: 1630708
Ion Ratio Lower Upper
202 100
101 13.3 8.9 13.3
100 10.1 6.7 10.1



#82
Pyrene
Concen: 46.447 ng/u1
RT: 19.880 min Scan# 2821
Delta R.T. 0.006 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01

Tgt Ion:202 Resp: 1624213
Ion Ratio Lower Upper
202 100
101 15.9 9.7 14.5#
100 13.1 8.0 12.0#





#83

Butylbenzylphthalate

Concen: 62.922 ng/ul

RT: 20.769 min Scan# 2972

Delta R.T. -0.000 min

Lab File: BN021795.D

Acq: 16 Sep 2022 19:01

Instrument :

BNA_N

ClientSampleId :

SSTD020339





Tgt Ion:149 Resp: 832807

Ion Ratio Lower Upper

149 100

91 79.5 58.8 88.2

206 19.9 18.1 27.1



#84

3,3'-Dichlorobenzidine

Concen: 39.016 ng/ul

RT: 21.563 min Scan# 3107

Delta R.T. -0.000 min

Lab File: BN021795.D

Acq: 16 Sep 2022 19:01




Tgt Ion:252 Resp: 435117


Ion Ratio Lower Upper

252 100

254 67.8 50.8 76.2

126 14.2 8.2 12.4#





#85

Benzo(a)anthracene

Concen: 37.127 ng/ul

RT: 21.633 min Scan# 3119

Delta R.T. 0.006 min

Lab File: BN021795.D

Acq: 16 Sep 2022 19:01

Instrument :

BNA_N

ClientSampleId :

SSTD020339





Tgt Ion:228 Resp: 1307529

Ion Ratio Lower Upper

228 100

229 19.1 16.3 24.5

226 27.0 23.2 34.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 60.558 ng/ul

RT: 21.545 min Scan# 3104

Delta R.T. -0.000 min

Lab File: BN021795.D

Acq: 16 Sep 2022 19:01




Tgt Ion:149 Resp: 1227203


Ion Ratio Lower Upper

149 100

167 27.0 22.5 33.7

279 3.6 3.5 5.3





#87

Chrysene

Concen: 36.217 ng/ul

RT: 21.686 min Scan# 31

Delta R.T. -0.000 min

Lab File: BN021795.D

Acq: 16 Sep 2022 19:01

Instrument :

BNA_N

ClientSampleId :

SSTD020339





Tgt Ion:228 Resp: 1234911

Ion Ratio Lower Upper

228 100

226 28.1 24.2 36.2

229 19.7 16.2 24.2



#89

Di-n-octyl phthalate

Concen: 65.069 ng/ul




RT: 22.504 min Scan# 3267


Delta R.T. 0.006 min

Lab File: BN021795.D

Acq: 16 Sep 2022 19:01

Tgt Ion:149 Resp: 1941132









#90
Benzo(b)fluoranthene
Concen: 35.900 ng/ul
RT: 23.386 min Scan# 3417
Delta R.T. -0.000 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01




Instrument :
BNA_N
ClientSampleId :
SSTD020339


Tgt Ion:252 Resp: 1112328
Ion Ratio Lower Upper
252 100
253 21.5 16.5 24.7
125 12.0 7.8 11.8#



#91
Benzo(k)fluoranthene
Concen: 40.133 ng/ul
RT: 23.386 min Scan# 3417
Delta R.T. -0.047 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01

Tgt Ion:252 Resp: 1112328
Ion Ratio Lower Upper
252 100
253 21.5 18.2 27.4
125 12.0 8.2 12.4









#93
Benzo(a)pyrene
Concen: 32.442 ng/ul
RT: 24.045 min Scan# 31
Delta R.T. -0.000 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01




Instrument :
BNA_N
ClientSampleId :
SSTD020339


Tgt Ion:252 Resp: 961733
Ion Ratio Lower Upper
252 100
253 20.9 18.7 28.1
125 13.3 9.2 13.8



#94
Indeno(1,2,3-cd)pyrene
Concen: 30.959 ng/ul
RT: 26.804 min Scan# 3998
Delta R.T. 0.006 min
Lab File: BN021795.D
Acq: 16 Sep 2022 19:01

Tgt Ion:276 Resp: 1126710
Ion Ratio Lower Upper
276 100
138 27.7 19.5 29.3
277 21.7 21.4 32.0





#95

Dibenzo(a,h)anthracene

Concen: 33.144 ng/ul

RT: 26.821 min Scan# 4100

Delta R.T. 0.012 min

Lab File: BN021795.D

Acq: 16 Sep 2022 19:01

Instrument :

BNA_N

ClientSampleId :

SSTD020339





Tgt Ion:278 Resp: 1017973

Ion Ratio Lower Upper

278 100

139 18.4 13.8 20.8

279 22.6 18.3 27.5



#96

Benzo(g,h,i)perylene

Concen: 31.105 ng/ul

RT: 27.621 min Scan# 4137

Delta R.T. 0.006 min

Lab File: BN021795.D

Acq: 16 Sep 2022 19:01

Tgt Ion:276 Resp: 987834

Ion Ratio Lower Upper

276 100

138 27.0 17.0 25.6#

277 25.5 17.8 26.6

