



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091622\
 Data File : BN021783.D
 Acq On : 16 Sep 2022 11:00
 Operator : CG/JU
 Sample : DFTPP251
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP251

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM091522.M
 Title : SVOA CALIBRATION
 Last Update : Fri Sep 16 02:39:54 2022



AutoFind: Scans 2481, 2482, 2483; Background Corrected with Scan 2475

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.2	138189	PASS
68	69	0.00	2	1.8	2605	PASS
69	198	0.00	100	41.9	147656	PASS
70	69	0.00	2	0.5	770	PASS
127	198	10	80	53.1	187221	PASS
197	198	0.00	2	0.5	1611	PASS
198	198	100	100	100.0	352427	PASS
199	198	5	9	6.7	23531	PASS
275	198	10	60	24.9	87632	PASS
365	198	1	100	3.4	12012	PASS
441	198	0.01	100	11.7	41397	PASS
442	198	50	100	65.0	228949	PASS
443	442	15	24	21.5	49160	PASS

m/z	Abundance
39.00	161.0
39.90	906.0
41.00	152.0
44.00	673.0
48.90	234.0
73.00	955.0
74.90	169.0
76.90	261.0
78.10	172.0
83.80	199.0
91.00	161.0
96.00	341.0
132.90	183.0
147.10	339.0
191.00	259.0
193.00	257.0
207.00	2741.0
208.00	461.0
209.00	266.0
221.00	173.0
281.00	517.0

Instrument :
BNA_N
ClientSampleId :
DFTPP251

m/z	Abundance
36.90	643.0
38.00	1582.0
39.00	9613.0
40.00	1100.0
41.10	340.0
42.90	165.0
43.90	728.0
44.90	300.0
49.00	1014.0
50.00	34056.0
51.00	135360.0
52.00	6973.0
53.00	239.0
55.00	693.0
56.00	4160.0
57.00	9339.0
58.00	385.0
59.00	164.0
61.00	1499.0
62.00	1885.0
63.00	5475.0
64.00	894.0
65.00	2711.0
66.00	327.0
67.00	174.0
68.00	2412.0
69.00	141760.0
70.00	690.0
73.00	1786.0
74.00	14427.0
75.00	23744.0
76.00	8408.0
77.00	170560.0
78.00	12335.0
79.00	10307.0
80.00	8154.0
81.00	11564.0
82.00	2933.0
83.00	2568.0
83.90	556.0
85.00	1845.0
86.00	3320.0
87.00	1543.0
88.00	581.0
89.00	337.0
91.00	2381.0
92.00	3051.0
93.00	18712.0
94.00	1167.0
94.90	337.0
96.00	1117.0
97.10	335.0
98.00	14421.0
99.00	13526.0
100.00	1052.0
101.00	6851.0
101.90	338.0
103.00	2180.0
104.00	3823.0
105.00	3815.0
106.00	1206.0
107.00	49160.0
108.00	7600.0
109.00	1311.0
110.00	86296.0
111.00	14898.0
111.90	1628.0
112.90	568.0
115.00	234.0
116.00	2354.0
117.00	42312.0
118.00	3750.0
118.90	587.0
119.90	606.0
121.00	267.0
121.90	3060.0
123.00	4447.0
123.90	2218.0
124.90	2006.0
126.00	546.0
127.00	166528.0
128.00	12608.0
129.00	63136.0
130.00	5602.0
131.00	965.0
132.00	616.0
132.90	345.0
134.00	1813.0
135.00	5056.0
136.00	2113.0
137.00	3263.0
138.00	547.0

Instrument :
BNA_N
ClientSampleId :
DFTPP251

138.90	261.0
140.00	804.0
141.00	7517.0
141.90	2745.0
143.00	1826.0
144.00	499.0
144.90	442.0
145.90	1507.0
147.00	4245.0
148.00	10580.0
149.00	2577.0
150.00	837.0
151.10	971.0
152.00	611.0
153.00	2300.0
154.00	1706.0
155.00	4722.0
156.00	6769.0
157.00	1233.0
157.90	1463.0
158.90	1051.0
160.00	2378.0
161.00	3645.0
162.00	1193.0
163.00	389.0
164.00	353.0
165.00	2879.0
166.00	2322.0
167.00	17840.0
168.00	13609.0
169.00	1873.0
170.00	542.0
171.00	603.0
172.00	1383.0
173.00	1700.0
174.00	3541.0
175.00	6793.0
176.00	1711.0
177.00	2888.0
177.90	893.0
179.00	11827.0
180.00	9700.0
181.00	3778.0
182.00	533.0
182.90	308.0
184.00	847.0
185.00	5526.0
186.00	42224.0
187.00	12530.0
188.10	1425.0
188.90	2607.0
189.90	378.0
191.00	1453.0
192.00	3749.0
193.00	4261.0
194.00	847.0
195.00	421.0
196.00	9100.0
196.80	1343.0
198.00	301056.0
199.00	18984.0
200.00	1714.0
201.40	1568.0
203.00	1906.0
204.00	11031.0
205.00	17504.0
206.10	76320.0
207.10	11659.0
208.00	2944.0
209.00	1183.0
210.20	1328.0
211.00	3207.0
211.80	637.0
213.00	221.0
215.00	727.0
216.00	1842.0
217.00	21216.0
218.00	2568.0
219.10	278.0
221.00	11679.0
221.70	2928.0
223.00	4570.0
224.00	42672.0
225.00	11122.0
226.00	1341.0
227.00	17504.0
228.00	2526.0
229.00	3530.0
229.90	442.0
231.00	1412.0
232.00	231.0
233.00	358.0
234.00	1184.0
235.00	1415.0
235.90	992.0
237.00	2595.0

Instrument :
BNA_N
ClientSampleId :
DFTPP251

238.00	265.0
239.00	601.0
240.00	620.0
241.00	782.0
242.00	2218.0
243.00	2355.0
244.00	34760.0
245.00	4436.0
246.00	5326.0
247.00	1452.0
248.00	270.0
249.00	1205.0
249.90	226.0
250.80	199.0
252.00	334.0
252.90	513.0
254.10	1131.0
255.00	157120.0
256.00	25296.0
257.00	2415.0
258.00	9042.0
259.00	1639.0
259.90	305.0
261.00	236.0
264.10	290.0
265.00	3542.0
265.90	679.0
270.00	254.0
270.90	423.0
272.00	455.0
273.00	4788.0
274.00	10481.0
275.00	67904.0
276.00	9910.0
277.00	5966.0
278.00	1245.0
278.90	318.0
281.00	623.0
281.90	258.0
283.10	795.0
284.00	389.0
285.00	918.0
286.10	251.0
289.00	216.0
289.90	208.0
291.00	206.0
291.80	320.0
293.00	1499.0
294.10	627.0
295.00	515.0
296.00	17424.0
297.00	2996.0
297.90	267.0
301.90	315.0
303.00	2327.0
304.00	755.0
307.90	265.0
309.10	187.0
310.00	254.0
313.00	155.0
314.00	906.0
315.00	2115.0
316.00	1112.0
316.90	199.0
321.00	637.0
322.00	256.0
323.10	5828.0
324.00	869.0
325.10	157.0
326.30	161.0
326.90	1237.0
328.00	624.0
329.00	177.0
331.90	393.0
332.90	528.0
334.10	3805.0
335.00	1053.0
336.00	198.0
341.00	722.0
342.00	249.0
345.90	1268.0
347.00	288.0
352.00	1676.0
353.10	1314.0
354.00	1767.0
355.10	395.0
365.00	8371.0
365.90	1221.0
370.00	169.0
371.00	385.0
372.00	2506.0
373.00	800.0
383.00	988.0
383.90	205.0
390.00	399.0
391.00	309.0

Instrument :
BNA_N
ClientSampleId :
DFTPP251

392.10	199.0
400.90	205.0
402.00	1201.0
403.00	1828.0
404.00	786.0
405.00	297.0
421.00	1337.0
422.00	1446.0
423.10	10706.0
424.10	4388.0
425.10	711.0
440.00	150.0
441.20	23256.0
442.10	172736.0
443.10	32736.0
444.10	3015.0
445.00	185.0

Instrument :
BNA_N
ClientSampleId :
DFTPP251

m/z	Abundance
37.00	657.0
38.00	1928.0
39.00	11714.0
40.00	1140.0
40.90	394.0
43.00	175.0
44.00	730.0
45.00	374.0
48.00	153.0
49.10	1278.0
50.00	41760.0
51.00	159232.0
52.00	8343.0
53.00	370.0
54.90	812.0
56.00	4706.0
57.00	10251.0
58.00	519.0
59.00	217.0
61.00	2038.0
62.00	2143.0
63.00	6625.0
64.00	748.0
65.00	3571.0
66.00	221.0
66.90	265.0
68.00	3067.0
69.00	170944.0
70.00	930.0
73.00	1884.0
74.00	17752.0
75.00	28464.0
76.00	10054.0
77.10	202944.0
78.00	14436.0
79.00	13613.0
80.00	9467.0
81.00	14066.0
82.00	3445.0
83.00	3249.0
84.00	539.0
85.00	2713.0
86.00	4182.0
87.00	2268.0
88.00	686.0
89.00	338.0
91.00	3010.0
92.00	3499.0
93.00	23792.0
94.00	1662.0
94.90	494.0
96.00	1153.0
96.90	323.0
98.00	17656.0
99.00	16103.0
100.00	1374.0
101.00	7375.0
102.00	472.0
102.90	2528.0
104.00	4777.0
105.00	4406.0
106.00	1408.0
107.00	62848.0
108.00	9716.0
109.00	1778.0
110.00	111416.0
111.00	17768.0
112.00	2174.0
113.00	662.0
114.90	313.0
116.00	3628.0
117.00	55560.0
118.00	4402.0
119.00	643.0
119.90	813.0
120.80	253.0
122.00	4106.0
123.00	6360.0
124.00	2782.0
125.00	2391.0
126.00	738.0
127.00	218624.0
128.00	16201.0
129.00	81752.0
130.00	6819.0
131.00	1493.0
132.00	753.0
132.90	517.0
134.00	2198.0
134.90	6818.0
136.00	2770.0
137.00	3892.0

Instrument :
BNA_N
ClientSampleId :
DFTPP251

138.00	791.0
138.90	468.0
140.00	1185.0
141.00	10199.0
142.00	3248.0
143.00	2248.0
144.00	736.0
145.00	624.0
146.00	1880.0
147.00	5840.0
148.00	14021.0
149.00	3156.0
150.00	775.0
151.00	1207.0
151.80	950.0
153.00	3387.0
154.00	2446.0
155.00	6083.0
156.00	8984.0
157.00	1563.0
158.00	2056.0
159.00	1577.0
160.00	3084.0
161.00	5205.0
162.00	1504.0
163.00	343.0
164.00	624.0
165.00	3998.0
166.00	3299.0
167.00	24040.0
168.00	14814.0
169.00	2186.0
169.90	608.0
171.00	862.0
171.90	1639.0
173.00	2349.0
174.00	4504.0
175.00	8858.0
176.00	2319.0
177.00	4085.0
178.00	1338.0
179.00	15899.0
180.00	11497.0
181.00	5259.0
182.00	743.0
183.00	450.0
184.00	1170.0
185.00	7789.0
186.00	60728.0
187.00	16356.0
188.00	2084.0
189.00	3521.0
190.00	529.0
191.00	1912.0
192.00	5298.0
193.00	6057.0
194.00	1143.0
195.00	674.0
196.00	11348.0
196.90	1761.0
197.90	408768.0
199.00	28496.0
200.00	2025.0
201.50	1764.0
203.00	2937.0
204.00	14963.0
205.00	24144.0
206.00	101624.0
207.00	16512.0
208.00	3436.0
209.00	1563.0
210.10	1839.0
211.00	4255.0
213.00	372.0
214.90	1322.0
216.00	2109.0
217.00	26320.0
218.00	3726.0
219.00	390.0
221.10	17064.0
221.80	3940.0
223.00	6065.0
224.00	60792.0
225.00	15792.0
226.10	1886.0
227.00	24488.0
228.00	3393.0
229.00	5022.0
229.90	663.0
231.00	2070.0
232.00	388.0
233.00	447.0
234.00	1645.0
235.00	1832.0
236.00	1283.0
237.00	2533.0

Instrument :
BNA_N
ClientSampleId :
DFTPP251

238.10	445.0
239.00	1082.0
240.00	891.0
241.00	1364.0
242.00	3496.0
243.00	3479.0
244.00	47744.0
245.10	6275.0
246.00	9542.0
247.00	1646.0
248.00	390.0
249.10	1665.0
249.90	206.0
251.00	421.0
252.00	402.0
252.90	1082.0
254.00	1192.0
255.00	220800.0
256.00	31680.0
257.00	3041.0
258.00	13799.0
259.00	2027.0
260.00	333.0
261.00	415.0
263.00	179.0
263.90	261.0
265.00	5590.0
266.00	836.0
267.00	211.0
269.90	284.0
270.90	512.0
272.10	802.0
273.00	6627.0
274.00	17640.0
275.00	99216.0
276.00	13938.0
277.00	9133.0
278.00	1592.0
279.00	347.0
281.00	431.0
282.00	342.0
283.00	951.0
284.10	736.0
285.00	1293.0
286.00	373.0
288.90	303.0
290.00	208.0
290.90	262.0
292.00	312.0
293.00	2200.0
294.00	712.0
295.00	605.0
296.00	30064.0
297.10	3915.0
298.00	285.0
301.00	317.0
302.10	541.0
303.10	3513.0
304.00	740.0
305.00	183.0
307.90	391.0
309.00	282.0
310.00	410.0
313.00	239.0
314.00	1622.0
315.00	3358.0
316.00	1871.0
316.90	299.0
321.00	1050.0
322.10	487.0
323.00	9260.0
324.10	1721.0
325.10	162.0
325.80	222.0
327.00	2306.0
327.90	1132.0
328.90	234.0
332.00	797.0
333.00	961.0
334.00	5854.0
335.00	1765.0
336.00	209.0
340.90	1056.0
342.10	225.0
346.00	2043.0
346.80	345.0
350.90	205.0
352.00	3204.0
353.00	2035.0
354.00	3511.0
355.00	647.0
359.00	260.0
365.00	14452.0
366.00	1899.0
369.90	262.0
371.10	970.0

Instrument :
BNA_N
ClientSampleId :
DFTPP251

m/z	Abundance
35.00	177.0
37.00	489.0
38.00	1372.0
39.00	8619.0
40.00	923.0
41.00	364.0
44.00	765.0
45.00	319.0
49.00	1005.0
50.00	30744.0
51.00	119976.0
52.00	5941.0
53.00	207.0
55.00	692.0
56.00	3528.0
57.00	8652.0
58.00	428.0
58.90	163.0
61.00	1715.0
62.00	1833.0
63.00	4985.0
64.00	821.0
65.00	2636.0
66.00	156.0
67.00	193.0
68.00	2336.0
69.00	130264.0
70.00	690.0
73.00	1707.0
74.00	13839.0
75.00	20952.0
76.00	7506.0
77.00	163264.0
78.00	10406.0
79.00	9423.0
80.00	7273.0
81.00	10878.0
82.00	2554.0
83.00	2373.0
84.00	464.0
85.00	1915.0
86.00	3276.0
87.00	1250.0
87.90	608.0
88.90	198.0
90.90	2314.0
92.00	2874.0
93.00	17944.0
94.00	1231.0
94.90	305.0
96.00	964.0
97.00	352.0
98.00	13627.0
99.00	11940.0
100.00	925.0
101.00	6223.0
101.90	272.0
102.90	1833.0
104.00	3634.0
105.00	3356.0
106.00	1106.0
107.00	49104.0
108.00	7348.0
109.00	1270.0
110.00	87672.0
111.00	13173.0
112.00	1670.0
112.90	549.0
115.00	243.0
116.00	2545.0
117.00	43128.0
118.00	3158.0
118.90	499.0
119.90	583.0
120.90	244.0
122.00	3143.0
123.00	4430.0
124.00	2095.0
125.00	2110.0
126.00	517.0
127.00	176512.0
128.00	12943.0
129.00	65536.0
130.00	5403.0
131.00	1276.0
131.90	652.0
132.90	361.0
133.90	1744.0
135.00	5400.0
136.00	1997.0
137.00	3417.0
137.90	680.0

Instrument :
BNA_N
ClientSampleId :
DFTPP251

139.00	267.0
139.90	915.0
141.00	8714.0
142.00	2754.0
143.00	1927.0
144.00	532.0
144.90	509.0
146.00	1701.0
147.00	4703.0
148.00	10719.0
149.00	2262.0
150.00	758.0
151.20	1080.0
151.80	605.0
153.00	2864.0
154.00	2003.0
155.00	4871.0
156.00	6869.0
157.00	1589.0
157.90	1571.0
159.00	1395.0
159.90	3046.0
161.00	4358.0
161.90	1208.0
162.90	256.0
163.90	441.0
164.90	3080.0
166.00	2346.0
167.00	20216.0
168.00	12500.0
169.00	1660.0
170.00	537.0
170.90	567.0
172.00	1589.0
173.00	2264.0
174.00	4164.0
175.00	7294.0
176.00	1869.0
177.00	3192.0
178.00	1090.0
179.00	12665.0
180.00	9029.0
181.00	4460.0
181.90	843.0
183.00	260.0
183.90	1114.0
185.00	6123.0
186.00	46136.0
187.00	14382.0
188.00	1592.0
189.00	3077.0
189.90	500.0
191.00	1371.0
192.00	4285.0
193.00	4691.0
194.00	960.0
195.00	590.0
196.00	10007.0
196.80	1730.0
197.90	347456.0
199.00	23112.0
200.00	1883.0
201.50	1650.0
203.00	2420.0
204.00	13259.0
205.00	21672.0
206.00	91912.0
207.10	13374.0
208.00	3304.0
209.00	1171.0
210.10	1664.0
211.00	3695.0
212.90	271.0
215.00	1113.0
215.90	1771.0
217.00	23864.0
218.00	3372.0
219.00	328.0
221.00	14650.0
223.00	5145.0
224.00	47880.0
225.00	12122.0
226.10	1531.0
227.00	20184.0
228.00	2789.0
229.00	4837.0
230.00	723.0
231.00	1798.0
231.90	342.0
233.00	394.0
234.00	1436.0
235.00	1345.0
236.00	1141.0
237.00	2483.0
238.00	330.0
238.90	830.0

Instrument :
BNA_N
ClientSampleId :
DFTPP251

239.90	847.0
241.00	1130.0
242.00	2796.0
243.10	3117.0
244.00	41240.0
245.00	4983.0
246.00	7927.0
247.00	1450.0
247.90	416.0
249.00	1586.0
250.00	246.0
250.90	371.0
251.90	277.0
252.90	708.0
254.00	1196.0
255.00	219904.0
256.00	28648.0
257.00	2638.0
258.00	12405.0
259.00	1905.0
259.90	350.0
261.00	359.0
263.00	176.0
264.00	263.0
265.00	5270.0
265.90	724.0
267.00	183.0
270.00	193.0
270.90	553.0
272.00	480.0
273.00	7220.0
274.00	15959.0
275.00	95776.0
276.00	12340.0
277.00	6830.0
278.00	1575.0
279.00	296.0
281.00	490.0
281.90	321.0
283.00	1140.0
284.00	610.0
285.00	1148.0
286.10	187.0
289.00	289.0
290.00	311.0
290.90	178.0
292.00	360.0
293.00	2131.0
294.00	608.0
294.80	576.0
296.00	26256.0
297.00	4199.0
297.90	276.0
301.10	289.0
302.00	428.0
303.10	3205.0
304.00	966.0
305.00	181.0
307.90	477.0
308.90	288.0
310.00	364.0
313.10	214.0
314.10	1250.0
314.90	2590.0
316.00	1495.0
316.90	310.0
321.00	953.0
322.00	519.0
323.00	8729.0
324.00	1915.0
325.00	180.0
325.90	202.0
326.90	1626.0
328.00	1101.0
332.00	667.0
333.10	809.0
334.10	6239.0
335.00	1532.0
336.00	260.0
339.80	152.0
341.00	1019.0
342.00	343.0
346.00	1883.0
347.00	549.0
350.90	176.0
352.00	2875.0
353.00	2031.0
354.00	3517.0
355.00	691.0
359.00	257.0
365.00	13213.0
366.00	1812.0
370.10	337.0
371.00	799.0
372.00	5834.0
372.90	1224.0

Instrument :
BNA_N
ClientSampleId :
DFTPP251

373.90	193.0
377.00	154.0
383.10	1467.0
383.90	447.0
384.90	189.0
390.00	578.0
391.00	490.0
391.90	391.0
401.00	293.0
402.00	2452.0
403.00	3328.0
404.00	999.0
405.00	326.0
421.10	2622.0
422.00	2442.0
423.10	18592.0
424.10	5842.0
425.00	413.0
441.00	53608.0
442.00	252672.0
443.10	65416.0
444.00	5208.0
445.00	321.0

Instrument :
BNA_N
ClientSampleId :
DFTPP251

372.00	5474.0
373.00	1349.0
383.00	1613.0
384.00	336.0
385.00	157.0
390.00	590.0
390.90	419.0
392.00	314.0
401.00	297.0
402.00	1848.0
403.10	2801.0
403.90	898.0
405.00	263.0
415.00	154.0
421.00	2659.0
422.10	2333.0
423.20	14502.0
424.00	5344.0
425.00	923.0
441.10	47328.0
442.00	261440.0
443.00	49328.0
444.00	4576.0
444.90	250.0

Instrument :
BNA_N
ClientSampleId :
DFTPP251