

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091624\
 Data File : BN033992.D
 Acq On : 16 Sep 2024 17:37
 Operator : JU/RC
 Sample : PB163399BS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS399

Quant Time: Sep 16 18:43:49 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN082724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Sep 15 01:52:43 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.477	152	6262	0.400	ng/ul	0.00
4) Naphthalene-d8	10.233	136	16954	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.124	164	8227	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.877	188	15945	0.400	ng/ul	0.00
17) Chrysene-d12	21.095	240	12362	0.400	ng/ul	0.00
23) Perylene-d12	23.237	264	13821	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.101	96	5190	0.806	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.834	152	9468	0.389	ng/ul	-0.01
18) Fluoranthene-d10	18.920	212	15871	0.431	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.131	88	15136	2.119	ng/ul#	78
5) Naphthalene	10.283	128	17588	0.388	ng/ul	100
7) 2-Methylnaphthalene	11.911	142	10893	0.379	ng/ul	100
8) 1-Methylnaphthalene	12.131	142	12917	0.437	ng/ul	100
10) Acenaphthylene	13.837	152	13507	0.362	ng/ul	99
11) Acenaphthene	14.184	153	11029	0.409	ng/ul	99
12) Fluorene	15.183	166	11615	0.385	ng/ul	99
14) Pentachlorophenol	16.539	266	2482	0.495	ng/ul	99
15) Phenanthrene	16.919	178	17506	0.372	ng/ul	99
16) Anthracene	17.012	178	15423	0.365	ng/ul	100
19) Fluoranthene	18.948	202	19260	0.395	ng/ul	99
20) Pyrene	19.315	202	23149	0.447	ng/ul	99
21) Benzo(a)anthracene	21.080	228	17611	0.362	ng/ul	100
22) Chrysene	21.133	228	19256	0.392	ng/ul	99
24) Benzo(b)fluoranthene	22.602	252	18947	0.353	ng/ul	97
25) Benzo(k)fluoranthene	22.646	252	20936	0.390	ng/ul	99
26) Benzo(a)pyrene	23.143	252	17863	0.415	ng/ul	97
27) Indeno(1,2,3-cd)pyrene	25.354	276	24782	0.383	ng/ul#	99
28) Dibenzo(a,h)anthracene	25.368	278	19127	0.381	ng/ul	99
29) Benzo(g,h,i)perylene	25.995	276	19662	0.389	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091624\
Data File : BN033992.D
Acq On : 16 Sep 2024 17:37
Operator : JU/RC
Sample : PB163399BS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SLCS399

Quant Time: Sep 16 18:43:49 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN082724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Sep 15 01:52:43 2024
Response via : Initial Calibration

