

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091720\
 Data File : BN011821.D
 Acq On : 17 Sep 2020 13:21
 Operator : CG/JU
 Sample : SSTDICCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICCC0.4

Quant Time: Sep 17 19:00:14 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091720.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 17 16:24:49 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.715	152	1943	0.40	ng	0.00
7) Naphthalene-d8	10.491	136	7200	0.40	ng	# 0.00
13) Acenaphthene-d10	14.357	164	4146	0.40	ng	0.00
19) Phenanthrene-d10	17.102	188	8864	0.40	ng	0.00
29) Chrysene-d12	21.312	240	9310	0.40	ng	# 0.00
36) Perylene-d12	23.559	264	9120	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.307	112	2530	0.44	ng	0.00
5) Phenol-d6	6.878	99	3011	0.43	ng	0.00
8) Nitrobenzene-d5	8.856	82	3042	0.42	ng	0.00
11) 2-Methylnaphthalene-d10	12.091	152	5279	0.41	ng	0.00
14) 2,4,6-Tribromophenol	15.854	330	868	0.41	ng	0.00
15) 2-Fluorobiphenyl	12.974	172	7628	0.43	ng	0.00
27) Fluoranthene-d10	19.142	212	11405	0.41	ng	0.00
31) Terphenyl-d14	19.747	244	8928	0.42	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.270	88	1161	0.40	ng	90
3) n-Nitrosodimethylamine	3.576	42	1657	0.43	ng	# 80
6) bis(2-Chloroethyl)ether	7.143	93	2581	0.42	ng	97
9) Naphthalene	10.542	128	8415	0.42	ng	99
10) Hexachlorobutadiene	10.833	225	1951	0.42	ng	# 100
12) 2-Methylnaphthalene	12.167	142	5775	0.41	ng	98
16) Acenaphthylene	14.078	152	8084	0.40	ng	100
17) Acenaphthene	14.420	154	5667	0.41	ng	95
18) Fluorene	15.410	166	7191	0.41	ng	99
20) 4,6-Dinitro-2-methylph...	15.486	198	591	0.39	ng	# 65
21) 4-Bromophenyl-phenylether	16.299	248	2150	0.41	ng	# 64
22) Hexachlorobenzene	16.421	284	2683	0.41	ng	96
23) Atrazine	16.579	200	1879	0.40	ng	98
24) Pentachlorophenol	16.761	266	1040	0.40	ng	98
25) Phenanthrene	17.151	178	11474	0.41	ng	100
26) Anthracene	17.236	178	9724	0.40	ng	100
28) Fluoranthene	19.177	202	13143	0.40	ng	99
30) Pyrene	19.539	202	13365	0.41	ng	100
32) Benzo(a)anthracene	21.291	228	12944	0.41	ng	99
33) Chrysene	21.344	228	13190	0.41	ng	99
34) Bis(2-ethylhexyl)phtha...	21.227	149	6279	0.39	ng	97
35) Indeno(1,2,3-cd)pyrene	25.818	276	17209	0.43	ng	98
37) Benzo(b)fluoranthene	22.885	252	13788	0.41	ng	# 92
38) Benzo(k)fluoranthene	22.929	252	14840	0.42	ng	# 92
39) Benzo(a)pyrene	23.463	252	12352	0.41	ng	# 87
40) Dibenzo(a,h)anthracene	25.835	278	13856	0.42	ng	# 92
41) Benzo(g,h,i)perylene	26.506	276	15151	0.43	ng	# 96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091720\
Data File : BN011821.D
Acq On : 17 Sep 2020 13:21
Operator : CG/JU
Sample : SSTDICCC0.4
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Quant Time: Sep 17 19:00:14 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091720.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 17 16:24:49 2020
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.715 min Scan# 588
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:152 Resp: 1943
Ion Ratio Lower Upper
152 100
150 150.4 118.3 177.5
115 67.0 51.3 76.9



#2
1,4-Dioxane
Concen: 0.40 ng
RT: 3.270 min Scan# 36
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21

Tgt Ion: 88 Resp: 1161
Ion Ratio Lower Upper
88 100
58 72.0 63.3 94.9
43 33.2 33.0 49.4









#3
n-Nitrosodimethylamine
Concen: 0.43 ng
RT: 3.576 min Scan# 74
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion: 42 Resp: 1657
Ion Ratio Lower Upper
42 100
74 98.8 64.4 96.6#
44 4.7 6.7 10.1#



#4
2-Fluorophenol
Concen: 0.44 ng
RT: 5.307 min Scan# 289
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21

Tgt Ion: 112 Resp: 2530
Ion Ratio Lower Upper
112 100
64 61.9 49.0 73.6
63 35.5 31.8 47.6









#5
Phenol-d6
Concen: 0.43 ng
RT: 6.878 min Scan# 484
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	99	Resp:	3011
Ion	Ratio	Lower	Upper
99	100		
42	25.0	23.4	35.0
71	43.7	38.1	57.1



#6
bis(2-Chloroethyl)ether
Concen: 0.42 ng
RT: 7.143 min Scan# 517
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21


Tgt Ion:	93	Resp:	2581
Ion	Ratio	Lower	Upper
93	100		
63	72.7	61.1	91.7
95	31.3	25.5	38.3








#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.491 min Scan# 834
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4







Tgt Ion:	136	Resp:	7200
Ion	Ratio	Lower	Upper
136	100		
137	12.4	9.5	14.3
54	10.0	5.6	8.4#
68	6.7	3.2	4.8#



#8
Nitrobenzene-d5
Concen: 0.42 ng
RT: 8.856 min Scan# 705
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21

Tgt Ion:	82	Resp:	3042
Ion	Ratio	Lower	Upper
82	100		
128	38.4	34.2	51.2
54	57.6	43.1	64.7









#9
Naphthalene
Concen: 0.42 ng
RT: 10.542 min Scan# 838
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	128	Resp:	8415
Ion	Ratio	Lower	Upper
128	100		
129	11.6	9.7	14.5
127	14.3	11.0	16.4



#10
Hexachlorobutadiene
Concen: 0.42 ng
RT: 10.833 min Scan# 861
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21

Tgt Ion:	225	Resp:	1951
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.7	51.9	77.9









#11
2-Methylnaphthalene-d10
Concen: 0.41 ng
RT: 12.091 min Scan# 1044
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:152 Resp: 5279
Ion Ratio Lower Upper
152 100
151 20.5 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.41 ng
RT: 12.167 min Scan# 1061
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21

Tgt Ion:142 Resp: 5775
Ion Ratio Lower Upper
142 100
141 88.3 70.0 105.0
115 41.5 35.8 53.6









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.357 min Scan# 1283
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:164 Resp: 4146
Ion Ratio Lower Upper
164 100
162 100.9 83.6 125.4
160 49.8 41.2 61.8



#14
2,4,6-Tribromophenol
Concen: 0.41 ng
RT: 15.854 min Scan# 1401
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21

Tgt Ion:330 Resp: 868
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 38.7 33.7 50.5









#15
2-Fluorobiphenyl
Concen: 0.43 ng
RT: 12.974 min Scan# 1174
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:172 Resp: 7628
Ion Ratio Lower Upper
172 100
171 35.6 27.4 41.2
170 24.7 17.8 26.6



#16
Acenaphthylene
Concen: 0.40 ng
RT: 14.078 min Scan# 1261
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21

Tgt Ion:152 Resp: 8084
Ion Ratio Lower Upper
152 100
151 20.0 15.9 23.9
153 12.9 10.5 15.7









#17
Acenaphthene
Concen: 0.41 ng
RT: 14.420 min Scan# 1288
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:154 Resp: 5667
Ion Ratio Lower Upper
154 100
153 110.2 83.4 125.2
152 55.4 42.8 64.2



#18
Fluorene
Concen: 0.41 ng
RT: 15.410 min Scan# 1366
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21

Tgt Ion:166 Resp: 7191
Ion Ratio Lower Upper
166 100
165 98.9 78.8 118.2
167 13.1 11.2 16.8









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.102 min Scan# 1503
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:188 Resp: 8864
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 9.6 7.7 11.5



#20
4,6-Dinitro-2-methylphenol
Concen: 0.39 ng
RT: 15.486 min Scan# 1372
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21

Tgt Ion:198 Resp: 591
Ion Ratio Lower Upper
198 100
51 86.1 45.0 67.4#
105 54.0 30.4 45.6#









#21
4-Bromophenyl-phenylether
Concen: 0.41 ng
RT: 16.299 min Scan# 1437
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:248 Resp: 2150
Ion Ratio Lower Upper
248 100
250 90.3 82.7 124.1
141 101.7 33.7 50.5#



#22
Hexachlorobenzene
Concen: 0.41 ng
RT: 16.421 min Scan# 1447
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21

Tgt Ion:284 Resp: 2683
Ion Ratio Lower Upper
284 100
142 37.5 32.4 48.6
249 31.9 26.8 40.2





#23

Atrazine

Concen: 0.40 ng

RT: 16.579 min Scan# 1460

Delta R.T. 0.000 min

Lab File: BN011821.D

Acq: 17 Sep 2020 13:21





Tgt Ion:	200	Resp:	1879
Ion	Ratio	Lower	Upper
200	100		
173	23.8	18.7	28.1
215	48.8	40.2	60.4

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4



#24

Pentachlorophenol

Concen: 0.40 ng




RT: 16.761 min Scan# 1475


Delta R.T. 0.000 min

Lab File: BN011821.D

Acq: 17 Sep 2020 13:21

Tgt Ion:	266	Resp:	1040
Ion	Ratio	Lower	Upper
266	100		
264	64.7	51.2	76.8
268	62.7	52.4	78.6









#25
Phenanthrene
Concen: 0.41 ng
RT: 17.151 min Scan# 1507
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:178 Resp: 11474
Ion Ratio Lower Upper
178 100
176 18.6 15.1 22.7
179 15.5 12.3 18.5



#26
Anthracene
Concen: 0.40 ng
RT: 17.236 min Scan# 1514
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21

Tgt Ion:178 Resp: 9724
Ion Ratio Lower Upper
178 100
176 18.5 14.6 21.8
179 15.3 12.4 18.6









#27
Fluoranthene-d10
Concen: 0.41 ng
RT: 19.142 min Scan# 1805
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	212	Resp:	11405
Ion Ratio	Lower	Upper	
212	100		
106	9.3	6.4	9.6
104	5.5	3.9	5.9



#28
Fluoranthene
Concen: 0.40 ng
RT: 19.177 min Scan# 1812
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21

Tgt Ion:	202	Resp:	13143
Ion Ratio	Lower	Upper	
202	100		
101	8.1	5.4	8.2
203	17.1	13.8	20.6









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.312 min Scan# 2102
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:240 Resp: 9310
Ion Ratio Lower Upper
240 100
120 7.3 3.5 5.3#
236 27.1 20.5 30.7



#30
Pyrene
Concen: 0.41 ng
RT: 19.539 min Scan# 1885
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21

Tgt Ion:202 Resp: 13365
Ion Ratio Lower Upper
202 100
200 20.6 16.5 24.7
203 17.7 13.9 20.9









#31
 Terphenyl-d14
 Concen: 0.42 ng
 RT: 19.747 min Scan# 1927
 Delta R.T. 0.000 min
 Lab File: BN011821.D
 Acq: 17 Sep 2020 13:21




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.4


Tgt Ion:	244	Resp:	8928
Ion	Ratio	Lower	Upper
244	100		
212	9.1	7.3	10.9
122	10.8	7.8	11.6



#32
 Benzo(a)anthracene
 Concen: 0.41 ng
 RT: 21.291 min Scan# 2100
 Delta R.T. 0.000 min
 Lab File: BN011821.D
 Acq: 17 Sep 2020 13:21

Tgt Ion:	228	Resp:	12944
Ion	Ratio	Lower	Upper
228	100		
226	27.0	21.0	31.4
229	20.2	16.0	24.0









#33
Chrysene
Concen: 0.41 ng
RT: 21.344 min Scan# 2105
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:228 Resp: 13190
Ion Ratio Lower Upper
228 100
226 30.0 23.3 34.9
229 19.8 15.8 23.8



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.39 ng
RT: 21.227 min Scan# 2094
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21


Tgt Ion:149 Resp: 6279
Ion Ratio Lower Upper
149 100
167 27.9 23.6 35.4
279 5.0 4.0 6.0








#35
Indeno(1,2,3-cd)pyrene
Concen: 0.43 ng
RT: 25.818 min Scan# 3202
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21


Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4





Tgt Ion:276 Resp: 17209
Ion Ratio Lower Upper
276 100
138 17.8 13.0 19.4
277 24.5 20.0 30.0




#36
Perylene-d12
Concen: 0.40 ng
RT: 23.559 min Scan# 2542
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21



Tgt Ion:264 Resp: 9120
Ion Ratio Lower Upper
264 100
260 25.2 19.6 29.4
265 83.0 38.6 58.0#









#37
Benzo(b)fluoranthene
Concen: 0.41 ng
RT: 22.885 min Scan# 2345
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:252 Resp: 13788
Ion Ratio Lower Upper
252 100
253 27.5 19.1 28.7
125 11.7 6.2 9.2#



#38
Benzo(k)fluoranthene
Concen: 0.42 ng
RT: 22.929 min Scan# 2358
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21

Tgt Ion:252 Resp: 14840
Ion Ratio Lower Upper
252 100
253 27.4 19.3 28.9
125 11.6 6.0 9.0#









#39
Benzo(a)pyrene
Concen: 0.41 ng
RT: 23.463 min Scan# 2514
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:252 Resp: 12352
Ion Ratio Lower Upper
252 100
253 30.5 19.8 29.6#
125 15.0 7.3 10.9#



#40
Dibenzo(a,h)anthracene
Concen: 0.42 ng
RT: 25.835 min Scan# 3207
Delta R.T. 0.000 min
Lab File: BN011821.D
Acq: 17 Sep 2020 13:21

Tgt Ion:278 Resp: 13856
Ion Ratio Lower Upper
278 100
139 13.8 8.6 13.0#
279 28.9 19.8 29.8





#41
 Benzo(g,h,i)perylene
 Concen: 0.43 ng
 RT: 26.506 min Scan# 3403
 Delta R.T. 0.000 min
 Lab File: BN011821.D
 Acq: 17 Sep 2020 13:21

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICCC0.4

Tgt Ion:	276	Resp:	15151
Ion Ratio	Lower	Upper	
276	100		
277	25.0	19.4	29.0
138	17.1	11.3	16.9#

