

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091720\
 Data File : BN011820.D
 Acq On : 17 Sep 2020 12:45
 Operator : CG/JU
 Sample : SSTDICC0.2
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2

Quant Time: Sep 17 18:59:55 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091720.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 17 16:24:49 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.715	152	1898	0.40	ng	0.00
7) Naphthalene-d8	10.491	136	7094	0.40	ng	# 0.00
13) Acenaphthene-d10	14.357	164	4016	0.40	ng	0.00
19) Phenanthrene-d10	17.102	188	8695	0.40	ng	0.00
29) Chrysene-d12	21.312	240	9206	0.40	ng	# 0.00
36) Perylene-d12	23.562	264	8868	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.307	112	1098	0.20	ng	0.00
5) Phenol-d6	6.878	99	1313	0.19	ng	0.00
8) Nitrobenzene-d5	8.856	82	1364	0.19	ng	0.00
11) 2-Methylnaphthalene-d10	12.091	152	2394	0.19	ng	0.00
14) 2,4,6-Tribromophenol	15.854	330	366	0.18	ng	0.00
15) 2-Fluorobiphenyl	12.974	172	3445	0.20	ng	0.00
27) Fluoranthene-d10	19.147	212	5273	0.19	ng	0.00
31) Terphenyl-d14	19.752	244	4090	0.19	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.270	88	556	0.20	ng	93
3) n-Nitrosodimethylamine	3.584	42	734	0.20	ng	# 77
6) bis(2-Chloroethyl)ether	7.143	93	1187	0.20	ng	97
9) Naphthalene	10.542	128	3943	0.20	ng	98
10) Hexachlorobutadiene	10.846	225	919	0.20	ng	# 98
12) 2-Methylnaphthalene	12.167	142	2637	0.19	ng	98
16) Acenaphthylene	14.078	152	3685	0.19	ng	99
17) Acenaphthene	14.420	154	2566	0.19	ng	93
18) Fluorene	15.410	166	3197	0.19	ng	99
20) 4,6-Dinitro-2-methylph...	15.486	198	238	0.16	ng	# 8
21) 4-Bromophenyl-phenylether	16.299	248	1007	0.19	ng	# 64
22) Hexachlorobenzene	16.421	284	1245	0.19	ng	98
23) Atrazine	16.579	200	878	0.19	ng	# 96
24) Pentachlorophenol	16.761	266	477	0.19	ng	96
25) Phenanthrene	17.151	178	5292	0.19	ng	99
26) Anthracene	17.236	178	4382	0.18	ng	99
28) Fluoranthene	19.176	202	6045	0.19	ng	98
30) Pyrene	19.539	202	6210	0.19	ng	100
32) Benzo(a)anthracene	21.290	228	5730	0.18	ng	97
33) Chrysene	21.344	228	6163	0.19	ng	98
34) Bis(2-ethylhexyl)phtha...	21.237	149	3352	0.21	ng	94
35) Indeno(1,2,3-cd)pyrene	25.822	276	7403	0.19	ng	98
37) Benzo(b)fluoranthene	22.888	252	6244	0.19	ng	# 78
38) Benzo(k)fluoranthene	22.933	252	6377	0.19	ng	# 78
39) Benzo(a)pyrene	23.467	252	5460	0.19	ng	# 69
40) Dibenzo(a,h)anthracene	25.835	278	5654	0.17	ng	# 77
41) Benzo(g,h,i)perylene	26.513	276	6554	0.19	ng	# 92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091720\
Data File : BN011820.D
Acq On : 17 Sep 2020 12:45
Operator : CG/JU
Sample : SSTDIC0.2
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2

Quant Time: Sep 17 18:59:55 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091720.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 17 16:24:49 2020
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.715 min Scan# 588
 Delta R.T. 0.000 min
 Lab File: BN011820.D
 Acq: 17 Sep 2020 12:45




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2


Tgt Ion:152 Resp: 1898
 Ion Ratio Lower Upper
 152 100
 150 155.5 118.3 177.5
 115 65.7 51.3 76.9



#2
 1,4-Dioxane
 Concen: 0.20 ng
 RT: 3.270 min Scan# 36
 Delta R.T. -0.000 min
 Lab File: BN011820.D
 Acq: 17 Sep 2020 12:45

Tgt Ion: 88 Resp: 556
 Ion Ratio Lower Upper
 88 100
 58 71.2 63.3 94.9
 43 43.2 33.0 49.4









#3
n-Nitrosodimethylamine
Concen: 0.20 ng
RT: 3.584 min Scan# 75
Delta R.T. 0.008 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion: 42 Resp: 734
Ion Ratio Lower Upper
42 100
74 101.5 64.4 96.6#
44 13.5 6.7 10.1#



#4
2-Fluorophenol
Concen: 0.20 ng
RT: 5.307 min Scan# 289
Delta R.T. -0.000 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45

Tgt Ion: 112 Resp: 1098
Ion Ratio Lower Upper
112 100
64 63.5 49.0 73.6
63 37.4 31.8 47.6









#5
Phenol-d6
Concen: 0.19 ng
RT: 6.878 min Scan# 484
Delta R.T. -0.000 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2

Tgt Ion: 99 Resp: 1313




Ion	Ratio	Lower	Upper
99	100		
42	24.6	23.4	35.0
71	41.3	38.1	57.1




#6
bis(2-Chloroethyl)ether
Concen: 0.20 ng
RT: 7.143 min Scan# 517
Delta R.T. -0.000 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45

Tgt Ion: 93 Resp: 1187

Ion	Ratio	Lower	Upper
93	100		
63	72.9	61.1	91.7
95	31.8	25.5	38.3









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.491 min Scan# 834
Delta R.T. -0.000 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2

Tgt Ion: 136 Resp: 7094




Ion	Ratio	Lower	Upper
136	100		
137	12.1	9.5	14.3
54	9.7	5.6	8.4#
68	6.8	3.2	4.8#




#8
Nitrobenzene-d5
Concen: 0.19 ng
RT: 8.856 min Scan# 705
Delta R.T. -0.000 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45

Tgt Ion: 82 Resp: 1364

Ion	Ratio	Lower	Upper
82	100		
128	37.9	34.2	51.2
54	60.5	43.1	64.7









#9
Naphthalene
Concen: 0.20 ng
RT: 10.542 min Scan# 838
Delta R.T. -0.000 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:128 Resp: 3943
Ion Ratio Lower Upper
128 100
129 12.6 9.7 14.5
127 14.9 11.0 16.4



#10
Hexachlorobutadiene
Concen: 0.20 ng
RT: 10.846 min Scan# 862
Delta R.T. 0.013 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45

Tgt Ion:225 Resp: 919
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.5 51.9 77.9









#11
2-Methylnaphthalene-d10
Concen: 0.19 ng
RT: 12.091 min Scan# 1044
Delta R.T. -0.000 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:152 Resp: 2394
Ion Ratio Lower Upper
152 100
151 21.9 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.19 ng
RT: 12.167 min Scan# 1061
Delta R.T. -0.000 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45

Tgt Ion:142 Resp: 2637
Ion Ratio Lower Upper
142 100
141 89.6 70.0 105.0
115 43.1 35.8 53.6









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.357 min Scan# 1283
Delta R.T. -0.000 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:164 Resp: 4016
Ion Ratio Lower Upper
164 100
162 103.8 83.6 125.4
160 49.5 41.2 61.8



#14
2,4,6-Tribromophenol
Concen: 0.18 ng
RT: 15.854 min Scan# 1401
Delta R.T. -0.000 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45

Tgt Ion:330 Resp: 366
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 47.8 33.7 50.5









#15
2-Fluorobiphenyl
Concen: 0.20 ng
RT: 12.974 min Scan# 1174
Delta R.T. -0.000 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:	172	Resp:	3445
Ion	Ratio	Lower	Upper
172	100		
171	36.1	27.4	41.2
170	24.4	17.8	26.6



#16
Acenaphthylene
Concen: 0.19 ng
RT: 14.078 min Scan# 1261
Delta R.T. -0.000 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45

Tgt Ion:	152	Resp:	3685
Ion	Ratio	Lower	Upper
152	100		
151	18.9	15.9	23.9
153	13.2	10.5	15.7









#17
Acenaphthene
Concen: 0.19 ng
RT: 14.420 min Scan# 1288
Delta R.T. -0.000 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:	154	Resp:	2566
Ion	Ratio	Lower	Upper
154	100		
153	112.3	83.4	125.2
152	57.7	42.8	64.2



#18
Fluorene
Concen: 0.19 ng
RT: 15.410 min Scan# 1366
Delta R.T. -0.000 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45

Tgt Ion:	166	Resp:	3197
Ion	Ratio	Lower	Upper
166	100		
165	100.0	78.8	118.2
167	14.2	11.2	16.8









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.102 min Scan# 1503
Delta R.T. -0.000 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:	188	Resp:	8695
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	9.3	7.7	11.5



#20
4,6-Dinitro-2-methylphenol
Concen: 0.16 ng
RT: 15.486 min Scan# 1372
Delta R.T. -0.000 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45

Tgt Ion:	198	Resp:	238
Ion Ratio	Lower	Upper	
198	100		
51	132.5	45.0	67.4#
105	81.9	30.4	45.6#









#21
4-Bromophenyl-phenylether
Concen: 0.19 ng
RT: 16.299 min Scan# 1437
Delta R.T. -0.000 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:248 Resp: 1007
Ion Ratio Lower Upper
248 100
250 88.8 82.7 124.1
141 99.4 33.7 50.5#



#22
Hexachlorobenzene
Concen: 0.19 ng
RT: 16.421 min Scan# 1447
Delta R.T. -0.000 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45

Tgt Ion:284 Resp: 1245
Ion Ratio Lower Upper
284 100
142 39.2 32.4 48.6
249 34.1 26.8 40.2









#23
Atrazine
Concen: 0.19 ng
RT: 16.579 min Scan# 1460
Delta R.T. -0.000 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:	200	Resp:	878
Ion Ratio	Lower	Upper	
200	100		
173	28.7	18.7	28.1#
215	49.3	40.2	60.4



#24
Pentachlorophenol
Concen: 0.19 ng
RT: 16.761 min Scan# 1475
Delta R.T. -0.000 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45

Tgt Ion:	266	Resp:	477
Ion Ratio	Lower	Upper	
266	100		
264	62.7	51.2	76.8
268	70.2	52.4	78.6









#25
Phenanthrene
Concen: 0.19 ng
RT: 17.151 min Scan# 1507
Delta R.T. -0.000 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:	178	Resp:	5292
Ion	Ratio	Lower	Upper
178	100		
176	19.3	15.1	22.7
179	16.0	12.3	18.5



#26
Anthracene
Concen: 0.18 ng
RT: 17.236 min Scan# 1514
Delta R.T. -0.000 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45

Tgt Ion:	178	Resp:	4382
Ion	Ratio	Lower	Upper
178	100		
176	18.1	14.6	21.8
179	16.2	12.4	18.6









#27
Fluoranthene-d10
Concen: 0.19 ng
RT: 19.147 min Scan# 1806
Delta R.T. 0.005 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2

Tgt Ion:212 Resp: 5273




Ion	Ratio	Lower	Upper
212	100		
106	9.5	6.4	9.6
104	5.8	3.9	5.9




#28
Fluoranthene
Concen: 0.19 ng
RT: 19.176 min Scan# 1812
Delta R.T. -0.000 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45

Tgt Ion:202 Resp: 6045

Ion	Ratio	Lower	Upper
202	100		
101	7.8	5.4	8.2
203	17.6	13.8	20.6









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.312 min Scan# 2102
Delta R.T. -0.000 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:240 Resp: 9206
Ion Ratio Lower Upper
240 100
120 7.9 3.5 5.3#
236 26.9 20.5 30.7



#30
Pyrene
Concen: 0.19 ng
RT: 19.539 min Scan# 1885
Delta R.T. -0.000 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45

Tgt Ion:202 Resp: 6210
Ion Ratio Lower Upper
202 100
200 20.4 16.5 24.7
203 17.5 13.9 20.9









#31
Terphenyl-d14
Concen: 0.19 ng
RT: 19.752 min Scan# 1928
Delta R.T. 0.005 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:244 Resp: 4090
Ion Ratio Lower Upper
244 100
212 10.2 7.3 10.9
122 10.3 7.8 11.6



#32
Benzo(a)anthracene
Concen: 0.18 ng
RT: 21.290 min Scan# 2100
Delta R.T. -0.000 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45

Tgt Ion:228 Resp: 5730
Ion Ratio Lower Upper
228 100
226 29.0 21.0 31.4
229 20.3 16.0 24.0









#33
Chrysene
Concen: 0.19 ng
RT: 21.344 min Scan# 2105
Delta R.T. -0.000 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:228 Resp: 6163
Ion Ratio Lower Upper
228 100
226 31.0 23.3 34.9
229 19.9 15.8 23.8



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.21 ng
RT: 21.237 min Scan# 2095
Delta R.T. 0.010 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45


Tgt Ion:149 Resp: 3352
Ion Ratio Lower Upper
149 100
167 26.1 23.6 35.4
279 4.4 4.0 6.0








#35
 Indeno(1,2,3-cd)pyrene
 Concen: 0.19 ng
 RT: 25.822 min Scan# 3203
 Delta R.T. 0.003 min
 Lab File: BN011820.D
 Acq: 17 Sep 2020 12:45

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2



Tgt Ion:276 Resp: 7403




Ion	Ratio	Lower	Upper
276	100		
138	16.7	13.0	19.4
277	23.7	20.0	30.0




#36
 Perylene-d12
 Concen: 0.40 ng
 RT: 23.562 min Scan# 2543
 Delta R.T. 0.003 min
 Lab File: BN011820.D
 Acq: 17 Sep 2020 12:45

Tgt Ion:264 Resp: 8868

Ion	Ratio	Lower	Upper
264	100		
260	25.2	19.6	29.4
265	86.0	38.6	58.0









#37
Benzo(b)fluoranthene
Concen: 0.19 ng
RT: 22.888 min Scan# 2346
Delta R.T. 0.003 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:252 Resp: 6244
Ion Ratio Lower Upper
252 100
253 34.0 19.1 28.7#
125 16.5 6.2 9.2#



#38
Benzo(k)fluoranthene
Concen: 0.19 ng
RT: 22.933 min Scan# 2359
Delta R.T. 0.003 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45

Tgt Ion:252 Resp: 6377
Ion Ratio Lower Upper
252 100
253 34.5 19.3 28.9#
125 16.5 6.0 9.0#









#39
Benzo(a)pyrene
Concen: 0.19 ng
RT: 23.467 min Scan# 2515
Delta R.T. 0.003 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:252 Resp: 5460
Ion Ratio Lower Upper
252 100
253 39.3 19.8 29.6#
125 22.1 7.3 10.9#



#40
Dibenzo(a,h)anthracene
Concen: 0.17 ng
RT: 25.835 min Scan# 3207
Delta R.T. -0.000 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45


Tgt Ion:278 Resp: 5654
Ion Ratio Lower Upper
278 100
139 16.8 8.6 13.0#
279 37.5 19.8 29.8#





#41
Benzo(g,h,i)perylene
Concen: 0.19 ng
RT: 26.513 min Scan# 3405
Delta R.T. 0.007 min
Lab File: BN011820.D
Acq: 17 Sep 2020 12:45

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2



Tgt Ion:276 Resp: 6554
Ion Ratio Lower Upper
276 100
277 27.1 19.4 29.0
138 18.8 11.3 16.9#

