

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091720\
 Data File : BN011824.D
 Acq On : 17 Sep 2020 15:10
 Operator : CG/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Quant Time: Sep 17 19:01:09 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091720.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 17 16:24:49 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.715	152	2065	0.40	ng	0.00
7) Naphthalene-d8	10.491	136	7841	0.40	ng	# 0.00
13) Acenaphthene-d10	14.357	164	4668	0.40	ng	0.00
19) Phenanthrene-d10	17.102	188	9869	0.40	ng	0.00
29) Chrysene-d12	21.312	240	10978	0.40	ng	# 0.00
36) Perylene-d12	23.559	264	10388	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.307	112	19320	3.17	ng	0.00
5) Phenol-d6	6.878	99	25026	3.35	ng	0.00
8) Nitrobenzene-d5	8.856	82	25771	3.26	ng	0.00
11) 2-Methylnaphthalene-d10	12.091	152	45679	3.28	ng	0.00
14) 2,4,6-Tribromophenol	15.854	330	8102	3.41	ng	0.00
15) 2-Fluorobiphenyl	12.974	172	63069	3.12	ng	0.00
27) Fluoranthene-d10	19.147	212	103228	3.30	ng	0.00
31) Terphenyl-d14	19.747	244	78846	3.15	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.262	88	9277	3.03	ng	92
3) n-Nitrosodimethylamine	3.568	42	13195	3.23	ng	# 79
6) bis(2-Chloroethyl)ether	7.143	93	20931	3.18	ng	99
9) Naphthalene	10.542	128	70769	3.21	ng	98
10) Hexachlorobutadiene	10.833	225	15774	3.10	ng	# 99
12) 2-Methylnaphthalene	12.167	142	49867	3.27	ng	97
16) Acenaphthylene	14.078	152	74772	3.32	ng	99
17) Acenaphthene	14.420	154	51742	3.30	ng	98
18) Fluorene	15.410	166	64251	3.29	ng	99
20) 4,6-Dinitro-2-methylph...	15.486	198	6532	3.87	ng	# 69
21) 4-Bromophenyl-phenylether	16.299	248	19196	3.27	ng	# 67
22) Hexachlorobenzene	16.421	284	23197	3.20	ng	96
23) Atrazine	16.579	200	17977	3.41	ng	96
24) Pentachlorophenol	16.761	266	10020	3.42	ng	97
25) Phenanthrene	17.151	178	103026	3.30	ng	100
26) Anthracene	17.236	178	92432	3.41	ng	100
28) Fluoranthene	19.171	202	121696	3.33	ng	99
30) Pyrene	19.539	202	121956	3.18	ng	100
32) Benzo(a)anthracene	21.290	228	122173	3.29	ng	98
33) Chrysene	21.344	228	121459	3.18	ng	99
34) Bis(2-ethylhexyl)phtha...	21.237	149	58313	3.05	ng	98
35) Indeno(1,2,3-cd)pyrene	25.818	276	157698	3.31	ng	98
37) Benzo(b)fluoranthene	22.885	252	129562	3.35	ng	97
38) Benzo(k)fluoranthene	22.929	252	130898	3.27	ng	97
39) Benzo(a)pyrene	23.463	252	115414	3.35	ng	96
40) Dibenzo(a,h)anthracene	25.835	278	130656	3.44	ng	98
41) Benzo(g,h,i)perylene	26.506	276	133129	3.28	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091720\
Data File : BN011824.D
Acq On : 17 Sep 2020 15:10
Operator : CG/JU
Sample : SSTDICC3.2
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Quant Time: Sep 17 19:01:09 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN091720.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 17 16:24:49 2020
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.715 min Scan# 588
Delta R.T. 0.000 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 2065
Ion Ratio Lower Upper
152 100
150 150.5 118.3 177.5
115 66.1 51.3 76.9



#2
1,4-Dioxane
Concen: 3.03 ng
RT: 3.262 min Scan# 35
Delta R.T. -0.008 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10

Tgt Ion: 88 Resp: 9277
Ion Ratio Lower Upper
88 100
58 73.8 63.3 94.9
43 34.0 33.0 49.4









#3
n-Nitrosodimethylamine
Concen: 3.23 ng
RT: 3.568 min Scan# 73
Delta R.T. -0.008 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 42 Resp: 13195
Ion Ratio Lower Upper
42 100
74 100.3 64.4 96.6#
44 5.6 6.7 10.1#



#4
2-Fluorophenol
Concen: 3.17 ng
RT: 5.307 min Scan# 289
Delta R.T. -0.000 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10

Tgt Ion: 112 Resp: 19320
Ion Ratio Lower Upper
112 100
64 62.2 49.0 73.6
63 35.8 31.8 47.6









#5
Phenol-d6
Concen: 3.35 ng
RT: 6.878 min Scan# 484
Delta R.T. -0.000 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 99 Resp: 25026
Ion Ratio Lower Upper
99 100
42 24.9 23.4 35.0
71 43.2 38.1 57.1



#6
bis(2-Chloroethyl)ether
Concen: 3.18 ng
RT: 7.143 min Scan# 517
Delta R.T. -0.000 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10

Tgt Ion: 93 Resp: 20931
Ion Ratio Lower Upper
93 100
63 75.6 61.1 91.7
95 31.9 25.5 38.3









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.491 min Scan# 834
Delta R.T. -0.000 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	136	Resp:	7841
Ion	Ratio	Lower	Upper
136	100		
137	12.4	9.5	14.3
54	10.2	5.6	8.4#
68	6.6	3.2	4.8#



#8
Nitrobenzene-d5
Concen: 3.26 ng
RT: 8.856 min Scan# 705
Delta R.T. -0.000 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10

Tgt Ion:	82	Resp:	25771
Ion	Ratio	Lower	Upper
82	100		
128	37.0	34.2	51.2
54	53.5	43.1	64.7





#9

Naphthalene

Concen: 3.21 ng

RT: 10.542 min Scan# 838

Delta R.T. -0.000 min

Lab File: BN011824.D

Acq: 17 Sep 2020 15:10


Tgt Ion:128 Resp: 70769

Ion Ratio Lower Upper




128 100

129 10.9 9.7 14.5

127 13.4 11.0 16.4



Abundance



#10

Hexachlorobutadiene

Concen: 3.10 ng

RT: 10.833 min Scan# 861

Delta R.T. -0.000 min

Lab File: BN011824.D

Acq: 17 Sep 2020 15:10


Tgt Ion:225 Resp: 15774

Ion Ratio Lower Upper



225 100


223 0.0 0.0 0.0

227 64.0 51.9 77.9



Abundance









#11
2-Methylnaphthalene-d10
Concen: 3.28 ng
RT: 12.091 min Scan# 1044
Delta R.T. -0.000 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 45679
Ion Ratio Lower Upper
152 100
151 20.5 16.7 25.1



#12
2-Methylnaphthalene
Concen: 3.27 ng
RT: 12.167 min Scan# 1061
Delta R.T. -0.000 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10

Tgt Ion:142 Resp: 49867
Ion Ratio Lower Upper
142 100
141 89.1 70.0 105.0
115 40.9 35.8 53.6









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.357 min Scan# 1283
Delta R.T. -0.000 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:164 Resp: 4668
Ion Ratio Lower Upper
164 100
162 102.1 83.6 125.4
160 48.2 41.2 61.8



#14
2,4,6-Tribromophenol
Concen: 3.41 ng
RT: 15.854 min Scan# 1401
Delta R.T. -0.000 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10

Tgt Ion:330 Resp: 8102
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 40.1 33.7 50.5









#15
2-Fluorobiphenyl
Concen: 3.12 ng
RT: 12.974 min Scan# 1174
Delta R.T. -0.000 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:172 Resp: 63069




Ion	Ratio	Lower	Upper
172	100		
171	35.4	27.4	41.2
170	23.6	17.8	26.6




#16
Acenaphthylene
Concen: 3.32 ng
RT: 14.078 min Scan# 1261
Delta R.T. -0.000 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10

Tgt Ion:152 Resp: 74772

Ion	Ratio	Lower	Upper
152	100		
151	19.5	15.9	23.9
153	13.0	10.5	15.7









#17
Acenaphthene
Concen: 3.30 ng
RT: 14.420 min Scan# 1288
Delta R.T. -0.000 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:154 Resp: 51742
Ion Ratio Lower Upper
154 100
153 106.2 83.4 125.2
152 52.9 42.8 64.2



#18
Fluorene
Concen: 3.29 ng
RT: 15.410 min Scan# 1366
Delta R.T. -0.000 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10

Tgt Ion:166 Resp: 64251
Ion Ratio Lower Upper
166 100
165 97.9 78.8 118.2
167 13.4 11.2 16.8









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.102 min Scan# 1503
Delta R.T. -0.000 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	188	Resp:	9869
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	9.0	7.7	11.5



#20
4,6-Dinitro-2-methylphenol
Concen: 3.87 ng
RT: 15.486 min Scan# 1372
Delta R.T. -0.000 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10

Tgt Ion:	198	Resp:	6532
Ion Ratio	Lower	Upper	
198	100		
51	27.9	45.0	67.4#
105	26.3	30.4	45.6#









#21
4-Bromophenyl-phenylether
Concen: 3.27 ng
RT: 16.299 min Scan# 1437
Delta R.T. -0.000 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:248 Resp: 19196
Ion Ratio Lower Upper
248 100
250 91.1 82.7 124.1
141 94.7 33.7 50.5#



#22
Hexachlorobenzene
Concen: 3.20 ng
RT: 16.421 min Scan# 1447
Delta R.T. -0.000 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10

Tgt Ion:284 Resp: 23197
Ion Ratio Lower Upper
284 100
142 37.1 32.4 48.6
249 31.7 26.8 40.2









#23
Atrazine
Concen: 3.41 ng
RT: 16.579 min Scan# 1460
Delta R.T. -0.000 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	200	Resp:	17977
Ion Ratio	Lower	Upper	
200	100		
173	20.7	18.7	28.1
215	47.8	40.2	60.4



#24
Pentachlorophenol
Concen: 3.42 ng
RT: 16.761 min Scan# 1475
Delta R.T. -0.000 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10

Tgt Ion:	266	Resp:	10020
Ion Ratio	Lower	Upper	
266	100		
264	61.6	51.2	76.8
268	63.0	52.4	78.6









#25
Phenanthrene
Concen: 3.30 ng
RT: 17.151 min Scan# 1507
Delta R.T. -0.000 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	178	Resp:	103026
Ion Ratio	Lower	Upper	
178	100		
176	18.7	15.1	22.7
179	15.2	12.3	18.5



#26
Anthracene
Concen: 3.41 ng
RT: 17.236 min Scan# 1514
Delta R.T. -0.000 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10

Tgt Ion:	178	Resp:	92432
Ion Ratio	Lower	Upper	
178	100		
176	18.1	14.6	21.8
179	15.3	12.4	18.6









#27
Fluoranthene-d10
Concen: 3.30 ng
RT: 19.147 min Scan# 1806
Delta R.T. 0.005 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:212 Resp: 103228
Ion Ratio Lower Upper
212 100
106 9.4 6.4 9.6
104 5.5 3.9 5.9



#28
Fluoranthene
Concen: 3.33 ng
RT: 19.171 min Scan# 1811
Delta R.T. -0.005 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10

Tgt Ion:202 Resp: 121696
Ion Ratio Lower Upper
202 100
101 8.1 5.4 8.2
203 17.4 13.8 20.6









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.312 min Scan# 2102
Delta R.T. -0.000 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:240 Resp: 10978
Ion Ratio Lower Upper
240 100
120 7.1 3.5 5.3#
236 27.1 20.5 30.7



#30
Pyrene
Concen: 3.18 ng
RT: 19.539 min Scan# 1885
Delta R.T. -0.000 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10

Tgt Ion:202 Resp: 121956
Ion Ratio Lower Upper
202 100
200 20.6 16.5 24.7
203 17.5 13.9 20.9









#31
 Terphenyl-d14
 Concen: 3.15 ng
 RT: 19.747 min Scan# 1927
 Delta R.T. -0.000 min
 Lab File: BN011824.D
 Acq: 17 Sep 2020 15:10




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Tgt Ion:	244	Resp:	78846
Ion	Ratio	Lower	Upper
244	100		
212	8.5	7.3	10.9
122	10.7	7.8	11.6



#32
 Benzo(a)anthracene
 Concen: 3.29 ng
 RT: 21.290 min Scan# 2100
 Delta R.T. -0.000 min
 Lab File: BN011824.D
 Acq: 17 Sep 2020 15:10

Tgt Ion:	228	Resp:	122173
Ion	Ratio	Lower	Upper
228	100		
226	27.2	21.0	31.4
229	19.3	16.0	24.0









#33
Chrysene
Concen: 3.18 ng
RT: 21.344 min Scan# 2105
Delta R.T. -0.000 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:228 Resp: 121459
Ion Ratio Lower Upper
228 100
226 29.8 23.3 34.9
229 19.4 15.8 23.8



#34
Bis(2-ethylhexyl)phthalate
Concen: 3.05 ng
RT: 21.237 min Scan# 2095
Delta R.T. 0.010 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10

Tgt Ion:149 Resp: 58313
Ion Ratio Lower Upper
149 100
167 28.3 23.6 35.4
279 4.6 4.0 6.0









#35
 Indeno(1,2,3-cd)pyrene
 Concen: 3.31 ng
 RT: 25.818 min Scan# 3202
 Delta R.T. -0.000 min
 Lab File: BN011824.D
 Acq: 17 Sep 2020 15:10

Instrument :
 BNA_N
 ClientSampleId :
 SSTDIC3.2

Tgt Ion:276 Resp: 157698




Ion	Ratio	Lower	Upper
276	100		
138	18.4	13.0	19.4
277	25.0	20.0	30.0




#36
 Perylene-d12
 Concen: 0.40 ng
 RT: 23.559 min Scan# 2542
 Delta R.T. -0.000 min
 Lab File: BN011824.D
 Acq: 17 Sep 2020 15:10

Tgt Ion:264 Resp: 10388

Ion	Ratio	Lower	Upper
264	100		
260	25.4	19.6	29.4
265	79.7	38.6	58.0









#37
Benzo(b)fluoranthene
Concen: 3.35 ng
RT: 22.885 min Scan# 2345
Delta R.T. -0.000 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:252 Resp: 129562
Ion Ratio Lower Upper
252 100
253 22.1 19.1 28.7
125 7.8 6.2 9.2



#38
Benzo(k)fluoranthene
Concen: 3.27 ng
RT: 22.929 min Scan# 2358
Delta R.T. -0.000 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10

Tgt Ion:252 Resp: 130898
Ion Ratio Lower Upper
252 100
253 22.3 19.3 28.9
125 7.7 6.0 9.0









#39
Benzo(a)pyrene
Concen: 3.35 ng
RT: 23.463 min Scan# 2514
Delta R.T. -0.000 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:252 Resp: 115414
Ion Ratio Lower Upper
252 100
253 22.5 19.8 29.6
125 8.6 7.3 10.9



#40
Dibenzo(a,h)anthracene
Concen: 3.44 ng
RT: 25.835 min Scan# 3207
Delta R.T. -0.000 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10

Tgt Ion:278 Resp: 130656
Ion Ratio Lower Upper
278 100
139 12.2 8.6 13.0
279 23.9 19.8 29.8





#41
Benzo(g,h,i)perylene
Concen: 3.28 ng
RT: 26.506 min Scan# 3403
Delta R.T. -0.000 min
Lab File: BN011824.D
Acq: 17 Sep 2020 15:10

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:276 Resp: 133129
Ion Ratio Lower Upper
276 100
277 23.7 19.4 29.0
138 16.2 11.3 16.9

