

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091725\
 Data File : BN037761.D
 Acq On : 18 Sep 2025 03:58
 Operator : RC/JU
 Sample : Q3095-12
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW202D-20250909

Quant Time: Sep 18 05:28:59 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091125.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 11 14:20:01 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)	
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.833	152	7232	0.400 ng	0.00	
7) Naphthalene-d8	10.637	136	19420	0.400 ng	-0.01	
13) Acenaphthene-d10	14.484	164	8613	0.400 ng	-0.01	
19) Phenanthrene-d10	17.235	188	15444	0.400 ng	# 0.00	
29) Chrysene-d12	21.411	240	12865	0.400 ng	0.00	
35) Perylene-d12	23.741	264	13080	0.400 ng	#-0.01	
System Monitoring Compounds						
4) 2-Fluorophenol	5.399	112	2030	0.122 ng	0.00	
5) Phenol-d6	6.988	99	1452	0.067 ng	0.00	
8) Nitrobenzene-d5	8.982	82	4795	0.338 ng	-0.01	
11) 2-Methylnaphthalene-d10	12.232	152	7932	0.308 ng	-0.01	
14) 2,4,6-Tribromophenol	15.982	330	820	0.365 ng	0.00	
15) 2-Fluorobiphenyl	13.115	172	12551	0.348 ng	0.00	
27) Fluoranthene-d10	19.266	212	14609	0.377 ng	0.00	
31) Terphenyl-d14	19.861	244	10464	0.401 ng	0.00	
Target Compounds						
2) 1,4-Dioxane	3.304	88	413	0.055 ng	#	82
24) Pentachlorophenol	16.888	266	127	0.038 ng	#	83
28) Fluoranthene	19.294	202	1128	0.021 ng	#	81
30) Pyrene	19.657	202	1175	0.023 ng	#	94
34) Bis(2-ethylhexyl)phtha...	21.339	149	3713	0.279 ng		99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091725\
 Data File : BN037761.D
 Acq On : 18 Sep 2025 03:58
 Operator : RC/JU
 Sample : Q3095-12
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW202D-20250909

Quant Time: Sep 18 05:28:59 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN091125.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 11 14:20:01 2025
 Response via : Initial Calibration

