

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091825\
 Data File : PS031771.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Sep 2025 18:47
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 02:23:15 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS090925.M
 Quant Title : 8080.M
 QLast Update : Tue Sep 09 13:56:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.286	7.744	2613.1E6	680.9E6	725.556m	749.176
Target Compounds							
1) T	Dalapon	2.672	2.693	4036.5E6	1984.5E6	698.275	704.322
2) T	3,5-DICHL...	6.453	6.696	3882.5E6	1060.6E6	727.971	695.469m
3) T	4-Nitroph...	7.097	7.297	892.9E6	803.1E6	809.403	635.944
5) T	DICAMBA	7.470	7.941	11649.7E6	4698.9E6	746.934	722.141
6) T	MCPD	7.648	8.038	683.0E6	145.9E6	76.408	77.450
7) T	MCPA	7.797	8.285	826.9E6	208.6E6	74.426	74.476
8) T	DICHLORPROP	8.178	8.660	2598.4E6	1064.9E6	729.535	731.862
9) T	2,4-D	8.414	9.002	2601.8E6	1162.5E6	759.510m	706.616
10) T	Pentachlo...	8.709	9.510	41867.5E6	29817.1E6	788.556	751.240
11) T	2,4,5-TP ...	9.287	9.895	15821.5E6	10886.1E6	754.350	723.120
12) T	2,4,5-T	9.583	10.323	14517.0E6	9692.4E6	793.356	720.250
13) T	2,4-DB	10.158	10.892	1740.4E6	712.1E6	746.046m	665.516
14) T	DINOSEB	11.355	11.267	11423.0E6	7586.7E6	792.018	780.475
15) T	Picloram	11.184	12.387	14571.4E6	17103.8E6	809.026	722.189
16) T	DCPA	11.653	12.308	20636.5E6	17740.2E6	765.737	735.982

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091825\
Data File : PS031771.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2025 18:47
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 02:23:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS090925.M
Quant Title : 8080.M
QLast Update : Tue Sep 09 13:56:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

