




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091825\
Data File : BN037775.D
Acq On : 18 Sep 2025 13:53
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 18 14:26:24 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 18 14:26:16 2025
Response via : Initial Calibration



TIC: BN037775.D\data.ms

(70) Pentachlorophenol (C)




16.886min (0.000) 277.80 ng

response 307595

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	67.44
264.00	61.60	66.00
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091825\
Data File : BN037775.D
Acq On : 18 Sep 2025 13:53
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 18 14:26:24 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 18 14:26:16 2025
Response via : Initial Calibration



TIC: BN037775.D\data.ms

(77) Benzidine

19.474min (0.000) 0.00 ng

response 1155100

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.36
183.00	13.20	12.81
0.00	0.00	0.00