




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091825\
Data File : BN037794.D
Acq On : 19 Sep 2025 02:08
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 19 04:29:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 19 04:29:39 2025
Response via : Initial Calibration



TIC: BN037794.D\data.ms

(70) Pentachlorophenol (C)




16.886min (0.000) 32518.88 ng

response 284548

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	62.85
264.00	61.60	62.23
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091825\
Data File : BN037794.D
Acq On : 19 Sep 2025 02:08
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 19 04:29:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 19 04:29:39 2025
Response via : Initial Calibration



TIC: BN037794.D\data.ms

(77) Benzidine

19.474min (0.000) 0.00 ng

response 1527830

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.88
183.00	13.20	12.20
0.00	0.00	0.00