

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091922\
 Data File : BN021826.D
 Acq On : 20 Sep 2022 05:29
 Operator : CG/JU
 Sample : N4649-02
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 EXZ18

Quant Time: Sep 20 06:33:32 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 20 00:20:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.051	152	4876	0.400	ng/u1	0.00
4) Naphthalene-d8	10.878	136	16472	0.400	ng/u1	# 0.00
9) Acenaphthene-d10	14.707	164	9835	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.460	188	20019	0.400	ng/u1	# 0.00
17) Chrysene-d12	21.647	240	12956	0.400	ng/u1	# 0.00
23) Perylene-d12	24.152	264	10352	0.400	ng/u1	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.363	96	28107	4.685	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.467	152	9839	0.377	ng/u1	0.00
18) Fluoranthene-d10	19.487	212	21655	0.449	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091922\
Data File : BN021826.D
Acq On : 20 Sep 2022 05:29
Operator : CG/JU
Sample : N4649-02
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
EXZ18

Quant Time: Sep 20 06:33:32 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 20 00:20:21 2022
Response via : Initial Calibration

