

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091922\
 Data File : BN021831.D
 Acq On : 20 Sep 2022 09:59
 Operator : CG/JU
 Sample : PB147687BL
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK687

Quant Time: Sep 21 04:22:47 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 20 00:20:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.051	152	4719	0.400	ng/u1	0.00
4) Naphthalene-d8	10.878	136	15768	0.400	ng/u1	# 0.00
9) Acenaphthene-d10	14.708	164	8890	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.456	188	17898	0.400	ng/u1	0.00
17) Chrysene-d12	21.650	240	10459	0.400	ng/u1	# 0.00
23) Perylene-d12	24.155	264	8886	0.400	ng/u1	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.363	96	28174	4.853	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.473	152	7232	0.290	ng/u1	0.00
18) Fluoranthene-d10	19.487	212	12583	0.323	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091922\
Data File : BN021831.D
Acq On : 20 Sep 2022 09:59
Operator : CG/JU
Sample : PB147687BL
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK687

Quant Time: Sep 21 04:22:47 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 20 00:20:21 2022
Response via : Initial Calibration

