

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091922\
 Data File : BN021802.D
 Acq On : 19 Sep 2022 13:42
 Operator : CG/JU
 Sample : N4595-02
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 CB8G5

Quant Time: Sep 20 00:24:25 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 20 00:20:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.055	152	4868	0.400	ng/u1	0.00
4) Naphthalene-d8	10.883	136	16463	0.400	ng/u1 #	0.00
9) Acenaphthene-d10	14.707	164	9445	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.460	188	19153	0.400	ng/u1	0.00
17) Chrysene-d12	21.650	240	11579	0.400	ng/u1 #	0.00
23) Perylene-d12	24.155	264	9340	0.400	ng/u1 #	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.363	96	25215	4.210	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.473	152	7810	0.300	ng/u1	0.00
18) Fluoranthene-d10	19.487	212	16289	0.378	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN091922\
Data File : BN021802.D
Acq On : 19 Sep 2022 13:42
Operator : CG/JU
Sample : N4595-02
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
CB8G5

Quant Time: Sep 20 00:24:25 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN090822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 20 00:20:21 2022
Response via : Initial Calibration

