

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN092019\
 Data File : BN007760.D
 Acq On : 20 Sep 2019 12:41
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 20 22:53:06 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 18 18:24:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	178#	0.00
2	1,4-Dioxane	0.446	0.389	12.8	184#	0.00
3	n-Nitrosodimethylamine	0.204	0.220	-7.8	126	0.00
4 S	2-Fluorophenol	1.377	1.117	18.9	166#	0.00
5 S	Phenol-d6	1.962	1.445	26.4#	165#	0.00
6	bis(2-Chloroethyl)ether	1.133	1.251	-10.4	223#	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	182#	0.00
8 S	Nitrobenzene-d5	0.350	0.324	7.4	173#	0.00
9	Naphthalene	1.123	1.125	-0.2	183#	-0.01
10	Hexachlorobutadiene	0.236	0.236	0.0	180#	-0.01
11 SURR	2-Methylnaphthalene-d10	0.674	0.683	-1.3	190#	0.00
12	2-Methylnaphthalene	0.759	0.780	-2.8	192#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	184#	-0.01
14 S	2,4,6-Tribromophenol	0.218	0.172	21.1	174#	-0.01
15 S	2-Fluorobiphenyl	1.647	1.527	7.3	184#	0.00
16	Acenaphthylene	2.121	1.855	12.5	179#	-0.01
17	Acenaphthene	1.368	1.238	9.5	182#	-0.01
18	Fluorene	1.753	1.601	8.7	183#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	179#	-0.01
20	4-Bromophenyl-phenylether	0.243	0.245	-0.8	184#	-0.01
21	Hexachlorobenzene	0.265	0.266	-0.4	181#	-0.01
22	Pentachlorophenol	0.098	0.081	17.3	169#	-0.01
23	Phenanthrene	1.246	1.253	-0.6	182#	-0.01
24	Anthracene	1.125	1.098	2.4	181#	-0.01
25 SURR	Fluoranthene-d10	1.348	1.353	-0.4	182#	-0.01
26	Fluoranthene	1.549	1.572	-1.5	184#	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	180#	-0.01
28	Pvrene	1.364	1.352	0.9	185#	-0.01
29 S	Terphenyl-d14	0.983	0.969	1.4	184#	-0.01
30	Benzo(a)anthracene	1.473	1.414	4.0	180#	-0.01
31	Chrysene	1.436	1.425	0.8	183#	-0.01
32	Bis(2-ethylhexyl)phthalate	0.758	0.817	-7.8	217#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.797	1.713	4.7	176#	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	173#	-0.02
35	Benzo(b)fluoranthene	1.389	1.338	3.7	177#	-0.01
36	Benzo(k)fluoranthene	1.373	1.384	-0.8	182#	-0.01
37 C	Benzo(a)pyrene	1.305	1.245	4.6	174#	-0.01
38	Dibenzo(a,h)anthracene	1.335	1.277	4.3	173#	-0.02
39	Benzo(a,h,i)perylene	1.321	1.240	6.1	169#	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN092019\
Data File : BN007760.D
Acq On : 20 Sep 2019 12:41
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 20 22:53:06 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 18 18:24:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			