

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN092019\
 Data File : BN007760.D
 Acq On : 20 Sep 2019 12:41
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 20 22:53:06 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 18 18:24:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	178	0.00
2	1,4-Dioxane	0.400	0.349	12.8	184	0.00
3	n-Nitrosodimethylamine	0.400	0.431	-7.7	126	0.00
4 S	2-Fluorophenol	0.400	0.325	18.8	166	0.00
5 S	Phenol-d6	0.400	0.336	16.0	165	0.00
6	bis(2-Chloroethyl)ether	0.400	0.442	-10.5	223	-0.02
7 I	Naphthalene-d8	0.400	0.400	0.0	182	0.00
8 S	Nitrobenzene-d5	0.400	0.370	7.5	173	0.00
9	Naphthalene	0.400	0.401	-0.3	183	-0.01
10	Hexachlorobutadiene	0.400	0.399	0.3	180	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.405	-1.3	190	0.00
12	2-Methylnaphthalene	0.400	0.411	-2.7	192	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	184	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.316	21.0	174	-0.01
15 S	2-Fluorobiphenyl	0.400	0.371	7.3	184	0.00
16	Acenaphthylene	0.400	0.350	12.5	179	-0.01
17	Acenaphthene	0.400	0.362	9.5	182	-0.01
18	Fluorene	0.400	0.365	8.8	183	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	179	-0.01
20	4-Bromophenyl-phenylether	0.400	0.404	-1.0	184	-0.01
21	Hexachlorobenzene	0.400	0.401	-0.3	181	-0.01
22	Pentachlorophenol	0.400	0.330	17.5	169	-0.01
23	Phenanthrene	0.400	0.402	-0.5	182	-0.01
24	Anthracene	0.400	0.390	2.5	181	-0.01
25 SURR	Fluoranthene-d10	0.400	0.401	-0.3	182	-0.01
26	Fluoranthene	0.400	0.406	-1.5	184	-0.01
27 I	Chrysene-d12	0.400	0.400	0.0	180	-0.01
28	Pvrene	0.400	0.396	1.0	185	-0.01
29 S	Terphenyl-d14	0.400	0.394	1.5	184	-0.01
30	Benzo(a)anthracene	0.400	0.384	4.0	180	-0.01
31	Chrysene	0.400	0.397	0.8	183	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.431	-7.7	217	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.381	4.8	176	-0.02
34 I	Perylene-d12	0.400	0.400	0.0	173	-0.02
35	Benzo(b)fluoranthene	0.400	0.385	3.8	177	-0.01
36	Benzo(k)fluoranthene	0.400	0.403	-0.8	182	-0.01
37 C	Benzo(a)pyrene	0.400	0.382	4.5	174	-0.01
38	Dibenzo(a,h)anthracene	0.400	0.383	4.3	173	-0.02
39	Benzo(a,h,i)perylene	0.400	0.375	6.3	169	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN092019\
Data File : BN007760.D
Acq On : 20 Sep 2019 12:41
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 20 22:53:06 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN091819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 18 18:24:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				